FEDERAL SPECIFICATION

SPECULUM SET, EAR

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.

1. SCOPE

1.1. This specification covers one type of ear speculum, suitable for use in visual examination and treatment of the ear.

2. APPLICABLE SPECIFICATIONS AND STANDARDS

2.1 The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

Federal Specifications:

GG–I–526 — Instruments, Dental and Surgical; General Specifications.

QQ–B–626 — Brass, Leaded and Non-Leaded; Rods, Shapes, Forgings and Flat Products with Finished Edges (Bars, Flat Wire and Strips).

Federal Standards:

Fed. Std. No. 102 — Preservation, Packaging, and Packing Levels.

Fed. Std. No. 123 — Marking for Domestic Shipment (Civilian Agencies)

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications, Standards, and Handbooks and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U. S. Government Printing Office, Washington 25, D. C.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D. C.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications, Standards, and Handbooks from established distribution points in their agencies.)

Military Specifications:

MIL–P–12902 — Packaging of Surgical and Veterinary Instruments, Diagnostic Equipment, and Related Items.

Military Standards:

MIL–STD–105 — Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129 — Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material.—The ear speculum shall be fabricated of class 4, composition N material in accordance with Federal Specification GG–I–526, or of composition 22 material, in accordance with Federal Specification QQ–B–626, chromium plated over nickel. Both materials shall be in accordance with plating requirements and corrosion tests requirements for class 4, composition N as specified in Federal Specification GG–I–526.

3.2 Style, design, and dimensions.—Shall be a set of four ear specula, Gruber design, sizes 0, 1, 2, and 3, nested in metal case, suitable for visual examination and treatment of the ear. Shall be as shown on figure 1.

3.3 Plating. — The entire surface of the specula and case shall be plated, except the omission of chromium plating in the deep lying inner surfaces shall be acceptable.


Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

3.4 Workmanship.—Workmanship shall be first class throughout. Speculum sets shall be free from defects which detract from their appearance or may impair their serviceability.

4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Supplier responsibilities for inspection.—Unless otherwise specified herein, the supplier is responsible for the performance of all inspection requirements prior to submission for Government inspection and acceptance. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. Inspection records of the examinations and tests shall be kept complete and available to the Government as specified in the contract or order.

4.1.1 Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances



SIZE	A	B	C	D	E	F
0	1 15/64	.565–.590	.490	3/32	.125	.093
1	1 5/16	.627–.640	.540	1/8	.156	.125
2	1 25/64	.681–.690	.590	5/32	.218	.156
3	1 15/32	.713–.740	.640	3/16	.250	.187

FIGURE 1.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:52Z
Check the source to verify that this is the current version before use.

or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 *Certificates of quality.*—When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

4.2 Sampling.

4.2.1 *For examination.* — Sampling for examination shall be conducted in accordance with Military Standard MIL–STD–105.

4.2.2 *For tests.* — Sampling for tests shall be conducted in accordance with Military Standard MIL–STD–105 (see table I of this specification).

4.2.3 *Raw materials.* — Four ounces of each lot of each metal used in the construction of the speculum set shall be selected for chemical analysis.

4.2.4 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination.—The speculum set shall be examined to determine compliance with all requirements contained in this specification.

4.4 Tests.—Tests shall be conducted, as necessary, to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

For civil agency procurement, Federal Standard No. 102 should be referred to for definitions and applications of the various levels of packaging and packing protection for supplies and equipment.

5.1 Packaging and packing.

5.1.1 *Unit of issue.* — One speculum set, as specified, constitutes one unit of issue.

5.1.2 *Packaging quantities.* — The number of units of issue indicated in table II shall be packaged in each unit, intermediate and exterior container, as applicable for the required level of protection specified in the procurement document.

5.1.2.1 *Packing variation permitted.* —If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which will insure safe delivery to destination.

TABLE II. — *Packaging quantities*

Unit package	*Intermediate package*	*Exterior container*
1 unit	12 units	360 units

TABLE I

Component	Characteristic	Req.	Test procedure	Unit of product	Insp. level	AQL
Speculum set (end item)----	Boil test	GG–I–526	GG–I–526	1-speculum set	L3	*
	Salt spray	GG–I–526	GG–I–526	1-speculum set	L3	*
	Thickness of coating	GG–I–526	GG–I–526	1-speculum set	L3	*

* Acceptance number shall be zero for all sample sizes.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.

5.1.3 *Level A.* — Packaging and packing shall conform to the requirements of Military Specification MIL–P–12902, Procedure Code No. 35.

5.1.4 *Level B.* — Packaging and packing shall conform to the requirements of Military Specification MIL–P–12902, Procedure Code No. 21.

5.1.5 *Level C.* — The subject commodity shall be packaged and packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.2 Marking—nonmilitary.

5.2.1 Unless otherwise specified in the contract or order unit packages and shipping containers shall be marked in accordance with Federal Standard No. 123.

5.3 Marking—military.

5.3.1 *Interior packages.* — Marking shall conform to the requirements of Military Specification MIL–P–12902, except that the ASMPA stock number will not be required.

5.3.2 *Exterior container.* — Exterior container shall be marked in accordance with Military Standard MIL–STD–129.

6. NOTES

6.1 Ordering data. — Purchasers should exercise any desired options offered herein and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Selection of applicable levels of packaging and packing (see 5.1).

(1) Level B preservation and packing is intended to provide economical but limited protection, and should be specified only when it is determined the speculum sets will be held in covered storage approximately 12 months from date of initial packaging.

6.2 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet Federal Government requirements.

6.3 Transportation description.—Transportation description and minimum weights applicable to this item are:

Rail:

Surgical instruments, not otherwise indexed by name.

Carload minimum weight 10,000 pounds, subject to Rule 34, Uniform Freight Classification.

Motor:

Surgical instruments, not otherwise indexed.

Truckload minimum weight (w) 10.6 pounds, (w) subject to Rule 34, National Motor Freight Classification.

Notice. — When Government drawings, specifications, or other data are used for any purpose other than in connection with a definitely related Government procurement operation, the United States Government thereby incurs no responsibility nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data, is not to be regarded, by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use, or sell any patented invention that may in any way be related thereto.

CUSTODIANS:

Army—Army Medical Service

Navy—Bureau of Medicine and Surgery

Air Force.

Copies of this specification may be purchased for 5 cents each.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:52Z
Check the source to verify that this is the current version before use.