A-A-53708
26 September 1988

COMMERCIAL ITEM DESCRIPTION

STONE, SHARPENING (Fine India)

The General Services Administration has authorized the use of this Commercial Item Description.

This commercial item description covers an artificial aluminum oxide stone suitable for sharpening dental instruments.

Salient characteristics.

The stone shall be a round edge, slip style, fine India sharpening stone. The stone shall be constructed of aluminum oxide abrasive suitably ball milled, blended with bonding agents, pressed, and kiln fired to produce a fine grit stone. The stone shall be pre-filled with oil and shall be capable of producing an initial sharpened edge. The abrasive content of the stone shall be 65 percent to 85 percent. The grit size of the stone shall be 280 to 320. The specific gravity of the stone shall be 1.95 to 2.57. Style, design and dimensions of the stone shall be in accordance with figure 1.

Identification marking. Each stone shall be prominently labeled, in a permanent and legible manner with the manufacturer's name or registered trademark. In addition, the stone or immediate container shall be marked or labeled to indicate that lubrication with oil is required when using the stone.


Workmanship. The sharpening stone shall possess sufficient strength to withstand normal service without chipping, breaking, or otherwise becoming damaged. The stone shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One sharpening stone, as specified, constitutes one unit.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A: Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.



NOTE:
UNLESS OTHERWISE SPECIFIED
ALL DIMENSIONS ARE IN INCHES

FIGURE 1 STONE, SHARPENING

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Certificates of quality. Certificates of quality, supplied by the manufacturer of the component or materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government.

Sampling.

For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 2.5 (percent defective) and an inspection level of S-2. The unit of product for examination purposes shall be one stone.

For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 at inspection level S-1, AQL 1.0 (percent defective). The unit of product for test purposes shall be one stone. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

Examination. The stones shall be examined to determine compliance with specification requirements. The stones shall be examined for defects including, but not limited to, those listed below:

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:23Z
Check the source to verify that this is the current version before use.

Defects.

Style or design of stone not as specified.
Stone has cracks, breaks, chips, or other defects affecting function or serviceability.
Dimensions of stone not as specified.
Stone not pre-filled with oil.
Identification marking or labeling not as specified.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Abrasive content. Prepare and weigh a sample of stone and place in a lead dish. Overspread sample with hydrofluoric acid; cover and let stand 24 hours. Pour off acid, break remains with a stirring rod, wash by decantation, dry and weigh residuum. The abrasive percentage shall be calculated as follows:

$$\text{Abrasive percentage} = \frac{B}{A} \times 100$$

A = Weight of Sample
B = Weight of Residuum

Grit size. Grit size shall be determined in accordance with Department of Commerce Commercial Standard PS8.

Specific gravity. Weigh a dry sample of stone. Dip sample in molten paraffin wax, until completely coated, and weigh again. Reweigh coated sample under water. Calculate the specific gravity of the stone as follows:

$$\text{Specific gravity} = \frac{A}{B-\left(\frac{B-A}{S} + C\right)}$$

A = Weight of sample
B = Weight of coated sample
C = Weight of submerged sample
S = Specific gravity of paraffin

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:23Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-263-2559	STONE, SHARPENING, Dental Instrument, Fine India, Aluminum Oxide, Round Edge, Slip Style, 4.5 inches (11.4 cm) Long, 1-3/4 inches (4.4 cm) Wide.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:23Z
Check the source to verify that this is the currentversion before use.

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

CIVIL AGENCY COORDINATING ACTIVITIES:
VA-OSS
PHS
FDA-MPQAS

61328

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

PROJECT NO. 6520-2074

LOCATION: NCR Disk No. 7

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/10490

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO. 4966 WASHINGTON D C.

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
2800 South 20th Street
Philadelphia, Pennsylvania 19101-8419

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
A-A-53708	STONE, SHARPENING (Fine India)

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)*:

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.