A-A-53715
24 October 1988

COMMERCIAL ITEM DESCRIPTION

STOOL, DENTAL LABORATORY, MOBILE, ADJUSTABLE

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a mobile dental laboratory stool suitable for use by laboratory technicians, dentists or dental assistants.

Salient Characteristics:

The laboratory stool shall be mobile with an adjustable seat and backrest body support. The stool shall consist of five (5) casters, a base assembly with a fixed footrest ring, a seat height adjustment mechanism, a seat support spindle, a seat and a backrest assembly.

The casters shall be a 2 inch ball bearing type, with replaceable wheels. The base shall be 14 gauge tubular steel, forming four legs with a diagonal base spread of 22 inches. Welded to the juncture of the four legs shall be a 14 gauge base barrel which shall be constructed so as to accurately receive the seat support spindle. The fixed footrest ring shall be 17 inches size diameter, of bright chrome finished steel and shall be attached to the base by hardened steel screws. The base assembly, except for the chrome footrest ring, shall be enameled silver in color. The seat height adjustment mechanism shall be a spring-loaded, fingertip lever activated, lock-plunger type and shall be welded to and near the top of the base barrel. The spring shall have a nominal force of 30 lbs. This mechanism shall provide safe and fast seat height adjustment. The seat support spindle shall fit accurately into the barrel base and be non-rotating. The seat support spindle shall contain holes to receive the spring load plunger of the seat height adjustment mechanism so as to provide secure and positive seat height adjustments in a range of 17.5 to 23.0 inches. The top of the spindle shall be radiused to form a ball bearing on which the seat can rotate. There shall be a grooved notch in the spindle beneath the ball bearing to receive a seat retaining set screw so as to allow full seat rotation without interference.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.




FIGURE 1 STOOL, DENTAL LABORATORY, MOBILE

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

The seat dimensions shall be 17.5 by 14.5 by 2.25 inches. The seat base shall be one piece aluminum casting constructed to accurately fit the radiused seat support spindle and be padded with two inch polyurethane foam. The backrest shall be 13 by 8 inch by 1 inches and attached to the chair by "L" shaped uprights which shall provide three inches of vertical and horizontal adjustment. The backrest shall also be padded with polyurethane foam and shall be attached to the top of the "L" shaped upright by a spring balanced pivoting box which will allow the backrest to pivot to the action range of the user and automatically return to the correct posture position. The seat and backrest covers shall be flame retardent black vinyl. The style and design of the stool shall be in accordance with Figure 1. The weight of the stool shall be 27 pounds nominal.

Workmanship. The stool shall be free from burrs, cracks, tears, or any other defects which will impair its serviceability, proper function or intended use.

Unit. Each (EA). One stool, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Sampling for inspection. Sampling for inspection shall be conducted in accordance with MIL-STD-105. Where feasible, the same sample shall be used for the determination of two or more inspection characteristics. The unit of product for inspection purposes shall be one dental stool. Unless otherwise specified, sampling for examination shall be conducted at inspection level II, except for dimensions which shall be conducted at inspection level S-1, AQL 2.5, and sampling for tests (when applicable) shall be conducted at inspection level S-1, AQL 1.0.

Examination. Examination shall include but not be limited to the following:

Defects.

Dimensions of the stool and accessories not as specified.
Chair legs out of plumb and not level.
Welding of legs to base barrel not uniform.
Aluminum seat plate has significant casting defects.
Vertical alignment of spindle not as specified.
Material components not as specified.
Metal finish not as specified.
Chair spindle has rough edges.
Plunger pin difficult to operate.
Chair seat assembly does not move freely on spindle.
Weight of stool not as specified.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

National Stock Number	Item Identification
6520-01-121-0696	STOOL, DENTAL LABORATORY, MOBILE, ADJUSTABLE.

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

49119

Project No. 6520-2019

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

(TO DETACH THIS FORM, CUT ALONG THIS LINE)

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE
A-A-53715	STOOL, DENTAL LABORATORY, MOBILE, ADJUSTABLE

3a. NAME OF SUBMITTING ORGANIZATION

b. ADDRESS *(Street, City, State, ZIP Code)*

4 TYPE OF ORGANIZATION *(Mark one)*

- ☐ VENDOR
- ☐ USER
- ☐ MANUFACTURER
- ☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil – Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-RSTE
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:23Z
Check the source to verify that this is the current version before use.