INTERIM COMMERCIAL ITEM DESCRIPTION

DRAINAGE UNIT, PLEURAL CAVITY

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a package of disposable pleural cavity drainage units.

Salient characteristics:

The drainage unit shall be a one (1) bottle pleural cavity drainage unit. It shall be sterile and disposable and shall provide a water seal suitable for use in closed drainage procedures with or without suction.

The bottle shall have a capacity of 2400 ML (+/- 5 percent) and be graduated in increments of 50 ML up to 2000 ML. It shall be fabricated of suitable transparent, clear plastic material. The bottle shall be provided with all necessary access ports, air vent fittings, a drain tube, and a dust cover.

The drain tube shall be fabricated of suitably flexible, transparent plastic material and be provided with a chest tube connector. It shall be 72 inches (+/- 10 percent, 183 cm) long. The tube shall be firmly affixed to the drainage bottle.

The unit shall be provided with an integral base for stability.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the current version before use.

Labeling shall meet the requirements of the Federal, Food, Drug and Cosmetic Act and shall be that usually furnished by the supplier as commercial practice.

The drainage unit shall be sterilized in a manner acceptable to the Food and Drug Administration.

The unit shall be packaged in a suitable container which shall maintain sterility until manually opened.

Workmanship. The drainage unit shall be free from defects which detract from its/their appearance or impair its/their serviceability.

Unit. Package (PG). One package containing six individually packaged sterile drainage apparatus, as specified constitutes one unit. The individual package shall be in a container capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of S2 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one drainage unit, as specified in this description.

DEFECTS

Not packaged 6 per unit.

Each drainage unit is not individually sealed in a separate sterile sub-package.

Graduation markings on bottle are not present in 50 ml increments up to 2000 ml.

Unit does not have a base plate.

Sealed package is not clean.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-01-023-2903

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4498

PGC 46892

Location: CID/DRAINAGE/H12-25

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/10561

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.