A-A-53746
17 October 1988

COMMERCIAL ITEM DESCRIPTION

WATCH GLASS, LABORATORY

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification DD-W-131C for Class B watch glasses.

This Commercial Item Description covers soda-lime watch glasses intended for use as coverings for laboratory ware.

Salient characteristics.

Shall be a watch glass intended for general laboratory uses, such as covering beakers, jars, dishes, etc.

Shall be of the following millimeter (mm) diameter sizes:

25, 35, 50, 65, 75, 90, 100, 125, 137.5, 150, 175, 200

A tolerance of plus or minus 2 millimeters shall be applicable to each diameter size.

Shall be circular, with the edge around the circumference finely ground or uniformly polished.

Shall be concave to prevent sliding when used as a laboratory wear covering.

Shall be fabricated of soda-lime glass conforming to the requirements of Type II glass in ASTM Specification E 438 (Standard Specification for GLASSES IN LABORATORY APPARATUS).

Shall be suitably annealed to demonstrate the following capabilities:

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Showing a ring strain not greater than that of temper No. 1, based on testing procedures for determining annealing strain contained in ASTM Specification C 148 (Standard Specification for POLARISCOPE EXAMINATION OF GLASS CONTAINERS). However, any longitudinal strain that may be present shall be faint and highly diffused.

Withstanding the shock (i.e. not breaking) caused by an initial suspension in boiling water for five (5) minutes followed by an immediate removal and plunging into water of 16 to 20 degrees (C).

Withstanding the effects of suspension in boiling water (i.e. distilled) for six (6) hours without any evidence of corrosion, scumming, chipping or cracking.

Workmanship. The watch glass shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Package (PG). One package containing twelve watch glasses with the same diameter size, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspections and tests. Inspections and tests shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Sampling for inspection and tests. Sampling for inspection and tests shall be conducted in accordance with commercial practice.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Reference standard. Where standard is referenced herein, the use of that standard in effect on the date of solicitation shall be used to determine compliance with stated requirements.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/ contractor shall comply, and be responsible for compliance by its

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	Item Identification	Specification designated size
6640-00-447-4400	WATCH GLASS, LABORATORY	50 mm
6640-00-447-4600	WATCH GLASS, LABORATORY	137.5 mm

Ordering data. (Diameter size required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order).

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6640-1525

Location: Disk No. 16
CID/WATCH/GLASS/LAB/S31

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.