A-A-53753
28 October 1988

COMMERCIAL ITEM DESCRIPTION

MANIKIN, RESUSCITATION TRAINING

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a life-like training manikin suitable for the purpose of training personnel in resuscitation methods, the control of bleeding in a wound, and leg splinting/traction.

Salient characteristics.

The manikin shall resemble a life size human body. The manikin shall have a construction of primarily molded plastic with a surface which has the appearance of human skin.

The manikin shall be approximately 68 inches in length. Total weight shall not exceed 65 pounds. The head and shoulders shall be proportionately the same weight as a human being.

The manikin shall be capable of simulating carotid pulse pressure so that it may be properly measured.

The head of the manikin shall be moveable and the jaws articulating. The eyes shall be capable of simulating a dilated and contracted pupil. The head and facial features shall be constructed to resemble a male. The head section shall be removeable to demonstrate how the air passage can be obstructed and why natural and artificial respiration depends on the movement of free air.

AMSC/NA FSC 6910

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

The manikin shall have a visual device which will indicate when heart lung resuscitation is being performed correctly. The air pathway shall be such that air blown into the mouth by the trainee shall be vented away from the trainee. Lung capacity and resistance shall be approximately that of an unconscious human being. The manikin shall have approximately the same breast bone reaction as an average adult unconscious person.

The manikin shall articulate at the shoulders, elbows, hips, and knees. One arm shall be constructed in such a manner as to simulate a penetrating wound. The wound shall be complete with arterial and venous bleeding. The arm shall be constructed to provide for bleeding control through the application of pressure to the wound. One leg shall be constructed to allow the instructor to simulate a broken femur for the purpose of teaching proper splinting/traction.

The manikin shall be clothed in neat fitting, simple apparel. Under clothing and socks are not required.

The manikin and clothing shall readily lend themselves to cleaning by ordinary methods. Each manikin shall be supplied with plastic film faces for the maintenance of clean and sanitary conditions throughout a normal lung resuscitation instruction period.

The manikin shall be furnished with at least, two fabric carrying cases. Each manikin shall be supplied with complete instructions for use and all necessary accessories for training.

Workmanship. The manikin shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One manikin, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:24Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defines as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specification, standard, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

NSN	ITEM IDENTIFICATION
6910-00-516-5436	MANIKIN, RESUSCITATION TRAINING, Plastic, Adult, Male

Ordering data. Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6910-0018

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:24Z
Check the source to verify that this is the current version before use.