
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.
This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

USER ACTIVITIES:
ARMY - ME
NAVY - MC, YD
AIR FORCE - 99

REVIEWER ACTIVITIES:
ARMY - MD
NAVY - MS
AIR FORCE - 03

FED. SUP CLASS
6630



DASH NUMBER	TYPE	SPECIFIC GRAVITY RANGE	SPECIFIC GRAVITY INTERVALS	BAUME RANGE	BAUME INTERVALS	SALINITY (PERCENT) RANGE	SALINITY (PERCENT) INTERVALS	LENGTH MM
-1	I	1.000-2.000	0.010	0-72°	1.0°	---	---	300
-2	II	0.890-1.000	0.002	26-10°	0.5°	---	---	150
-3	III	---	---	---	---	0-100	1	300
-4				D E L E T E D				
-5	III	---	---	---	---	0-26.4	0.5	300
-6	I	---	---	19-31°	0.1°	---	---	300
-7	I	---	---	0-35°	0.5°	---	---	300
-8	I	---	---	0-70°	1.0°	---	---	300
Ⓑ -9	*	0.700-2.000	0.010	10-70L 0-72H	1.0°	---	---	380
-10	II	0.690-0.800	0.002	72-45°	0.5°	---	---	150
-11	II	0.790-0.900	0.002	46-25°	0.5°	---	---	150
-12	I	1.000-1.220	0.002	0-26°	0.2°	---	---	300
-13	I	1.200-1.420	0.002	24-43°	0.2°	---	---	300
-14	I	1.400-1.620	0.002	42-55°	0.2°	---	---	300
-15	I	1.600-1.820	0.002	54-65°	0.2°	---	---	300
-16	I	1.000-1.225	0.005	0-26°	0.5°	---	---	150
-17	I	1.200-1.425	0.005	24-43°	0.5°	---	---	150
-18	I	1.400-1.625	0.005	42-55°	0.5°	---	---	150
-19	I	1.000-2.000	0.020	0-72°	1.0°	---	---	150

*Universal Range

MATERIAL: Ⓑ Hydrometer stems and bodies shall be of borosilicate glass in accordance with ASTM E 438, Type I, Class A.

Ballast shall be of mercury, lead shot or fused metal alloy.

Scales shall be printed on No. 1 white sulfite or white ledger paper.

Graduation lines and inscriptions shall be made with permanent black marking material, such as India ink.

DESIGN: Hydrometers shall be of the constant-mass, variable displacement type, and shall have a stem and body.

Ⓑ Scale shall be firmly cemented to the glass stem. Every tenth division shall be numbered.

Ⓑ Ballast shall be confined to the hydrometer bulb and be prevented from shifting by a glass partition, or by a cement that will not soften at temperatures of normal use.

Hydrometers shall be calibrated to read directly the Specific Gravity and Baume Gravity at 60°F.

Accuracy rating of the hydrometers shall be porm 1.5 times the minimum graduation intervals.

APPROVED 7 OCT 1968 REVISED Ⓐ 15 JULY 1980 Ⓑ 18 JUN 1986

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		HYDROMETER, GRADUATED SCALE, GENERAL USE	MS36538
Procurement Specifications GG-H-925	SUPERSEDES:		SHEET 1 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE PROJECT NUMBER 6630-0438-1 REVISION

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:42Z
Check the source to verify that this is the current version before use.

FED. SUP CLASS
6630

The hydrometers shall be of the following types:

Type I. For liquids heavier than water.

Type II. For liquids lighter than water.

Type III. For brine testing (used in curing meat products).

For liquids lighter than water, use cylinders with a minimum of 80 mL of liquid for submersion. For liquids heavier than water, use cylinders with a minimum of 250 mL of liquid for submersion. For universal hydrometer, use a cylinder with a minimum of 325 mL of liquid for submersion. The cylinders shall permit the hydrometer to float freely.

The hydrometers shall be permanently marked with an inscription which will indicate the use of the instrument, the type of liquid for which the instrument is intended and the temperature at which it is used.

NOTES:

1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements as specified herein.

2. For design feature purposes, this standard takes precedence over procurement document referenced herein.

3. Referenced documents shall be of the issue in effect on date of invitations for bid.

Ⓑ 4. FSCMs: 05668, 15481, 22353, 22527, 53629, 59728, 64484, and 87664.

USER ACTIVITIES:

REVIEWER ACTIVITIES:

REVISED

APPROVED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		HYDROMETER, GRADUATED SCALE, GENERAL USE	MS36538
Procurement Specifications GG-H-925	SUPERSEDES:		SHEET 2 OF 2

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

☆U.S. GOVERNMENT PRINTING OFFICE: 1986—605-036/43858

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:42Z
Check the source to verify that this is the current version before use.