[INCH-POUND]
A-A-59380
3 December 1998
SUPERSEDING
MIL-T-38760
1 April 1966

COMMERCIAL ITEM DESCRIPTION

TENSIOMETER, WIRE ROPE

The General Services Administration has authorized the use of this Commercial Item Description for all federal agencies.

1. SCOPE. This Commercial Item Description (CID) outlines the minimum requirements for three types of hand-held wire rope (cable) tensiometers. The tensiometers will be used to measure static tension in aircraft control cables. The cables range from 1/16 inch to 5/16 inch diameter. These commercially-available tensiometers measure tension up to 200, 450, and 1600 pounds within ± 5 percent of the applied load. Only those products that meet or exceed the requirements outlined in this CID are acceptable.

2. CLASSIFICATION. The tensiometers are classified by tension ranges as follows:

Type I	300	to	1600 pounds
Type II	10	to	200 pounds, direct reading
Type III	150	to	450 pounds, direct reading

3. SALIENT CHARACTERISTICS.

3.1 General. The tensiometer shall be easy to operate and read. It shall clamp quickly and securely onto cables. The tensiometer shall be capable of being operated using one hand and with bulky gloves. It shall have a zero adjustment setting. The tension reading shall be capable of being locked, so that the tension can be read after the tensionmeter is removed from the cable.

3.2 Type I. The Type I tensiometer may include a calibration chart outlining the readings for each increment of load and each cable diameter. The chart, if used, shall be dated, certified, and attached to the back of the tensiometer or to the inside of the carrying case. It shall be protected by transparent material.

Beneficial comments, recommendations, additions, deletions, clarifications, etc., and any data which may improve this document should be sent to: San Antonio Air Logistics Center, SA-ALC/TILDD, Building 171, Post A-16, 485 Quentin Roosevelt Road., Kelly AFB, TX 78241-6425 by using the DD Form 1426, Standardization Document Improvement Proposal, appearing at the end of this document or by letter.

AMSC N/A FSC 6635

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:12Z
Check the source to verify that this is the current version before use.

3.2.1 Adapters. Interchangeable cable adapters, if used with Type I tensiometers, shall be securely attached to the tensiometer during use. Each adapter shall be marked to indicate which cable diameters it is used with.

3.2.2 Type I dimensions and weight. When the tensiometer is in the storage configuration, the dimensions shall be not greater than 6.5 inches wide x 4.5 inches long x 1.5 inches thick. Weight shall be not greater than 3 pounds.

3.3 Types II and III. The Type II and III tensiometers shall be direct reading. Interchangeable cable adapters shall not be used.

3.3.1 Cable diameter gage. The tensiometer shall include an integral gage that indicates the diameter of the cables.

3.3.2 Test bar. A test bar shall be provided to check the calibration of the tensiometer. When the tensiometer is set to the smallest cable diameter and 5 readings are averaged, the tensiometer shall indicate the load marked on the bar within ± 2 percent of the value marked. The tensiometer's serial number shall be marked on the bar.

3.3.3 Types II and III dimensions and weight. When the tensiometer is in use, the overall dimensions shall be not greater than 6 inches wide x 9.5 inches long x 2 inches thick. The head of the tensiometer, that clamps onto the cable, shall be not greater than 4.5 inches wide x 1.625 inches long (measured from the end of the tensiometer and including the cable diameter) x 1.25 inches thick. Weight shall be not greater than 3 pounds.

3.4 Cable characteristics. The tensiometer shall indicate static tension in 7 x 7 and 7 x 19 cables in accordance with MIL-W-83420 "Wire Rope, Flexible, for Aircraft Control" Type I (non-jacketed) and Table I, below. The tensionmeter shall not be damaged by cable tensions that exceed specified ranges.

TABLE I. Cable characteristics and tensiometer accuracy.

Type	Tension Range (lbs.)	Cable Diameter (Inches)								Accuracy +/- (percent of applied load)
		1/16	3/32	1/8	5/32	3/16	7/32	1/4	5/16	
I	300 to 550					X	X	X	X	5%
	550 to 1600					X	X	X	X	2.5%
II	10 to 200	X	X	X						5% or 2.5 lbs. [1]
	30 to 200				X	X				5% or 2.5 lbs. [1]
	80 to 200						X	X		5%
III	150 to 450			X	X	X	X	X		5%

[1] whichever is greater.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:12Z
Check the source to verify that this is the current version before use.

3.5 Accuracy. The tensiometer shall indicate the tension in the cables listed in Table I above within the accuracy specified in Table I. Example: Type II shall be accurate within ± 5 percent of the applied load, or ± 2.5 pounds, whichever is greater.

3.6 Calibration. Each tensiometer shall be calibrated in accordance with the procedure in Technical Order (T.O.) 33K6-4-2153-1. A calibration report, including serial number and test data, shall be provided with each tensiometer. A calibration label shall be applied to each tensiometer indicating who performed the calibration and the date completed.

3.7 Environmental requirements. The tensiometer shall operate and indicate accurately in environments between minus 20 and plus 120 degrees Fahrenheit with relative humidity between 10 and 95 percent.

3.8 Marking. Each tensiometer shall be marked with the manufacturer's name, the model number, serial number, cable diameters (in inches), tension range (in pounds), and accuracy. The marking shall be legible and permanent.

3.9 Operating instructions. Operating instructions shall be provided with each tensiometer.

3.10 Carrying case. The tensiometer shall be provided with a durable carrying case that shall protect the tensiometer and all accessories from damage during use, storage, and transport.

4. REGULATORY PROVISIONS.

In accordance with Section 23.403 of the Federal Acquisition Regulation (FAR), the Federal Government's policy is to acquire items composed of the highest percentage of recoverable materials practicable consistent with maintaining a satisfactory level of competition and without adversely affecting performance requirements or exposing the supplier's employees to undue hazards from the recovered materials.

5. QUALITY ASSURANCE PROVISIONS.

5.1 Product conformance. The products provided shall meet the salient characteristics of this CID, conform to the producer's own drawings, specifications, standards, and quality assurance practices, and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance.

5.2 Market acceptability. The offered tensiometer must have been on the commercial market for at least 2 years. At least 50 tensiometers must have been sold to commercial (non-military) customers.

5.3 Product sample(s). Product samples are required for examination and testing, to deter-mine whether the offered tensiometer meets the requirements of this CID, provides accurate and repeatable measurements, and has acceptable form, fit, function and suitability for use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:12Z
Check the source to verify that this is the current version before use.

6. PACKAGING.

6.1 Preservation, packing, or marking requirements shall be as specified in the contract or order.

7. NOTES.

7.1 Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, provided they fall within the tolerances specified in conversion tables contained in the latest version of FED-STD-376 and all other requirements of this CID are met. If the manufacturer elects to use metric dimensions as the basis upon which to derive or base a mechanical feature, a request shall be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

7.2 Source of documents.

7.2.1 Document ordering. Copies of federal and Department of Defense specifications, standards, and handbooks are available from the Defense Automated Printing Service, (DAPS), 700 Robbins Avenue, Building 4D, Philadelphia, PA 19111-5094. Phone 215/697-2197.

7.2.2 T.O. 33K6-4-2153-1. A copy of T.O. 33K6-4-2153-1 “Calibration Procedure for Aircraft Cable Tensiometers” may be obtained from: SA-ALC/LDEE, Kelly AFB, TX 78241-6426, Phone 210/925-7153 or AFMETCAL Det 1/MLLW, 813 Irving Wick Dr. W., Ste 4M, Heath, OH 43056-6116.

7.3 Ordering data. The following information should be provided at the time of order:
a. Title, CID document number, and date of document.
b. Type of tensiometer required (see Section 2).
c. Number of product samples required. (see 5.3).
d. Packaging requirements (see 6.1).

7.4 Cross-reference. The following is a cross-reference between the classification types in MIL-T-38760 and this CID:

TABLE II. Type cross-reference.

MIL-T-38760 Type	A-A-59380 Type	Range (lbs.)
C7	I	300 to 1600
C8	II	10 to 200
C9	III	150 to 450

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:12Z
Check the source to verify that this is the current version before use.

7.5 Part or Identifying Number (PIN). The following part identification numbering procedure is for government purposes and does not constitute a requirement for the contractor:

AA59380 – 1

Type: 1 - Type I 300 to 1600 pounds
2 - Type II 10 to 200 pounds
3 - Type III 150 to 450 pounds

CID Document Number

7.6 Possible acceptable products. The following commercial products may meet the salient characteristics of this CID:

TABLE III. Possible acceptable products.

Type	Manufacturer	Model No.	National Stock Number
I	Opti Manufacturing Corp.	T5-8005-110-000	6635-00-530-1130
II	Sigmation, Inc. Opti Manufacturing Corp.	19738 (Type C8) T60 1001C8-1A	6635-01-455-6940 6635-00-530-1128
III	Sigmation, Inc. Opti Manufacturing Corp.	19739 (Type C9) T60 1001C9-1A	6635-01-375-2019 6635-00-530-1129

MILITARY INTERESTS:
Custodians:
Air Force – 99
Army – GL1
Navy – AS

Reviewer:
DLA - CC

CIVIL AGENCY
COORDINATING ACTIVITY:
GSA - FSS

Preparing Activity:
Air Force - 82

Agent:
Air Force - 99

Project 6635-0180

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:12Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	**1. DOCUMENT NUMBER** A-A--59380	**2. DOCUMENT DATE (YYMMDD)** 981203
3. DOCUMENT TITLE **TENSIOMETER, WIRE ROPE**		
4. NATURE OF CHANGE *(Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)*		
5. REASON FOR RECOMMENDATION		
6. SUBMITTER		
a. NAME *(Last, First, Middle Initial)*	b. Organization	
c. ADDRESS *(Include zip code)*	**d. TELEPHONE** *(Include Area Code)* (1) Commercial (2) DSN *(if applicable)*	**7. DATE SUBMITTED (YYMMDD)**
8. PREPARING ACTIVITY		
a. SA-ALC/LDEE	**b. TELEPHONE** *(Include Area Code)* (1) Commercial 210/925-7153 (2) DSN 945-7153	
c. 485 Quentin Roosevelt Rd. KELLY AFB, TX 78241-6425	**IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS, CONTACT:** **Defense Quality and Standardization Office** **5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466** **Telephone (703) 756-2340 DSN 289-2340**	

DD Form 1426, OCT 89 *Previous editions are obsolete.* 198/290

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:12Z
Check the source to verify that this is the current version before use.