W-P-001251A (DSA-DM)
30 October 1973
SUPERSEDING
Revision of
Int. Fed. Spec W-P-001251
26 October 1966

INTERIM FEDERAL SPECIFICATION

POWER SUPPLY, CAUTERY APPARATUS

This interim Federal specification was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pennsylvania 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

1. SCOPE

1.1 This specification covers one type and two sizes of power supply, 200 watts and 300 watts, to supply required power for operation of various medical examining and surgical cautery uses.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

(Activities outside the Federal Government may obtain copies of Federal Specifications and Standards as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C., 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Services Administration Regional Offices in Boston, New York, Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, Seattle, and Washington, D.C.

(Federal Government activities may obtain copies of Federal Specifications and Standards and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

SPECIFICATIONS

Military

MIL-L-10547	Liners, Case and Sheet, Overwrap; Water Vaporproof or Waterproof, Flexible
MIL-M-7298	Manual, Technical: Commercial Equipment.
MIL-P-116	Preservation, Methods of

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

(Copies of specifications, standards, and drawings required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:34Z
Check the source to verify that this is the currentversion before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

C73.11 For Plugs and Receptacles, 125 Volts. 15 Amperes, 2-Pole, 3-Wire, grounding type.

(Applications for copies should be addressed to American National Standards Institute, 1430 Broadway, New York, N.Y. 10018.)

UNDERWRITERS' LABORATORIES, INC. (UL)

Standards for Safety

UL817 Cord Sets and Power Supply Cords.

(Copies of the above UL Standard are available from Underwriters' Laboratories, Inc., 207 East Ohio Street, Chicago, Illinois 60611, 333 Pfingsten Road, Northbrook, Ill., 60062, 1285 Walt Whitman Road, Melville, L.I., N.Y. 11746, 1655 Scott Boulevard, Santa Clara, Calif. 95050.)

NATIONAL FIRE PROTECTION ASSOCIATION (NFPA)

Standard NFPA 56 Code for the use of FLAMMABLE ANESTHETICS.

(Copies of the above are avilable from National Fire Protection Association, 60 Batterymarch Street, Boston, Mass., 02110.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit proof, prior to delivery, that the items to be delivered under the contract conform to the applicable requirements of UL 817. Proof shall consist of one of the following:

A. UL label or UL listing of the product, provided the contractor agrees to have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the Contracting Officer, upon the latter's request.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

B. Laboratory report from a qualified independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The qualification of such laboratory shall be determined by the Government which reserves the right to require a facilities report and to conduct a facilities survey. Any laboratory report submitted shall include actual test methods, data taken, and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches, and parts list, all appropriately identified. Compliance with this paragraph does not excuse compliance with all other requirements of the contract and specifications.

3.2 Materials. Materials shall be as specified herein, of the highest quality throughout, and suitable in all respects for the purpose intended.

3.3 Design and construction. The cautery apparatus power supply shall be a compact table model providing variable current suitable for rapid heating of various cautery electrodes and diagnostic lamps. It shall be arranged to permit the use of a diagnostic lamp and a cautery electrode either singly or simultaneously. The transformer and associated control components shall be enclosed in a sturdy housing, the top of which shall serve as the control panel on which shall be mounted the jacks, the switch and the control knobs. The jacks shall be placed adjacent to their respective controls.

3.3.1 Housing. The housing shall be fabricated from sheet steel or aluminum, and laminated thermosetting plastic. The sides and ends of the housing shall be formed from one continuous piece of sheet steel, not less than 0.031 inch thick (#22 United States Standard Gage), or not less than 0.039 inch thick aluminum. The joint shall be securely riveted or welded. The top and bottom of the housing shall be laminated thermosetting plastic, not less than 5/16 inch thick. The base shall be fitted with rubber pads at the four corners. The top shall be permanently marked with numbered or lettered dial graduations, "ON" and "OFF" positions of the line switch, and identifying words "LIGHT" and "CAUTERY" in close proximity to the appropriate controls and jacks.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

* 3.3.2 Output jacks. Jacks shall be brass, chromium plated and of the friction grip type sized to provide a secure connection with cord tip pins 0.075 to 0.130 inch diameter and carry the maximum rated current without overheating under all normal operating techniques. Two pairs of jacks shall be provided, one pair for electrocautery and one pair for diagnostic lamps.

* 3.3.3 Transformer. The transformer shall be one of the following sizes as indicated in Table I, and designed to deliver the rated current without overheating under all normal operating techniques:

Table I

Size	Rated Power (watts)	Minimum output	
		Cautery (amperes at 4.0 volts)	Lamp (volts at 0.5 amperes)
I	300	75	6
II	200	50	6

* 3.3.4 Controls. Controls shall include a control for the cautery output, a control for the lamp output and line switch. Lamp and cautery controls shall operate independently of each other, permitting either simultaneous or independent use of electrocautery and a diagnostic lamp. The controls shall provide smooth even adjustment of the respective currents. No adjustment shall be required for maintaining the selected level under continuous operation. The line switch shall be effective to both sides of the incoming power circuit.

* 3.3.5 Power cord. The apparatus shall be equipped with a permanently attached, flexible 3-conductor connecting cord and attachment plug. The cord shall be not less than 6 feet long. A strain relief bushing shall be provided where the cord enters the housing. The attachment plug shall be in accordance with ANSI C73.11 for 125 volt, 2-pole, 3-wire grounding type. An adapter with grounding pigtail shall be provided for use at installations equipped with 2-prong receptacles. The cord shall withstand a pull of 30 pounds when tested as specified. There shall be continuity between the ground prong of the attachment plug and the metal housing when tested as specified.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:34Z
Check the source to verify that this is the current version before use.

3.4 Dielectric strength. There shall be no electrical breakdown when the apparatus is subjected to the Dielectric test specified herein.

3.5 Identification marking. The housing shall be legibly and permanently marked in an inconspicuous location with the following information:

National Stock Number
Manufacturer's name or registered trademark
Contract or purchase order number
Electrical characteristics
U.S.

Marking secured by adhesives shall not be used.

3.6 Finish. Sheet steel and aluminum shall be thoroughly cleaned and primed with an appropriate primer. The unit shall then be given one or more coats of black enamel in accordance with best commercial practice. Hardware shall be plated or otherwise finished to resist corrosion.

3.7 Service Data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation, and maintenance and lists of component parts. The manual shall comply with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the Contracting Officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract, or purchase order. In addition, the Contractor shall furnish 15 copies of the approved technical manual to the Contracting Officer for distribution. Should the supplier have furnished the date of the contract and propose to furnish manuals identical with those previously accepted, then the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the Contracting Officer. Waiver of submission of drafts and submission of 15 additional copies shall be at the option of the Government.

3.8 Workmanship. Power supply unit shall be finished in a first class manner throughout and shall be free from defects or imperfections which may affect its appearance or serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

* 4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the power supply may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Classification of Inspections.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and Table II.

Source: https://assist.dla.mil -- Downloaded: 202640-26T11:34Z
Check the source to verify that this is the currentversion before use.

Table II

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	S-4	1.0
Total defects (Major and Minor combined)	S-4	4.0
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 for an Inspection level of S-1 and and AQL of 4.0.

4.3 Examination. The units shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

Table III. Categories and defects*

Major	
101	Any component defective, not present, not useable.
102	Controls not adjustable, not operative.
103	Output less than minimum.
104	Top marking not complete, not permanent.
105	Jacks not sized to grip cord tip pins.
Minor	
201	Identification marking not present, not legible, not permanently secured to housing.
202	Surfaces and edges not free of burrs, sharpness, nicks, scratches.
203	Power Supply does not rest evenly on flat level surface.
204	Service data not present, not as specified.

* Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Dielectric strength. The apparatus shall be subjected to a gradually increasing 60-cycle potential of 900 volts, AC, which shall be applied between the current carrying parts and the metal enclosure and maintained at the maximum value for a period of 1 minute.

4.4.2 Ground continuity. Electrical continuity shall be determined between the ground prong of the plug and metal enclosure using a suitable continuity meter or indicator.

4.4.3 Strain relief. The supply cord shall be tested in accordance with UL817.

4.4.4 Output. The power supply shall be connected to a suitable source and the output measured while operating for 5 minutes. The apparatus shall be observed for overheating.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C, as specified (see 6.2).

5.1.1 Unit of issue. One power supply, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV

Unit package	Intermediate package	Exterior container
1 unit	Not required	4 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:34Z
Check the source to verify that this is the currentversion before use.

5.1.3 Level A.

5.1.3.1 Cleaning and drying. Each unit shall be thoroughly cleaned and dried in accordance with the applicable method as outlined in MIL-P-116.

5.1.3.2 Preservative. Preservative shall not be applied.

5.1.3.3 Intimate wrap. Each unit shall be wrapped in neutral tissue or other suitable anticorrosive materials.

5.1.3.4 Unit package. Each unit shall be packaged in accordance with MIL-P-116, method 1C-2.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Fiberboard boxes shall conform to the special requirements specified in PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed by method II of PPP-B-636. Fiberboard boxes shall conform to special requirements specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use.

6.1.1 Intended use. This specification covers power supply units, for use with the following:

Stock No.	Description
6515-00-321-0600	Cable assembly
6515-00-321-0640	Cable assembly
6515-00-321-0680	Cable assembly
6515-00-318-5800	Cautery handle
6515-00-318-5900	Cautery handle

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

Stock No.	Description (continued)
6515-00-318-6100	Cautery handle
6515-00-323-7400	Cystoscope
6515-00-323-7800	Cystoscope
6515-00-323-8500	Cystoscope
6515-00-324-2200	Panendoscope
6515-00-359-8000	Resectoscope

6.1.2 Type II, class B power supply unit shall also be for use with Ophthalmoscope, NSN 6515-00-299-8326.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Size required (see 3.3.3).
(c) Selection of applicable levels of packing (see 5.1 and 5.2).

6.3 This specification covers the following items appearing in the Federal Supply Catalog.

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION	SIZE
6515-00-318-6800	Power Supply Cautery Apparatus, Large, 110 volt, 60 cycle, AC	I
6515-00-318-6840	Power Supply, Cautery Apparatus Medium, 110 volt, 60 cycle, AC	II

6.4 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet requirements.

6.6 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the last previous issue.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:34Z
Check the source to verify that this is the current version before use.

Custodians:

Army - MD
Navy - MS
Air Force - 03

Review activities:

Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1725

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 202610-26T11:34Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

S/N 0102-014-1802

Source: https://assist.dla.mil -- Downloaded: 2026l0-26T11:34Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Defense Personnel Support Center
2800 S. 20th St.
Philadelphia, PA 19101
Attn: ATTC-2

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:34Z
Check the source to verify that this is the current version before use.