MIL-T-21669 (WP)
Amendment-2
7 December 1964
Superseding
Amendment-1
21 January 1960

MILITARY SPECIFICATION

TEST SET, MARK 25 MOD 1

This amendment forms a part of Military Specification MIL-T-21669 (NOrd) of 12 November 1958, and has been approved by the Bureau of Naval Weapons, Department of the Navy.

By this amendment basic specification MIL-T-21669 (NOrd) is changed to read MIL-T-21669 (WP), and throughout this specification Bureau of Ordnance shall be changed to read Bureau of Naval Weapons.

* This amendment deletes requirements for the furnishing of data not considered essential and relocates essential data. No attempt has been made to update any other part of this specification.

* Page 3, paragraph 3.4: Relocate and designate 6.2.1.

* Page 5, paragraph 4.1.2, line 8: Delete ", together with a report of inspection,"

Page 6, paragraph 4.4.1.1: Delete sub-paragraphs (b), (c), (d), (e), and (f) and substitute:

"(b) One (1) variable decade precision resistor, with resistance ranges of 0-10 ohms, 0-100 ohms, 0-1,000 ohms, and 0-10,000 ohms, all at least plus or minus 1 percent, 1/2 watt.

(c) One (1) milliammeter, 0-30 milliampere range, accuracy within plus or minus 2 percent. "

Delete letter (g) and substitute (d)

Page 7, paragraph 4.5.3, line 5: Delete (g) (in item 4.4.1.1(g)) and substitute (d)

FSC 6625

Source: https://assist.dla.mil -- Downloaded: 20260-19T10:16Z
Check the source to verify that this is the current version before use.

Page 7, paragraph 4.5.4: Delete sub-paragraphs (d), (e), and (f) and substitute:

"(d) Connect the decade precision resistor (item 4.4.1.1 (b)) across the Current-Ohms binding posts with the 1,000 ohm range set on 1,000 ohms. Adjust the resistor until test set meter reads exactly 1,000 ohms. Decade resistor shall indicate 1,000 plus or minus 20 ohms.

(e) Repeat step (d) using settings of 4,000 ohms and 10,000 ohms. In each case decade resistor shall indicate 4,000 plus or minus 80 ohms and 10,000 plus or minus 200 ohms respectively.

(f) Remove resistor. "

Pages 8 and 9, paragraph 4.5.5: Delete sub-paragraphs (a) through (k), and substitute:

"(a) Set 3,000 ohms on the precision decade resistor (item 4.4.1.1 (b)).

(b) Set test set selector at 1 MA.

(c) Observing polarity, connect 1.5 volt battery (item 4.4.1.1 (a)), decade resistor (item 4.4.1.1 (b)) and external milliammeter (item 4.4.1.1 (c) 30 MA full scale) in series across the Current-Ohms binding posts.

(d) Vary decade resistor to obtain half-scale and full-scale deflection on the test set milliammeter and read external milliammeter on each setting. Test set meter shall indicate same as external milliammeter within plus or minus 2 percent in each instance.

(e) Repeat step (d) for 3 MA, 10 MA, and 30 MA ranges. Test set meter shall indicate same as external milliammeter within plus or minus 2 percent in each instance.

(f) Remove milliammeter, decade resistor and battery."

* Page 12: Insert new paragraph after 6.1.1:

"6.2 Data - For the information of Contractors and Contracting Officers, any of the data specified in (a) subparagraphs below, (b) applicable documents listed in Section 2 of this specification, or (c) referenced lower-tier documents need not be prepared for the Government and shall not be furnished to the Government unless specified in the contract or order. The data to be furnished shall be listed on DD Form 1423 (Contractor Data Requirements List), which shall be attached to and made a part of the contract or order. NavWeps Form 4200/15 (Drawings Lists and Specifications Required) shall be attached where applicable."

* Items marked with asterisk added by this amendment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-19T10:16Z
Check the source to verify that this is the current version before use.