INCH-POUND

A-A-30190
January 8, 1993

COMMERCIAL ITEM DESCRIPTION

PAD, PETROLATUM AND BISMUTH TRIBROMOPHENATE

The General Services Administration has authorized the use of this commercial description.

This commercial item description describes a petrolatum and bismuth tribromophenate impregnated pad suitable for medical and surgical use.

Salient characteristics. Pad, Petrolatum and Bismuth Tribromophenate Impregnated, shall be a sterile 3 percent Bismuth Tribromophenate and Petrolatum blend impregnated gauze dressing, suitable for medical and surgical use.

The gauze shall comply with the requirements of the USP, Type I.

The impregnated gauze shall be subjected to a suitable sterilization process which shall render the gauze sterile. The gauze shall contain not less than 2.5 percent and not more than 4.3 percent bismuth tribromophenate in a petrolatum blend when determined by a suitable, accurate and reproducible method.

The blended mixture of bismuth tribromophenate in petrolatum shall be evenly and uniformly distributed throughout the gauze. The mixture of bismuth tribromophenate and petrolatum shall have a melting range of not less than 100°F (38°C) and not more than 140°F (60°C) and a consistency of not less than 150 and not more than 240 when tested using USP methods in the monograph for White Petrolatum, USP.

All ingredients used in the manufacture of the product shall be in accordance with the tests, standards and requirements of the USP/NF including any supplements or revisions thereto, or if not included in this compendium, they shall be of the highest pharmaceutical grade suitable for use in the manufacture of the product.

Testing and examination shall be conducted to determine compliance with all requirements of this document.

FSC 6510

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Each impregnated gauze shall be supplied in an individually wrapped immediate container that is overwrapped and sterilized in a suitable package capable of maintaining sterility of contents unless damaged or opened.

Not more than 6 months shall be elapsed from the date of manufacture to the date of delivery to the Government.

Size 1 - 5 inches wide by 9 inches long, (12.7 cm wide by 22.9 cm long). The tolerance for the width is plus or minus 1/8 inch (.32 cm) and plus or minus 1/2 inch for length (1.27 cm).

Size 2 - 1 inch wide by 8 inches long (2.54 cm wide by 20.3 cm long). The tolerance for the width is plus or minus 1/8 inch (.32 cm) and plus or minus 1/2 inch for length (1.27 cm).

Workmanship. Product shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 72 individually packaged pads, as specified, constitutes one unit for NSN 6510-01-120-2642 and 6510-01-115-4729. Unit packaging for VA NSN 6510-00-900-6187 shall appear in the solicitation.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract of purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the government. The government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with this commercial item description requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of S-1.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description, and conforms to the producer's own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling and marking. Unless otherwise specified, preservation, packaging and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE. The following National Stock Number is covered by this document:

NATIONAL STOCK NUMBER	NOMENCLATURE	SIZE
6510-00-900-6187	PAD, PETROLATUM & BISMUTH TRIBROMOPHENATE	1
6510-01-120-2642	PAD, PETROLATUM & BISMUTH TRIBROMOPHENATE, 72s	1
6510-01-115-4729	PAD, PETROLATUM & BISMUTH TRIBROMOPHENATE, 72s	2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.

Copies of Federal Standard No. 376, Preferred Metric Units for General Use by the Federal Government and MIL-STD-105, Sampling Procedures and Tables for Inspection by Attributes may be obtained from the Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20420.

Copies of Federal Test Method Standard 191 may be obtained from your local General Services Administration.

Ordering data. Purchasers should select the preferred options permitted herein, and include the following information in procurement documents:

(a) Title, number, and date of this commercial item description
(b) Size

MILITARY INTERESTS:

Military Coordinating Activity:

DoD-MB

AGENT:

DLA-DM

Custodians:

Army-MD
Navy-MS
Air Force-03

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
PHS

PREPARING ACTIVITY:

VA-OSS VANAC (904)
Agent VA-OSS VACO (92A)

Project No. 6510-0848

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T08:37Z
Check the source to verify that this is the current version before use.