DATA ITEM DESCRIPTION
Title: RADIOACTIVE SOURCE CERTIFICATE OF CALIBRATION

Number: DI-ENVR-82387
AMSC Number: N10339
DTIC Applicable: N/A
Preparing Activity: SH
Applicable Forms: N/A

Approval Date: 20230130
Limitation: N/A
GIDEP Applicable: N/A
Project Number: ENVR-2018-003

Use/relationship: The Radioactive Source Certificate of Calibration will be used to provide the government with the certification use and disposition of radioactive sources.

This Data Item Description (DID) contains the format and content preparation instructions for the data product generated by the specific and discrete task requirement as delineated in the contract.

Requirements:

1. Reference Documents. N/A

2. Format. The Radioactive Source Certificate of Calibration shall be in contractor format.

3. Content. The Radioactive Source Certificate of Calibration shall contain the following:

    a. A certification statement by the manufacturer that certifies the radioactive source is National Institute of Standards and Technology (NIST) traceable.
    b. A certification statement by the manufacturer that certifies the source contains no hazardous material that would affect its disposal as waste, other than its radioactive content.
    c. The activity of the source
    d. The activity of each of the nuclides in the source
    e. The uncertainty value, reference time and date
    f. The activity of any gamma emitting impurities detected (expressed as a percentage of the activity of the main nuclide as reference time and date)
    g. Density and volume of active volume
    h. The vendor assigned source serial number
    i. International Organization for Standardization (ISO) Classification
    j. The results and date of leakage testing

End of DI-ENVR-82387

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T11:03Z
Check the source to verify that this is the current version before use.