NNN-B-300A
August 19, 1977
SUPERSEDING
Int. Fed. Spec. NNN-B-00300(DSA-DM)
December 22, 1964

FEDERAL SPECIFICATION

BLOOD COLLECTING-CULTURE UNIT, LABORATORY

This specification was approved by the Commissioner, Federal Supply Service, General Services Administration for the use of all Federal agencies.

1. SCOPE

1.1 Scope. This specification covers one type of blood collecting-culture unit, suitable for collecting and culturing blood and other body fluids aerobically and anaerobically.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

Federal Specifications:

GG-I-526 - Instruments, Dental and Surgical; General Specifications.
PPP-B-585 - Boxes, Wood, Wirebound.
PPP-B-601 - Boxes, Wood, Cleated-Plywood.
PPP-B-621 - Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636 - Boxes, Shipping, Fiberboard.

Federal Standards:

Fed. Std. No. 123 - Marking for Shipment (Civil Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, US Government Printing Office, Washington, DC 20402.

FSC 6630

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

(Single copies of this specification and other Federal Specification required by activities outside the Federal Government for bidding purposes are available without charge from Business Service Centers at the General Services Administration Regional Offices in Boston, New York, Washington, DC, Atlanta, Chicago, Kansas City, MO, Fort Worth, Denver, San Francisco, Los Angeles, and Seattle, WA.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

Military Specifications:

MIL-L-10547 - Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.
MIL-N-36124 - Neddles, Hypodermic, Disposable.

Military Standards: Sa

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

(Copies Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

U.S. Department of Health, Education and Welfare, Food and Drug Administration(FDA):

Federal Food, Drug, and Cosmetic Act and Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U.S. Pharmacopeial Convention, Inc.:

The United States Pharmacopeia (USP).

(Application for copies should be addressed to the Macking Publishing Company Easton, PA 18042.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Design and construction. Each blood-collecting culture unit shall be an assembly consisting of the following components:

3.2.1 An evacuated culture bottle containing 50 ± 3 milliliters (ml) of soybean-casein digest medium under a partial atmosphere of carbon dioxide, with sufficient vacuum to permit taking not less than 30 ml of blood. Bottle shall be rubber stoppered and capped.

3.2.2 A sterile bleeding set consisting of a disposable 20-gage 1-1/2 ± 1/16 inch blood collecting needle and a 20-gage, 1 ± 1/16 inch stopper-puncturing needle connected by a 12-1/4 ± 1/4 inch length of polyvinyl chloride tubing.

3.2.3 Bottles shall be fabricated of good quality flint or soda-lime glass Bottles shall maintain the integrity of the medium and permit both aerobic and anaerobic cultivation for culturing blood and other body fluids.

3.2.3.1 Bottle. Each bottle shall be graduated and commercially labeled in accordance with FDA regulations for In Vitro Diagnostic Products. In addition, the expiration date shall be shown. This information may be shown on pressure sensitive label.

3.2.4 Needles.

3.2.4.1 Cannula. The cannula shall be fabricated from 20-gage (outside diameter 0.035 ± 0.001 inch, inside diameter 0.022 inch minimum) corrosion-resisting steel tubing, conforming to AISI 304 of MIL-N-36124. Tubing shall be seamless drawn or welded and drawn.

3 2.4.2 Hub. The hub shall be fabricated from brass or aluminum. Brass shall be nickel or chromium plated. The hub shall be so constructed that the cannula goes through, and is firmly secured along, the central longitudinal axis of the hub. Hub shall be secured to the cannula by swaging. One end of the hub shall be of such size that it will form a tight leakproof connection with the plastic tubing. When subjected to the test specified in 4.4.1, the tubing shall not become detached from its respective hub.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

3.2.4.3 Blood-collecting needle. The cannula of the blood-collecting needle shall be a straight tube with a point at one end and a hub at the other. The venipuncture end shall have a regular bevel point, shall be ground concave or flat, and may have a center ridge at the distal end to strengthen the tip. The venipuncture point shall be so ground at the outer end as to enhance the keeness of the tip and cutting edge. The tip of the venipuncture point shall be straight and strong and shall be of maximum keenness.

3.2.4.3.1 Guard. The sterile blood-collecting needle shall be protected by a polyvinyl chloride plastic sealed guard.

3.2.4.4 Stopper-puncturing needle. The cannula of the stopper-puncturing needle shall be a straight tube with a point at one end and a hub at the other. The end for the stopper-puncture shall have a sharp, short, beveled point with a flat grind suitable for puncturing through the rubber stopper.

3.2.4.4.1 Guard. The sterile stopper-puncturing needle shall be protected with a polyvinyl chloride plastic needle guard containing a bacteria-retentive filter.

3.2.4.5 Adherence. The cannula shall not separate from the hub, when tested as specified in 4.4.4.1.

3.2.4.6 Resistance to breakage. There shall be no breakage, when the needle is tested as specified in 4.4.4.2.

3.2.4.7 Stiffness. The maximum deflection of the cannula shall not exceed 0.021 inch, when tested as specified in 4.4.4.3.

3.2.4.8 Corrosion resistance. The cannula shall meet the requirements for resistance to boiling specified in GG-I-526, when tested as specified in 4.4.4.4.

3.2.5 Tubing. The tubing of the blood collecting-culture unit shall be fabricated from transparent flexible polyvinyl chloride plastic. The tubing shall be sufficiently rigid to prevent collapse during normal use.

3.2.6 Rubber stopper. Each bottle shall be stoppered with a rubber internal-sealing stopper. Rubber shall be nonblooming, nonoxidizable, non-bleeding, and self-sealing diaphragm type capable of retaining the required vacuum.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

3.2.7 Protective cap. Each bottle shall be protected with a black plastic cap.

3.3 Performance.

3.3.1 Vacuum. Vacuum in the bottle at the time of shipment shall be such that a volume of distilled water shall be drawn into the bottle above the medium to not less than the maximum graduation, when the bottle is tested as specified in 4.4.2.

3.3.1.1 Vacuum guarantee. The bottles shall be properly evacuated and shall be guaranteed to draw the quantity of distilled water specified in 3.3.1 for a period of 18 months from the date of manufacture.

3.3.2 Soybean-casein digest medium.

3.3.2.1 Appearance. The color of the broth shall be light to medium yellow and clear in appearance, when examined as specified in 4.4.5.1.

3.3.2.2 pH requirement. The pH reading of the sterile broth shall be 7.3 $\pm$ 0.2, when tested as specified in 4.4.5.2.

3.3.2.3 Growth of microorganisms. The soybean-casein digest medium shall be a replicable culture medium for the promotion of growth of aerobic and anaerobic microorganisms when tested as specified in 4.4.5.3.

3.4 Blood-collecting set. The sterile blood-collecting set shall be packaged in a sealed envelope with appropriate identification on the envelope.

3.5 Instruction sheet. Each individual package shall contain a detailed instruction sheet for taking specimens. When an alternate unit package as described in 5.1.1.2 is used, each unit package shall contain a detailed instruction sheet.

3.6 Sterility. Blood collecting culture assembly shall comply with sterility test requirements specified in 4.4.3.

3.6.1 Certificate of sterility. A certificate of sterility, for the bleeding sets and bottles, stating that samples from each sterilizer charge have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item Identification.
b. National Stock Number.
c. Contract or purchase order number.
d. Sterility lot or control number.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:48Z
Check the source to verify that this is the current version before use.

e. Method of sterilization.
f. Date of sterilization.
g. The biological indicator used with test samples
h. Date sterility test was started.
i. Specific test results.
j. Statement that samples comply (or do not comply) with sterility tests.
k. Signature of head of laboratory or a designated qualified alternate.
l. Quantity in sterilizer load.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, Defense Personnel Support Center, ATTN: DPSC-ATQ, 2800 South 20th Street, Phila., PA 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance representative for submission to the procuring activity.

3.6.2 Lot size. A sterility lot size shall consist of a manufacturer's batch subjected to the same sterilizer load. Each component package shall be clearly marked with a lot number as specified in 5.3.2.1 which shall be traceable to the pertinent batch and date of sterilization.

3.7 Workmanship. Workmanship shall be first class throughout. Blood collecting-culture unit shall be clean, and shall be free from defects which detract from its appearance or may impair its serviceability.

3.8 Expiration dating period. The expirating dating period shall be 18 months.

3.9 Delivery Not less than 16 months of the expiration dating period shall remain at the time of delivery to the Government.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:48Z
Check the source to verify that this is the currentversion before use.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemcial properties) of the item.

4.1.2 Certificate of quality. Certificate of quality, supplied by the manufacturers of the corrosion-resisting steel and plastic, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government The Certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and, (i) as to any expiration dated item, for the expiration dating period specified by the contractor for such item, or such longer period as may be required by regulation of any federal agency; or (ii) as to non-dated items, for not less than 3 years after delivery of the item to the Government.

4.2 Sampling

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105.

TABLE I. Sampling for examination

	Inspection level	AQL (percent defective)
For visual examination		
Major A defects	S-3	1.0
Major B defects	S-3	1.0
Minor A defects	S-3	2.5
Minor B defects	S-3	2.5
For dimensional examination	S-2	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

Source: https://assist.dla.mil -- Downloaded: 202640-26T12:48Z
Check the source to verify that this is the currentversion before use.

TABLE II. Sampling for tests

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL (percent defective)
Set	Tubing to needle hub	3.2.2.2	4.4.1	S-3	1.5
	Vacuum	3.3.1	4.4.2	S-2	1.5
	Sterilization	3.6	4.4.3	1/	2/
Needle	Adherence	3.2.2.5	4.4.4.1	S-1	2/
	Resistance to breakage	3.2.2.6	4.4.4.2	S-1	2/
	Stiffness	3.2.2.7	4.4.4.3	S-1	2/
	Boil test	3.2.2.8	4.4.4.4	S-1	2/
Soybean casein digest medium	Appearance	3.3.2.1	4.4.5.1	S-1	2/
	pH requirement	3.3.2.2	4.4.5.2	S-1	2/
	Growth of micro-organisms	3.3.2.3	4.4.5.3	S-1	2/

1/ Sampling shall be as specified in 4.4.3.
2/ Acceptance number shall be zero.

4.2.2.1 Where possible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The blood collecting-culture unit shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following classification of defects:

TABLE III. Classification of defects

Categories	Defects 1/
Major A (7X magnification)	
(Needle)	
101	Point not sharp.
102	Point edges not free of feathering.
103	Point not free of hooks and bends
104	Point not symmetrical.
105	Lumen not free of constriction.
106	Cannula not straight (deviation from longitudinal axis exceeds (3°).
107	Cannula not free of cracks.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

TABLE III. Classification of defects (Continued)

Categories	Defects 1/
108	Cannula not free of scale and corrosion.
109	Cannula not free of slivers.
110	Cannula not free of foreign matter.
111	Cannula not free of burs.
Major B	
(Needle)	
151	Threading not free of breaks.
152	Hub not free of distortion.
153	Hub not free of cracks.
(Set)	
154	Bottle not free of cracks.
155	Stopper not secured in bottle.
156	Blood-collecting bottle does not contain culture medium.
157	Tubing not flexible.
158	Tubing is tacky or sticky.
Minor A (7X magnification)	
(Needle)	
201	Cannula not free of pits.
202	Cannula not free of tool marks.
203	Hub not free of foreign matter.
Minor B	
(Needle)	
231	Hub not free of fins.
(Set)	
233	Bottle not free of bubbles.
234	Stopper not free of bloom.

1/ Examination shall not be restricted to the classified possible defects listed above.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:48Z
Check the source to verify that this is the currentversion before use.

4.3.2 Dimensional examination. The blood collecting-culture unit shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.3.3 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing and marking comply with the requirements of section 5. Defects shall be scored as specified in table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 expressed in terms of percent defective.

TABLE IV. Examination of preparation for delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Packaging and packing materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence or method of application.

4.3.3.1 Examination of closure, waterproofing and banding of containers. Each shipping container is required to comply with PPP-B-585, PPP-B-601, or PPP-B-636. Examination for defects in closure, waterproofing and banding shall be in accordance with the appendix to the applicable box specification.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Tubing to needle hub. The cannula of the needle shall be attached in a clamping device. The tubing shall be suspended, hanging vertically downward, and a 5-pound weight applied gradually shall be suspended from the tubing for 5 seconds.

4.4 2 Vacuum. The blood collecting needle shall be immersed in water The rubber stopper shall be punctured with the stopper-puncturing needle. The water drawn into the bottle shall fill the bottle to the maximum graduation mark indicated on the bottle.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

4.4.3 Sterility. Components of the blood-collecting culture assembly culture bottles 3.2 (a) and bleeding sets 3.2(b) shall be tested for sterility in accordance with the method specified in the USP "Procedural Details for Sterility Tests for Class B, Type I or Type II, Inoculated Products or Inoculated Carrier for Parenteral or Non-Parenteral Devices "

4.4.4 Needles.

4.4.4.1 Adherence. A straight pull of 15 pounds shall be applied between the cannula and the hub for a period of 1 minute.

4.4.4.2 Resistance to breakage. The needle hub shall be held in a fixed position, and a force shall be applied to the distal end of the cannula, 1 inch from the hub, in order to bend the cannula in a clockwise direction through an arc of 25^{0} from the original or neutral position. The operation shall then be reversed (i.e., bent through an arc of 25^{0} from the neutral position in a counterclockwise direction) and finally, the cannula shall be returned to the neutral position. This complete operation (or cycle) shall be repeated 20 times.

4.4.4.3 Stiffness. The cannula shall be placed across two test edges (0.002-inch radius maximum) 0.750 inch $\pm$ 0.003 inch apart. A load of 50 ounces $\pm$ 1/4 ounce shall be applied to the center of the span, and the resulting maximum deflection of the cannula shall be measured using a gage whose minimum graduations are not greater than 0.001 inch. The force shall be applied through an edge with a maximum radius of 0.036 inch, and shall not be farther than 0.003 inch from the center.

4.4.4.4 Boil test. Boil test shall be performed as specified in GG-I-526.

4.4.5 Soybean-casein digest medium. Samples of the sterile soybean-casein digest medium shall be compared with a lot previously tested and found satisfactory.

4.4.5.1 Appearance. The broth shall be visually examined for color, clarity, and general appearance.

4.4.5.2 Determination of pH. The sterile broth shall be tested for pH at 70^{0} F $\pm$ 2^{0} F.

4.4.5.3 Growth of microorganisms. The soybean-casein digest medium shall be tested for its ability to support early and rapid growth of aerobic and anaerobic microorganisms. After incubation for 7 days, the broth samples shall be comparable to that of the control material and the total amount of growth occurring in all samples shall be at least 90-percent of that of the control.

Source: https://assist.dla.mil -- Downloaded: 202610-26T12:48Z
Check the source to verify that this is the current version before use.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Individual package. Each individual package shall contain one blood-collecting culture bottle of medium, one bleeding set (see 3.4) and instruction sheet packaged in a suitable container. No individual package need be used if alternate as described in 5.1.1.2 is used.

5.1.1.2 Unit package. Ten (10) complete units shall be further packaged in a box of appropriate size. As an alternate, the ten (10) blood-collecting culture bottles of medium shall be packaged in a corrugated box containing twelve (12) partitioning cells. The ten (10) envelopes containing the sterile bleeding sets and instruction sheet shall be placed in the two center cells. Closure of the unit package shall be adequate to prevent accidental opening.

5.1.2 Level C. Units shall be packaged in standard commerical containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.1).

5.2.1 Level A. Eight (8) unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed as specified in the appendix of PPP-B-636.

5.2.2 Level B. Eight (8) unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Civil agencies. Marking for shipment shall be in accordance with Fed. Std. No. 123. In addition each unit and shipping container shall be marked: "STORE AT CONTROLLED ROOM TEMPERATURE 59° - 86° F".

5.3.2 Military activities.

5.3.2.1 Individual package. Each individual package shall be commercially identified. Marking shall include the lot or control number.

5.3.2.2 Unit packages. Each unit packages shall be marked in accordance with MIL-STD-129. Lot (control) number, contract or purchase order number, and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-6/76G". Final sterilization shall be on or after indicated packaging date. In addition, Type I Shelf-Life markings shall be shown as specified in MIL-STD-129. Marking shall also include the legend: "STORE AT CONTROLLED ROOM TEMPERATURE 15° - 30° C (59° - 86° F)".

5.3.2.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number and Type I Shelf-Life markings shall be shown as specified in MIL-STD-129. The method of sterilization (G-Gas, R-Radiation, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-6/76G". Final sterilization shall be on or after indicated packaging date. In addition, marking shall include the following legend: "STORE AT CONTROLLED ROOM TEMPERATURE 15° - 30° C (59° - 86° F)".

6. NOTES

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in procurement documents:

(a) Title, number and date of this specification.
(b) The National Stock Number.
(c) Selection of applicable levels of packaging and packing (see 5.2 and 5.3).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

6.2 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6630-00-985-5315	BLOOD COLLECTING-CULTURE UNIT, LABORATORY, 10s

MILITARY CUSTODIANS:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITES:
GSA-FSS
HEW-FDA, NIH
VA-DMS

Project No. 6630-0275

Orders for this publication are to be placed with General Services Administration, acting as an agent for the Superintendent of Documents. See Section 2 of this specification to obtain extra copies and other documents referenced herein. Price 35 cents each

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.