NOT MEASUREMENT SENSITIVE

MIL-DTL-2/18D
w/Amendment 1
29 September 2006
SUPERSEDING
MIL-DTL-2/18D
25 September 2002

DETAIL SPECIFICATION SHEET

VALVE, CYLINDER, GAS: CARBON MONOXIDE, OUTLET 350 (PRESSURES THROUGH 3000 psig (20680 kPa) AT 120 ºF (48.9 ºC))

This specification is approved for use by all Departments and Agencies of the Department of Defense.

The requirements for acquiring the product described herein shall consist of this specification sheet, MIL-DTL-2 and QPL-2.

REQUIREMENTS:

TYPE DESIGNATION: V18-350-2

TABLE I. Valve requirements.

Style	Class	Composition	Inlet size	Outlet application	Handwheel	PRD[1]
IIB	30	A	6	350	Yes	5
III	30	A	6	350	Yes	5

[1] Pressure Relief Device

SPECIAL REQUIREMENTS:

1. Mark the valve body: CARBON MONOXIDE.
2. A valve outlet cap shall be provided.

INTENDED USE: The valves covered by this specification sheet are intended for use with carbon monoxide cylinders.

NOTES:

1. Special requirements will be mandatory for all valve procurements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:37Z
Check the source to verify that this is the current version before use.

AMENDMENT NOTATIONS: The margins of this specification are marked with vertical lines to indicate where modifications from this amendment were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations.

Custodians:
Army - AT
Navy - SH
Air Force - 68

Preparing Activity:
DLA - GS7

(Project 8120-2006-010)

Review Activities:
Army - AV
Navy - AS, MC

NOTE: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST database at http://assist.daps.dla.mil/.

Source: https://assist.dla.mil -- Downloaded: 2026-10-15T15:37Z
Check the source to verify that this is the current version before use.