Not Measurement Sensitive
1 December 1993
MIL-C-49513(TM)

MILITARY SPECIFICATION
CATALOGS, SUPPLY: MEDICAL SETS, KITS AND OUTFITS

This specification is approved for use by Department of the Army and is available for use by medical departments and agencies of the Department of Defense.

1. SCOPE.

1.1 **Scope.** This specification contains the requirements for preparation of Medical Supply Catalogs (SC), including illustrated components list and accountability aid for medical sets, kits, and outfits (SKO).

1.2 **Applicability.** This specification is applicable only to those Medical SKOs that are type classified and are assigned a Line Item Number (LIN) and National Stock Number (NSN).

1.3 **Examples/figures.** The figures used in this specification are examples only. The text of this document takes precedence over the figures.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commander, US Army Medical Materiel Agency, ATTN: SGMMA-OCU, Frederick, MD 21702-5001 by letter or by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document.

AMSC NO A6774 FSC 6545

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

2. APPLICABLE DOCUMENTS

2.1 Government Documents.

2.1.1 Specifications and standards. The following specifications form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents are those listed in the issue of the Department of Defense Index of Specifications and Standards (DODISS) and supplement thereto, cited in the solicitation.

SPECIFICATIONS

Military

MIL-M-38784	Manuals, Technical: General Style and Format Requirements
MIL-M-85337	Manuals, Technical: Quality Assurance Program; Requirements for

STANDARDS

MIL-STD-129	Marking for Shipping and Storage

(Unless otherwise indicated, copies of federal and military specifications, standards, and handbooks are available from the Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia, PA 19111-5094.)

2.1.2 Other government documents, drawings, and publications.

The following other government documents, drawings, and publications form a part of this document to the extent specified herein. Unless otherwise specified, the issues are those cited in the solicitation.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

HANDBOOK

CATALOG HANDBOOK H4/H8 (SB708-43)	Section A - Commercial and Government Entity Code - Name of Code Section B - Commercial and Government Entity Code - Code to Name
FEDERAL SUPPLY CATALOGING HANDBOOK H6	Federal Item Name Directory for the Federal Cataloging Catalog System

(Application for copies of cataloging handbooks should be addressed to Commander, Defense Logistics Information System, ATTN: DLIS-WXA, Federal Center, Battle Creek, MI 49017-3084.)

2.2 **Non-government publications**. The following document(s) form a part of this document to the extent specified herein. Unless otherwise specified, the issues of the documents which are DOD adopted are those listed in the issues of the DODISS cited in the solicitation. Unless otherwise specified, the issues of documents not listed in the DODISS are the issues of the documents cited in the solicitation (see 6.2).

ASTM D3951 (DOD adopted)	Standard Practice for Commercial Packaging

(DOD activities can obtain copies of ASTM D3951 from Standardization Documents Order Desk, Building 4D, 700 Robbins Avenue, Philadelphia, PA 19111-5094. Other government activities, contractors and other private concerns must procure copies from the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103).

2.3 **Order of precedence**. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification takes precedence. Nothing in this specification, however, supersedes applicable laws and regulations unless a specific exemption has been obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 **General**. Unless otherwise specified, style and format shall be in accordance with MIL-M-38784. The data elements provided in the SC that are contained in the Army Master Data File (AMDF) and specific unique codes within the Medical Logistics System shall be in accordance with the latest issue, and the date of that issue shall be reflected below the title of Section I.

3.2 **Content**. Medical supply catalogs for SKOs shall consist of:

a. Cover.

b. Section I, Introduction.

c. Section II, Illustrated Component List/Hand Receipt.

d. Section III, Illustrated Associated Support Items of Equipment (ASIOE).

e. Section IV, American Hospital Formulary Service Therapeutic Classification.

f. Section V, NSN/Illustration and Page Index.

g. Section VI, NSNs Deleted From Previous Publication and Applicable Replacement. (Note: This section is for other than first issue catalogs).

h. Concluding material.

3.2.1 **Cover**. The cover shall consist of the applicable SC publication number; the title "DEPARTMENT OF THE ARMY SUPPLY CATALOG"; the designator "SETS, KITS, AND OUTFITS"; the standard item nomenclature; NSN for the respective item; the line item number (LIN); when applicable, the supersession notice; the authority line "HEADQUARTERS, DEPARTMENT OF THE ARMY"; the date of the publication (day, month, year); the applicable distribution statement (see Figure 1 for an example). The cover shall be salmon color on cover stock not less than 80 pound type.

3.2.2 **Section I, Introduction**. The introduction shall be prepared in accordance with figures 2 through 16 and shall include the following paragraphs:

Source: https://assist.dla.mil -- Downloaded: 202610-26T04:01Z
Check the source to verify that this is the current version before use.

a. General
b. Purpose
c. Responsibility
d. Explanation and use of columns (See Section II)
e. Abbreviations
f. Special Information
g. Accountability
h. Transportation information
i. Power requirements
j. Reporting errors and recommending improvements

3.2.2.1 **General**. Identifies the title of the specific SC being written and provides an explanation of the numeric/alpha designation assigned to the SC. It will state the reason for publication and contain the title of the document authorizing the SC.

3.2.2.2 **Purpose**. States the purpose and intended use of the SC. It shall contain the statement "Items listed in Section II constitute the minimum basic load to sustain this set/kit/outfit."

3.2.2.3 **Responsibility**. Identifies designated responsible proponents for all aspects of management of this SKO, to include referencing applicable Army regulations and other government documents.

3.2.2.4 **Explanation and use of columns (See Section II)**. This paragraph shall provide a description of the contents and layout of data in Section II.

3.2.2.5 **Abbreviations**. This paragraph shall be a listing of all nonstandard abbreviations that are not prescribed in AR 310-50 (Authorized Abbreviations and Brevity Codes) or Army Acronymns as defined in AR 40-61 and DOD Medical Catalog introduction (Authorized Abbreviations Referenced to Terms). Some words will be truncated (with vowels removed that do not meet the above references) in order to fit a limited-space nomenclature field.

3.2.2.6 **Special Information**. Defines and/or explains additional information deemed beneficial to the user, which logically does not belong in other specified paragraphs. It includes the paragraphs listed below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

3.2.2.6.1 **Associated Support Items of Equipment (ASIOE)**. Explains the need for this section and defines the parimeters for its use. ASIOE will only be listed in Section III of the SC for information. Items will be authorized separately.

3.2.2.6.2 **Therapeutic Classification Index**. Explains the need for this section and defines the parimeters for its use. It will identify any supporting authorization documents. These items will be listed in Section IV of the SC.

3.2.2.6.3 **Remarks Code**. This paragraph shall explain all applicable remarks codes.

3.2.2.6.4 **Component Listings**. Explains the availability of the component listing publication formats, both automated and hard copy. It will identify USAMMA (with address) as the proponent for distribution and resupply of the SC.

3.2.2.7 **Accountability**. Identifies procedures used for completing and maintaining Section II of the SC.

3.2.2.8 **Transportation Information**. Includes any appropriate information on packaging and handling as well as dimensions of the item, such as length, width, height, weight and cubic measurement of SKO.

3.2.2.9 **Power Requirements**. Contains a list of required electrical, pneumatic/hydraulic, or portable power requirements associated with the SKO. It will be identified by NSN, nomenclature, required voltage and wattage.

3.2.2.10 **Reporting of Errors and Recommending Improvements**. The following reporting of errors and recommending improvements statements shall be included in Section I.

"a. You can help improve this supply catalog. If you find any mistakes or if you know of a way to improve the information, please let us know. Mail your letter or DA Form 2028, (Recommended Changes to Publications and Blank Forms), direct to: Commander, US Army Medical Materiel Agency, ATTN: SGMMA-OCU, Frederick, MD 21702-5001. A reply will be furnished to you."

"b. If you have recommendations to improve the component makeup, satisfactory performance, quality or quantity of component authorization, submit SF 368, Quality Product Deficiency Report,

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

direct to: Commander, US Army Medical Materiel Agency, ATTN: SGMMA-OCU, Frederick, MD 21702-5001. A reply will be furnished to you."

3.2.3 **Section II, Illustrated Component List/Hand Receipt.** (Header Page Only). This section shall consist of component illustrations, item descriptions, quantity on hand, accountability blocks, and signature blocks (on the last page). Content and format shall be in accordance with the following paragraph (see Figure 17 through 19 for an example.) Shall contain the statement also "For use of this form, see Section I. The proponent agency is ODCSLOG."

3.2.3.1 **"FROM" Block (Header Page Only).** Identifies the element from which the property is issued. Enter the name of the organization, unit, section, or squad which issues the property.

3.2.3.2 **"TO" Block (Header Page Only).** Identifies the element to which the property is issued. Enter the name and rank of the person receiving the property. TOE/MTOE paragraph, line number, and job title may be added as optional information.

3.2.3.3 **"Hand Receipt Number" (Header Page Only).** Enter a locally assigned number.

3.2.3.4 **"End Item Stock Number" Block.** Enter the NSN of the SKO.

3.2.3.5 **"End Item Description" Block.** Enter a short description of the SKO as specified in the Federal Supply Cataloging Handbook H6.

3.2.3.6 **"Pub Number" Block.** The catalog shall contain a four-segment publication number. The first segment shall be "SC" to identify the publication type as a Supply Catalog. The second segment shall be the Federal Supply Classification (FSC), i.e., "6545". The third segment shall be a one-digit compiler code. For medical supply catalogs, this number will be "8".

3.2.3.7 **"Pub Date" Block.** The pub date shall be the publication date of the SC.

3.2.3.8 **"ACT" (Action Code) Column.** This column shall contain a one-character symbol indicating the type of publication action taken since the previous catalog. The symbols shall be defined as follows:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

C (Change) - Indicates a change to data other than the NSN since the previous publication.

N (New) - Indicates the NSN was not previously included in the SKO.

R(Replacement)- Indicates the NSN has been replaced or changed.

3.2.3.9 **"Stock Number" Column.** Each component of the host SKO and subset shall be identified by its National Stock Number (NSN) in this column. The host component NSN shall contain dashes within the NSN and the NSN of subset component shall be without dashes causing the NSN to be indented to the left three spaces.

3.2.3.10 **"Illustration Number" Column.** This column shall contain the number identity of illustrations on the facing page (see Figure 18) to agree with the NSN and corresponding description data. The numbering shall be sequencial beginning with "0001." The first facing illustration page shall contain a title page on its reverse side for Section II "SECTION II: ILLUSTRATED COMPONENT LIST/HAND RECEIPT". This title page allows the illustration page(even-numbered pages) to face the correct item description page (odd-numbered pages). When an illustration is not used, the NSN and the words "Not Illustrated" shall be placed on the facing illustration page in the location where the illustration would normally appear and the illustration number column would remain blank on the description data page.

3.2.3.11 **"Item Description" Column.** Items shall be arranged in alphabetical sequence in the item description column by federal item name with each accounting requirements code (ARC) group, i.e., N, D, or X. Item names shall be followed by an adequate description sufficient to identify the component. When type classified, the line item number (LIN) shall be listed in the item description column. Subset items shall be listed in the same manner as described above. When there is more than one subset to an SKO, the subset(s) components shall be identified by an alphabetic designator beginning with "A" after each subset component with the second subset component identified with a "B", etc.

3.2.3.12 **"RMKS" (Remarks) Column.** This column shall contain a listing of all applicable codes. The codes shall be explained in Section I under "Special Information".

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

3.2.3.13 **"ARC" (Accounting Requirements Code) Column.** This column shall contain the status code of all items. The codes shall be explained in Section I.

3.2.3.14 **"CIIC" (Controlled Item Inventory Code) Column.** The CIIC shall be recorded as specified in DA Pam 710-2-1.

3.2.3.15 **"UI" (Unit of Issue) Column.** This column shall indicate a two-position code which shows the smallest quantity of supply, per item, indicated by the Army Master Data File (AMDF) that an item can be requisitioned and issued.

3.2.3.16 **"Quantity Authorized" (Qty Aut) Column.** This is the quantity authorized to be on-hand for each end item or component of an end item.

3.2.3.17 **"Quantity on Hand" (A thru E) Column.** Five columns shall be on each description sheet to allow for recording the actual item quantities on hand for each component listed for five separate inventories. Each column shall be headed with a capital letter "A" thru "E". When quantities are not accountable, quantity on hand block shall be shaded.

3.2.3.18 **Last Page/Signature Blocks.** Sufficient blank space shall be allowed on the last page of Section II to provide for the addition of locally authorized items. Below this blank space, sufficient vertical space shall also be allowed in the quantity on hand columns for a signature, rank, and the date of the inventory (see Figure 19).

3.2.4 **Section III, Associated Support Items of Equipment (ASIOE).**

This section shall contain a title page "SECTION III: ASSOCIATED SUPPORT ITEMS OF EQUIPMENT (ASIOE)" which allows the illustration page to face the correct item description page. This section shall be an illustrated NSN item description section in a non-hand receipt format (see Figures 20 and 21). The items listed in Section III beginning with page III-003 are ASIOE end items dedicated to the operation and/or maintenance of the medical SKO. ASIOE are separately documented in Table of Organization and Equipment (TOE)/The Army Authorization Documents System (TAADS) and are not to be considered components of the medical set. They are listed in this section for information purposes only and do not

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

constitute an additional authorization. Total authorization is reflected on the unit's Modification Table of Organization and Equipment (MTOE).

3.2.4.1 **Contents, Section III**. The following block and column titles are used in formatting the Section III page and are the same as explained in Section I and Section II of this specification.

"End Item Stock No" Block
"End Item Description" Block
"Publication Number" Block
"Publication Date" Block
"Act" (Action Code) Column
"Stock Number" Column
"Illustration Number" Column
"Item Description" Column
"Rmks" (Remarks) Column
"ARC" (Accounting Requirements Code) Column
"CIIC" (Controlled Item Inventory Code) Column
"UI" (Unit of Issue) Column.
"Qty Aut" (Quantity Authorized)

"Qty on Hand" Column. There shall be no inventory columns provided for Section III of this SC. Items described here are not components of the SKO, are authorized only by TOE/MTOE documentation and are provided for information purposes only.

3.2.5 **Section IV, American Hospital Formulary Service Therapeutic Classification**. This section is a therapeutic index constructed in accordance with the American Hospital Formulary Service (AHFS) to assist physicians in prescribing items initially supplied with this medical set. This index lists the FSC 6505 components of the medical SKO with a therapeutic classification in NSN sequence within each therapeutic class (see Figure 22). Permission to use the pharmacological-therapeutic classification systems of the AHFS has been granted by the American Society of Hospital Pharmacists. For complete information concerning all drugs, consult the AHFS, copyright 1959-1993, American Society of Hospital Pharmacists, Inc., all rights reserved.

3.2.5.1 **Contents, Section IV**. The following block titles are used in formatting the Section IV page and are explained in Section I and II of this specification:

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

"End Item Stock Number" Block
"End Item Description" Block
"Publication Number" Block
"Publication Date" Block

3.2.6 **Section V, NSN/Illustration Number/Page Number Cross Reference.** This section shall cross-reference the NSN in NSN sequence of each component with the corresponding illustration number to the page where the item is listed in Section II and III (see Figure 23).

3.2.6.1. **Contents, Section V.** The following block titles are used to format this section and are explained in Section I and II of this specification:

"End Item Stock Number" Block
"End Item Description" Block
"Publication Number" Block
"Publication Date" Block

3.2.6.2 **Data Description:**

"NSN" Column. Each component NSN of the SC shall be listed here in NSN sequence.

"Ill No." Column. The illustration number, when applicable, is listed here to its corresponding NSN component of the SC.

"Page No." Column. The page number listed here is where the corresponding NSN component and its illustration number, when applicable, is located within the SC.

3.2.7 **Section VI, NSNS Deleted from Previous Publications and Applicable Replacements.** This section shall appear in all editions subsequent to the previous edition of the Supply Catalog. It shall include all deletions and replacements made to Section II since the previous edition (see Figure 24). The following note shall be included immediately after the last item "The actions reflected in this section are cummulative actions resulting from the annual update process to this SKO. The data shown above should be used to bring your set, kit, or outfit and management records up-to-date."

3.2.7.1 **Contents, Section VI.** The following block and column titles are reflected in this section and are explained in Section I and II of this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

"End Item Stock Number" Block
"End Item Description" Block
"Publication Number" Block
"Publication Date" Block

3.2.7.2 **Data Description:**

"Deleted NSN" Column. This column shall sequentially list the NSNs deleted since the previous publication of this SKO.

"Replacement NSN" Column. This column shall list the NSN of the replacement item since the previous publication of this SKO.

"Replacement NSN Description" Column. This column shall reflect the item description of the replacement NSN since the previous publication of this SKO.

3.2.8 **Concluding Material**. The concluding material in the SC shall include an authentication page (see Figure 25 for an example).

3.3 **Page numbering**. Page numbers shall be centered at the bottom of each page

3.3.1 **Cover and numbering**. The cover shall not be numbered on the front or the back.

3.3.2 **Section I, II, III, IV, V, VI page numbering**. Pages in these sections shall be numbered consecutively by Arabic numerals (1, 2, 3, 4, etc.) beginning with 1 and prefixed with the upper case Roman numeral representing the section number. For example, page number II-006 is the sixth page of Section II.

3.3.3 **Concluding material page numbering**. The concluding material pages of the SC shall not be numbered.

3.4 **Horizontal format**. Supply catalogs shall be prepared in horizontal format (across the broadside of the page). The catalogs shall be printed head to foot.

3.5 **Page size**. Unless otherwise specified, final trim size shall be 10-3/4 x 8-1/4 inches.

3.6 **Standard holes**. Standard 3 hole notebook drilling shall be used: 3 holes, 3/8" diameter, 4-1/4" center.

Source: https://assist.dla.mil -- Downloaded: 202610-26T04:01Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 **Responsibility for inspection.** Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in this specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 **Responsibility for compliance**. All items shall meet all requirements of sections 3 and 5. The inspection set forth in this specification shall become a part of the contractor's overall inspection system or quality program. The absence of any inspection requirements in the specification shall not relieve the contractor of the responsibility of assuring that all products or supplies submitted to the Government for acceptance comply with all requirements of the contract. Sampling inspection, as part of manufacturing operations, is an acceptable practice to ascertain conformance to requirements, however, this does not authorize submission of known defective material, either indicated or actual, nor does it commit the Government to acceptance of defective material.

4.2 **Quality assurance provisions**. The contractor shall be responsible for the performance of quality assurance provisions in accordance with MIL-M-85337 to the extent specified by the contracting activity.

5. PACKAGING

5.1 **Packaging requirements.** The requirements for packaging shall be in accordance with ASTM D3951.

5.2 **Marking.** Exterior of cartons containing supply catalogs shall be marked in accordance with MIL-STD-129 and shall contain the appropriate supply catalog number and publication date.

Source: https://assist.dla.mil -- Downloaded: 202610-26T04:01Z
Check the source to verify that this is the current version before use.

6. NOTES

(This section contains information of a general or explanatory nature that may be helpful, but is not mandatory.)

6.1 **Intended use**. This specification is intended for use in preparation of SCs which contain medical sets, kits, and outfits. The catalogs contain the technical data covering the components listed and are used for inventory/accountability of these components/items by operators and maintenance personnel.

6.2 **Acquisition requirements**. Acquisition documents must specify the following:

a. Title, number, and date of this specification
b. The supply catalog publication number
c. Distribution statement and destruction notice (as applicable)
d. Issue of DODISS to be cited in the solicitation, and if required, the specific issue of individual documents referenced (see 2.1.1)
e. Optional requirements in this specification, identification of option by name and paragraph number, details of the options
f. Any data deliverable as required by selected quality assurance provisions

6.3 **Data requirements**. The following data item descriptions (DIDs) must be listed, as applicable, on Contract Data Requirements Lists (DD Form 1423) when this specification is applied on a contract, in order to obtain the data, except where DOD FAR Supplement 27.475-1 exempts the requirements for a DD Form 1423.

Specification	DID Number	DID Title
MIL-M-85337 (NAVY) Paragraph Number		
	DI-M-2194	Technical Manual Quality Assurance Program Plan
	DI-M-2195	Validation Plan
	DI-M-2196	Validation Certification
	DI-M-2197	Evaluation Record
	DI-M-2198	Verification Plan

Source: https://assist.dla.mil -- Downloaded: 202610-26T04:01Z
Check the source to verify that this is the current version before use.

The listed DIDs were those cleared as of the date of this specification. The current issue of DOD 5010.12-L, Acquisition Management Systems and Data Requirements Control List (AMSDL) must be researched to ensure that only current, cleared DIDs are cited on DD Form 1423.

6.4 **Technical manual acquisition**. This specification must be listed on the Contract Data Requirements List (DD Form 1423) in order to acquire the technical manuals described by this specification, except where DOD FAR Supplement 27.475-1 exempts the requirements for a DD Form 1423.

6.5 **Definitions**. Definitions listed here are applicable to this specification.

6.5.1 **Sets, kits, outfits (SKO)**. A collection of medical materiel and equipment designed to perform a medical maintenance function or a specific medical mission. SKOs are depicted by and accounted for in a supply catalog (SC). Normally Associated Support Items of Equipment (ASIOE) type items are customarily approved through the basis of issue plan processes and fulfill the Army Authorization Documentation System (TAADS) requirements.

6.5.2 **Set**. A collection of medical equipment and supplies/tools used by a group (section, squad, platoon or unit) to perform an organizational medical mission. It is type classified, assigned a LIN, and the unit of issue is Set.

6.5.3 **Supply Catalog (SC).** A supply catalog provides supply management and item identification data necessary for the maintenance of Army supplies and equipment. These SCs furnish information relative to the components that comprise a "collection type" item of supply. Collection type items are those normally issued and identified as equipment, groups, kits, outfits, plants, sets, or systems. They are type classified and assigned a logistics code.

6.6 **Subject term (key word) listing.** The following is a list of subject terms that allows for identification of this SC during retrieval searches.

component list
type classified

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

*SC 6545-8-MG1

DEPARTMENT OF THE ARMY SUPPLY CATALOG

SETS, KITS, AND OUTFITS
COMPONENTS LIST/HAND RECEIPT

MEDICAL MATERIEL SET
MINIMAL CARE WARD DEPMEDS (L311)

NSN 6545-01-299-5619
LIN: M48055

*THIS PUBLICATION SUPERSEDES SC 6545-8-MG1 DATED 29 JUNE 1990, INCLUDING ALL CHANGES

HEADQUARTERS, DEPARTMENT OF THE ARMY

30 JUNE 1993

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Figure 1, Example - Supply Catalog Cover Page

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

SECTION I - INTRODUCTION

1. GENERAL.

a. This catalog provides an alphabetical listing of the Medical Materiel Set (MMS) Minimal Care Ward DEPMEDS (L311).*

b. The component listing/hand receipt is identified by a four-segment symbol:

(1) The first segment (SC) indicates Supply Catalog.

(2) The second segment (6545) identifies the Federal Supply Classification (FSC).

(3) The third segment (8) identifies the compiler.

(4) The fourth segment (MG1) distinguishes the SKO from others in the same FSC.

c. The title SKO is authorized by The Army Authorization Documents Systems (TAADS).

d. The term SKO in the SC introduction will be replaced by the specific title for each SC (i.e., MES, DES, MMS, etc.).

e. This Catalog is not an authorization for requisitioning NSN 6545-01-299-5619. Components listed may be requisitioned in quantities indicated when the set is authorized by Modification Table of Organization and Equipment (MTOE) or other Department of the Army (DA) authorization document(s).

2. PURPOSE.

This SC prescribes initial allowances and quantities of medical materiel necessary to provide ancillary nursing services for care and treatment of twenty (20) patients requiring minimal care/supervision and applicable Corps and CommZ-Medical Treatment Facilities. This constitutes the minimum basic load to sustain this SC.

3. RESPONSIBILITY.

a. The Defense Medical Standardization Board (DMSB) is responsible for the overall management of the DEPMEDS MMS. The DMSB approves the composition of the DEPMEDS MMS. The Surgeon General (TSG) is responsible for budgeting and for the initial issue of the DEPMEDS MMS to include all procurement appropriation funded medical equipment.

*This publication supersedes SC 6545-8-MG1 dated 29 June 1990.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 2, Example - Section I, Introduction Page 1

b. USAMMA coordinates requirements with the Defense Personnel Support Center (DPSC). USAMMA performs staff-management actions as directed by TSG and as outlined in AR 40-61.

c. USAMMA develops the requirements; OTSG coordinates programming and budgeting from the proper appropriation (OMA, OMAR, OMNG, OPA, or Stock Fund**) for procurement of newly developed or significantly revised MMS and for Associated Support Items of Equipment (ASIOE).

**Reference AR 40-61 dated 01 August 1989, Section I, Abbreviations Glossary.

d. After initial fielding, MACOMs (in addition to OCAR and the ARNG) are responsible for follow-up requisitioning, storing, and maintaining those MMS and equipment authorized to be on-hand in TOE units or TDA activities. MACOMs program and budget for replacement of consumable and expense equipment components, and for minor changes to MMS on-hand at TOE units.

4. EXPLANATION AND USE OF COLUMNS.

The following information is a brief description of the contents and layout of data provided to user(s) of this Supply Catalog:

a. ACT (Action Code). A one-position alphabetic code that records the action taken in a component list/hand receipt. They are:

C (CHANGE) Indicates a change to the AMDF unit of issue, set measurement quantity/unit of issue, ARC, or Remarks Code since the previous publication.

D (DELETION) Indicates the NSN has been deleted and not replaced.

N (NEW) Indicates the NSN was not previously included in the basic publication or changes.

R (REPLACEMENT) Indicates the NSN has been replaced or changed.

b. ILLUSTRATION/NATIONAL STOCK NUMBER (NSN). Not all the NSNs in this SC are represented by an illustration. The items that have an illustration show an illustration number in the "ILL NO" column of the CL/HR. For easy reference, the illustration number is also listed under each illustration.

(1) The illustrations may not be exact representations of the items in the assemblage since the products of the various manufacturers may vary in appearance.

(2) A cross-reference list of NSN-to-Illustration Number is provided as Section V. This list is in NSN sequence and will further assist in identifying NSNs that have an illustration.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 3, Example - Section I, Introduction Page 2

c. ITEM DESCRIPTION. The item descriptions are listed in alphabetical sequence by item name with additional descriptive information required for identification.

(1) The first description shown is that of the MMS. Subsequent descriptions are for the components by item names in alphabetical and ascending size sequence within each Accounting Requirements Code (ARC) group (i.e., N, D, and X).

(2) The statement "Item Not Illustrated" indicates that an illustration should not appear.

(3) Manufacturer's code, part number, and/or the applicable specification references follow the description, when applicable.

d. RMKS (REMARKS). Space reserved for notes which are explained in the "Special Information" paragraph.

e. ARC (ACCOUNTING REQUIREMENTS CODE). This is a one-position alphabetic code that identifies the way an item is accounted for by the Army. One of the following codes must be entered here.

N-NONEXPENDABLE - This code identifies identifies items which are not used up and keep their identity throughout their life and must be accounted for in the property book.

D-DURABLE - This code identifies items which are not used up. These items keep that identity throughout their life. Items identified by this code are accounted for by hand-receipt or other similar documents.

X-EXPENDABLE - This code identifies items which will be used up. These items do not need to be entered on hand receipts; however, preprinted hand receipts list expendable items for information purposes only.

f. CIIC (CONTROLLED ITEM INVENTORY CODE). This is a one-position alphabetic/numeric code which indicates the security classification, security risk, or pilferage controls required for storage and transportation of DOD assets, DA PAM 710-2-1 identifies those items that must be inventoried monthly. The CIIC Code applicable to components of this SC unless otherwise indicated in the CIIC column, is U (UNCLASSIFIED).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 4, Example - Section I, Introduction Page 3

g. UI (UNIT OF ISSUE). This is a two-position code which shows the smallest quantity that an item can be requisitioned and issued.

h. PEC (PRIMARY ESSENTIAL CODE). The revised editions of AR 40-61 and AR 200-1 will eliminate the requirements for PECs for Equipment On Hand (EOH) computation procedures. References to PECs in the future UA Listings (UAL) and Supply Catalogs (SCs) will not be shown.

h. QTY AUTH (QUANTITY AUTHORIZED). This is the quantity authorized to be on-hand for each end item or component of an end item.

i. QUANTITY ON HAND (A Thru F). These columns are for the actual quantity on-hand for each component listed.

5. ABBREVIATIONS:

Standard abbreviations are listed in AR 310-50 (Authorized Abbreviations and Brevity Codes) and the DoD Medical Catalog Introduction (Authorized Abbreviations Referenced to Terms). Some words will be truncated (with vowels removed that do not meet the above references) in order to fit a limited-space nomenclature field. Army acronyms used in this SC are defined in AR 40-61, Glossary, Section I Abbreviations. Nonstandard abbreviations listed are used in this publication:

Abbreviation	Explanation
ABS	Absorbent, Absorbable
ABSORB	Absorbable
AC	Alternating Current
ACET	Acetate
ACR	Acrylic
ACRY	Acrylic
ADH	Adhesive, Adhere
ADMIN	Administration
ADJ	Ajustable
ADL	Adult
ADU	Adjustable
AE	Antiembolism
AERS	Aerosol
AHP	Angle Handpiece
AL	Aluminum, Alloy
ALC	Alcohol
ALUM	Aluminum
AMPS	Amperes
ANALYZ	Analyzer
ANAS	Anastomosis
ANER	Aneroid
ANES	Anesthesia
ANG	Angle
ANGIO	Angioplasty
ANGLD	Angled
AORT	Aorta
APP	Apparatus
APPAR	Apparatus
APPROX	Approximately
APPRX	Approximately
ARM	Armed
ASIF	Association For The Study of Internal Fixation
ASST	Assortment
AUTO	Automatic
AUTOCL	Automatic Closure, Autoclavable

Figure 5, Example - Section I, Introduction Page 4

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Abbreviation	Explanation
BACITR	Bacitracin
BAL	Balloon
BAND	Bandage
BAT	Battery
BAT CHAR	Battery Charger
BAT/OP	Battery Operated
BEVL	Bevel
BIO	Biological
BIOLOG	Biological
BICAR	Bicarbonate
BILI	Bilirubin
BILIR	Bilirubin
BLD	Blood
BLDE	Blade
BLK	Black
BLN	Blunt
BNK	Bank
BR	Braided
BRST	Breast
BT	Bottle
C	Celsius, Cycle, Centigrade
CØ2	Carbon Dioxide
CAL	Calcium
CAP	Capsule, Capacity
CAPS	Capsules
CARB	Carbide
CARDIO	Cardiovascular
CART	Cartridge
CART-NDL	Cartridge Needle
CAS	Cassette
CASS	Cassette
CAT	Catheter
CATH	Catheter
CATH&NDL	Catheter and Needle
CATH-TUBE	Catheter-Tube
CH	Child
CHAM	Chamber
CHAMB	Chamber
CHEM	Chemical
CHI-ADU	Child-Adult
CHL	Chloride
CHLD	Child
CHLORAM	Chloramphenicol
CHLORAME	Chloramphenicol
CHRM	Chrome
CIR	Circle
CLIN	Clinical
CLINDAMY	Clindamycin
CLO	Closure
CLR	Clear
CM	Centimeter
COL	Collecting
COLL	Collection, Collect
COLLECTI	Collecting
COMB	Combination
COMP	Compressor
CONCEN	Concentration
CONN	Connector
CONSEC	Consecutive
COT	Cotton
CR	Cream
CRES	Crescent
CRM	Corrosion-Resisting Metal
CRS	Corrosion Resisting Steel
CRVD	Curved
CS	Carbon Steel
CTN	Cotton
CU	Cubic
CULT	Culture
CUSH	Cushion
CVD	Curved
CVR	Cover
CVRS	Covers

Source: https://assist.dla.mil -- Downloaded: 2026-02-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 6, Example - Section I, Introduction Page 5

Abbreviation	Explanation
CYL	Cylinder, Cycle, Cylindrical
CYLIN	Cylindrical
D	Deep
DA	Double Armed
DARKRM	Darkroom
DBL	Double
DC	Direct Current
DEG	Degree
DEN	Dental
DENT	Dental
DEPMEDS	Deployable Medical Systems
DES	Dental Equipment Set
DET	Detachable
DETACH	Detachable
DETERG	Detergent
DEVEL	Developer
DEX	Dexamethasone
DFLCTNG	Deflecting
DIA	Diameter
DIAG	Diagnostic, Diagonal
DIAGN	Diagnostic
DFLCTNG	Deflecting
DILUT	Diluting
DIS	Disposable
DISP	Disposable
DISPEN	Dispensers
DISPO	Disposable
DISPSNERS	Dispensers
DP	Deep, Durable Press
DRESS	Dressing
EA	Each
ECG	Electrocardiogram
ELAS	Elastic
ELEC	Electric
EDTA	Ethylenediaminetetra Acetic Acid
EKG	Electrocardiogram
ELECTROSURG	Electrosurgical
ENDOTRAC	Endotrachael
ENDOTRACH	Endotrachael
ENE	Enema
EPIN	Epinephrine
EQ	Equivalent
EQUI	Equivalent
EQUIP	Equipment
EQUIV	Equivalent
ERYTHROM	Erythromycin
EX	Examining
EXAM	Examining
EXP	Explosion
EXT	Extension, Extra
F/	For
FAB	Fabrication
FABRCTN	Fabrication
F/ALL	For All
FED	Federal
FIB	Fiber
FILT	Filter
FIX	Fixer, Fixed
FL	Fluid
FLD	Field
FLEX	Flexible
FLM	Film
FLR	Floor
FNGR	Finger
FOLD	Folding
FRICT	Friction
FOUNT	Fountain
FR	French
FRM	Form

Figure 7, Example - Section I, Introduction Page 6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Abbreviation	Explanation
FT	Feet/Foot
FST	Fastener
F/SZ	For Size
GA	Gage
GAL	Gallon
GEL	Gelatin
GEN	General
GL	Gallon
GLA	Glass
GLAS	Glass
GLU	Glucose
GM	Gram
GMS	Grams
GP	General Purpose
GRA	Gramicidin
GRP	Group
GYN	Gynecology
H	High
HCL	Hydrochloride
HEMO	Hemostatic
HEXAG	Hexagonal
HG	Height
HI	High/Highly
HNDL	Handle
HNDPC	Handpiece
HOR	Horizontal
HORIZ	Horizontal
HOSP	Hospital
HRS	Hours
HTD	Heated
HYDROCOR	Hydrocortisone
HYDROX	Hydroxide
HYP	Hypodermic
HYPO	Hypodermic
HZ	Hertz

Abbreviation	Explanation
I	Inch
ID	Inside Diameter, Identification
ILLUM	Illuminator
IMP	Impregnated
IMPREG	Impregnated
IN	Inch, Injection
IND	Individually/Individual
INDI	Indicating
INDIC	Indicating
INF	Infant
INF-CH-AD	Infant-Child-Adult
INFO	Information
INJ	Injection
INS	Inserts
INST	Instrument
INSTR	Instrument(s)
INSUL	Insulated
INTERMIT	Intermittent
INTRA	Intravenous
INTRAVEN	Intravenous
INTRO	Introduction
INVER	Inverted
INW	Inch Wide
IOTHALA	Iothalamate
I.S.	Individually Sealed
ISO	Isopropyl
IV	Intravenous
KET	Ketone(s)
KG	Kilogram
KILO	Kilogram
KT	Kit
L	Long
LAB	Laboratory
LACT	Lactated

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 8, Example - Section I, Introduction Page 7

Abbreviation	Explanation
LG	Length, Long, Large
LGTH	Length
LI	Liter
LIQ	Liquid
LO	Low
LPM	Liters Per Minute
LRG	Large
LRGE	Large
LSE	Leaf
LT	Light
M	Meter
MACH	Machine
MAG	Magnesium
MAGNES	Magnesium
MAN	Manual
MATL	Materials
MAX	Maximum
MCG	Microgram
MCGM	Microgram
MECH	Mechanical/Mechanic
MED	Medium, Medical, Medicinal
MEDROXYPR	Medroxyprogesterone
MEG	Megetholate
MEQ	Milliequivalents
MER	Mercury
MERC	Mercury
MERTH-OD-FORM	Merthiolate-Iodine-Formalin
METR	Metered
MG	Milligram
MICON	Miconazole
MICROSURG	Microsurgery
MIN	Minimum, Minute
MISS	Medical Instrument Supply Set
MIX	Mixing
ML	Milliliter
MM	Millimeter
MODIF	Modified
MONIT	Monitor
MONO	Monofilament
MR	Meter
MT	Mount
MTD	Mounted
NAS	Nasal
NASOPHAR	Nasopharyngeal
NDL	Needle
NEG	Negative
NF	National Formulary
NIT	Nitrate
NITR	Nitrite, Nitrate
NLD	Needle
NO	Number
NONADJ	Nonadjustable
NONABS	Nonabsorbable, Nonabsorbent
NONAUTO	Nonautoclavable
NONAUTOCL	Nonautoclavable
NONCORR	Noncorrosive
NONFLAM	Nonflammable
NONSTER	Nonsterile
NONWOV	Nonwoven
NONWVN	Nonwoven
NORM	Normal
NYST	Nystatin
OB	Obstetrical
OCCU	Occlusion
O/D	Outside Diameter
OD	Outside Diameter, Olive Drab

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 9, Example - Section I, Introduction Page 8

Abbreviation	Explanation
OP	Operated, Open
OIN	Ointment
OINT	Ointment
OPER	Operated, Operating
OPHTH	Ophthalmic
OPNG	Opening
OPT	Optic
O.R.	Operating Room
ORAN	Oranasal
ORTHO	Orthopedic
OSCILL	Oscillations
O/T	Outlet
OXY	Oxygen
OZ	Ounce
PØ2	Oxygen Partial Pressure
PAPR	Paper
PAR	Partial
PAT	Patient
PDWR	Powder
PED	Pediatric
PG	Package
PH	Potential Hydrogen
PHARYN	Pharyngeal
PHENYLEP	Phenylephrine
PHENYLPRO	Phenylpropanolamine
PHOS	Phosphate
PL	Plus, Plated
PLAS	Plastic
PLTD	Plated
PLYSTR	Polyester
PNEU	Pneumatic
POLY	Polyester, Polyurethane, Polyethylene,
POLYGL	Polyglycolic, Polyglactin
POLYMY	Polymyxin
POLYPROP	Polypropylene
POLYUR	Polyurethane
PORT	Portable
POS	Positive, Position
POSTOP	Post Operative
POT	Potassium
POVIDONE-IOD	Povidone-Iodine
POV-IOD	Povidone-Iodine
PR	Pair
PREP	Preparation
PRES	Pressure
PRF	Proof
PRIM	Primary
PROC	Procedure
PROCH	Prochlorperazine
PROHIB	Prohibits
PROT	Protein
PROTECT	Protective
PRTBLE	Portable
PRTO	Protein
PSI	Pounds Per Square Inch
PT	Pint, Points
PTBL	Portable
PTS	Points
PUR	Purpose
PURP	Purpose
PWDR	Powder
QS	Sufficient Quantity
QT	Quart
RAD	Radiological
RADGR	Radiographic
RADIO	Radiopaque, Radiological
RADIOP	Radiopaque
R-ANG	Right Angle
RAY	Rayon
RBBR	Rubber
RD	Round, Radiographic
REAG	Reagent

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 10, Example - Section I, Introduction Page 9

Abbreviation	Explanation
REC	Rectal
RECEP	Receptacle
RECIP	Reciprocating, Reciprocation
RECT	Rectal
REF	Refrigerator
REFRIG	Refrigeration, Refrigerator
REG	Regular, Regulator
REINF	Reinforced
REMOV	Removable
RES	Rest
RESP	Respiratory
REV	Reverse
REVERS	Reversible
RH	Rhodium
RNG	Ring
RO	Roll
RPM	Revolutions Per Minute, Rotations Per Minute
RUB	Rubber
RUBB	Rubber
SA	Single Armed
SE	Set
SEA	Sealed
SEAL	Sealed
SEL	Self
SEN	Sensitive
SENS	Sensitive
SENSI	Sensitive
SEP	Separate
SERA	Serum
SET	Setting
SFTY	Safety
SGL	Single
SH	Sheet
SHPE	Shaped, Shape
SIL	Silicone
SIMETH	Simethicone
SM	Small
SML	Small
SMPL	Sample
SNG	Single
SNGL	Single
SOD	Sodium
SOL	Solution, Soluble
SPD	Speed
SPEC	Specimen
SPHER	Spherical
SPHYGMO	Sphygmomanometer
SQ	Square
SS	Stainless Steel
STD	Standard
STER	Sterile, Sterilize, Sterilizing, Sterilized
STERIL	Sterilization
STERLZR	Sterilizer
STETHO	Stethoscope
STL	Steel
STR	Strength, Straight
SUBNORM	Subnormal
SUCT	Suction
SUL	Sulfate
SULF	Sulfate
SUPOS	Suppositories
SURG	Surgical, Surgery
SURGY	Surgery
SUS	Suspension
SUST REL	Sustained Release
SUT	Suture
SYN	Synthetic
SYNTH	Synthetic

Figure 11, Example - Section I, Introduction Page 10

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Abbreviation	Explanation
SYR	Syringe
SYR NDL	Syringe Needle
SYS	System
SZ	Size
TABS	Tablets
TEMP	Temperature
TENN	Tennessee
TERM	Terminal
TH	Thick
THK	Thick
THR	Thread
TOW	Towel
TP	Tip
TRACH	Tracheal,Tracheotomy
TREAT	Treatment
TUNGS	Tungsten
TY	Type
U	University
UN	Unit
UNC	Unified National Coarse
UNF	Unified National Fine
URE	Urethral
URETH	Urethral
UNIV	Universal
UNIVER	Universal
URIN	Urinalysis
UROBIL	Urobilinogen
USP	United States Pharmacopoeia
V	Volt(s)
VAG	Vaginal
VAL	Valve
VAR	Variable
VIS	Visual
VOL	Volume
VENT	Vented
VERT	Vertical
VET	Veterinary
VS	Versus
W/	With
W	Wide/Watts, Wide, With
W/BAT-CHAR	With Battery Charger
WBC	White Blood Cell
WD	Wide
W/FLEX	With Flexible
W/HDL	With Handle
WHL	Whole
W/O	Without
W/SEP	With Separate
WT	Weight
WVN	Woven
X	Extra
XLGE	Extra Large
XLR	Extra Large
YD/YDS	Yard(s)

6. SPECIAL INFORMATION.

a. The items listed in Section III of this SC beginning with Page III-03 are ASIOE end items dedicated to the operation and/or maintenance of this medical assemblage.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 12, Example - Section I, Introduction Page 11

b. ASIOE are separately documented into TOE/The Army Authorization Documents System (TAADS) and are not to be considered components of this medical assemblage.

c. They are listed in this section for information purposes only and do not constitute an additional authorization. Total authorization is reflected on the Unit's MTOE.

***2 indicators will still be reflected in the Remarks column of this report noting these items will appear in Section II of the Unit Assemblage List (UAL). THE ITEMS ON THIS REPORT WITH ***2 INDICATORS WILL NO LONGER APPEAR IN THE SECTION II HAND RECEIPT PORTION OF THIS SC. The previously used

***3 indicators have been removed and those items are now part of the Section II hand receipt portion of this SC.

d. To assist physicians in prescribing items initially supplied with this set, a Therpeutic Index constructed in accordance with the American Hospital Formulary Service (AHFS) is provided on page IV-01. The Therapeutic Index may not be required for all SCs. This index lists the FSC 6505 components of the major assemblage with a therapeutic classification in NSN sequence within each therapeutic class.

e. Permission to use the Pharmacologic-Therapeutic Classification Systems of the AHFS has been granted by the American Society of Hospital Pharmacists. The Society is not responsible for the accuracy of transpositions, or additions, or excerpts from the original context. For complete information concerning all drugs, consult the AHFS, copyright 1959-1987, American Society of Hospital Pharmacists, Inc., All Rights Reserved.

f. REMARKS COLUMN: The symbols shown in the Remarks Column apply to this SC only and include the following:

F - Subject to Damage by Freezing
G - Requires Refrigeration between 2 degrees to 8 degrees Centigrade (35 degrees to 46 degrees Fahrenheit)
I - Flammable or Oxidizing
P - Potency Type Items
***2 - ASIOE Items

g. COMPONENTS LISTINGS.

(1) To support Army Medical Department Activities, USAMMA also provides floppy disks for UA component listings in lieu of hard-copy versions. For each UA requested, the disk will consist of UA information in raw data format and a report file which is in the same two-section format as the current hard-copy component listing. A brief instruction guide on "How to Access and Print the Information" is included on the disk. Additionally, a separate enclosure with the data record layout of the raw data will be included to allow the user to merge the data with the user's existing software, i.e. DBase 3, Lotus, or Paradox.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 11, Example - Section I, Introduction Page 12

(2) USAMMA will not send a hard-copy version of the component listing to those activities that have the floppy disk system. Only one method of receiving listings is allowed. The disk will be furnished just once during the current year. Only the addition of a new set by MTOE update will permit the release of additional disks per Unit Identifier Codes (UIC). Commands responsible for more than one unit should indicate this and furnish the various UICs and Uas under their command on their request. It is recommended that this original USAMMA disk be copied to your hard drive and the USAMMA disk be retained as the official record.

(3) Once an activity receives their initial Uas on disk(s), their address will be retained for an automatic distribution of the next succeeding year's annual update via disk. Activities that have automated capabilities and choose the disk method, should send their complete address and UIC along with all UAs authorized under that UIC to:

Commander
U.S. Army Medical Materiel Agency
ATTN: SGMMA-OCU
Fort Detrick
Frederick, MD 21702-5001

(4) Component authorizations are published in medical materiel set computer listings/floppy disks available from USAMMA and in the appropriate supply catalogs (SC 6545-8 Series). Revisions to components are published annually in the DA SB 8-75 Series and are reflected in current year computer listings/floppy disks.

(5) Publication of component authorization changes in the SB 8-75 Series (and/or a current year medical materiel set computer listing/floppy disk) constitutes authority for updating assemblages in respective units and serves as the authorization document of assemblage component.

(6) Commanders of DEPMEDS Units will continue to utilize the UA list initially fielded with the unit until authorized for update by USAMMA. This may result in the use of several UA listings with different dates for different assemblages until a single, current year update is provided by USAMMA. Commanders should annually request a current year medical unit assemblage listing/floppy disk by submitting written requests to:

Commander
U.S. Army Medical Materiel Agency
ATTN: SGMMA-OCU
Fort Detrick
Frederick, MD 21702-5001
Telephones: DSN: 343-7203
COMM: 301-619-7203

The request should identify the unit assemblage (UA) number, Nomenclature, Line Item Number (LIN), or NSN for each medical materiel set required.

Figure 14, Example - Section I, Introduction Page 13

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

(7) Commanders who do not have automated capabilities and need the hard copy listing should continue to request a current-year medical materiel UA listing by submitting written requests to the aforementioned address. The request should identify the UA Number, Nomenclature, Line Item Number (LIN), or NSN for each medical equipment set required.

h. AMDF INCOMPATIBILITIES. If incompatibilities are discovered between this publication and the AMDF (Army Master Data File), the latter will govern.

i. Section VI, NSNS DELETED FROM PREVIOUS PUBLICATIONS AND APPLICABLE REPLACEMENTS.

(1) This Section appears in all editions subsequent to the previous edition of the Supply Catalog.

(2) It includes all deletions and replacements made to Section II since the previous edition.

7. ACCOUNTABILITY.

a. Page II-2, Section II (FROM:) identifies the element from which the property is issued. Enter the name of the organization, unit, section, or squad which issues the property.

b. The block (TO:) identifies the element to which the property is issued. Enter the name and rank of the person receiving the property. TOE/MTOE paragraph, line number, and job title may be added as optional information.

c. HAND RECEIPT NUMBER: Enter a locally assigned number.

d. END ITEM STOCK NUMBER: Enter the NSN of the MMS.

e. END ITEM DESCRIPTION: Enter a short description of the MMS.

f. PUBLICATION NUMBER: SC Number.

g. PUBLICATION DATE. SC Effective Date.

h. QUANTITY ON HAND: These columns are for the actual quantity-on-hand for each component listed.

(1) Enter the quantity on-hand for each component listed. Line out all unused blocks in columns with recorded quantities.

(2) Advance all quantities to the next column when quantities change. Quantities must be advanced when changing hand-receipt holders.

(3) The person receiving the property will sign, enter his/her rank, and date the proper quantity column on the last page (ink entry). The last page is the last-numbered page and may be an odd or even number. The last page may be reserved for signature only. (The original page will have an original signature; the copy may have a carbon signature.)

Figure 15, Example - Section I, Introduction Page 14

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

8. TRANSPORTATION INFORMATION.

a. Supplies will be packed in containers suitable for operations under field conditions.

b. Packages or containers will ordinarily be of such size and weight as to permit handling by two men.

c. When practicable, crating will be such as to permit equipment to be readily accessible by removing the front, side, or top of the container.

d. To reduce the overall bulk when this assemblage is packed by the Depot, empty chests which appear as components on the equipment list will be utilized wherever possible as shipping containers.

e. The following information is applicable to this MMS.

Total Weight and Volume of this MMS:

Weight: 2,982.12
Volume: 366.769

9. POWER REQUIREMENTS. An external power source is required to support this MMS. The following NSNs operate on 115V, single-phase power with wattage requirements as shown:

NSN	ITEM	VOLTAGE	WATTS
6515-01-187-6606	Thermometer, Oral/Rect	115/230	12
6545-01-302-0228	Sink Unit, Surg Scrub	110	1200
6645-01-046-8848	Clock Wall, 1.5V DC	1.5	0.3
7910-01-177-0622	Cleaner, Vacuum, Elec	120/230	900

10. REPORTING ERRORS AND RECOMMENDING IMPROVEMENTS.


You can help improve this supply catalog. If you find any mistakes or know a way to improve the information, please let us know. Mail your letter or DA Form 2028 (Recommended Changes to Publication and Blank Forms) direct to:

Commander
US Army Medical Materiel Agency
ATTN: SGMMA-OCU
Frederick, MD 21702-5001

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.


Figure 16, Example - Section I, Introduction Page 15

6545-01-299-5619
NOT ILLUSTRATED



ILL. NO. 00001


6530-01-220-7186
NOT ILLUSTRATED



ILL. NO. 00002

6545-00-914-3510
NOT ILLUSTRATED

7910-01-177-0622
NOT ILLUSTRATED




ILL. NO. 00003

II-002

Figure 17, Example - Section II, Header Page/Component Illustrations

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.



SECTION II. ILLUSTRATED
COMPONENT LIST/HAND RECEIPT
FOR USE OF THIS HAND RECEIPT
SEE SECTION I.

FROM:

TO:

HAND RECEIPT NO:

END ITEM STOCK NUMBER	END ITEM DESCRIPTION	PUBLICATION NUMBER	PUBLICATION DATE	QUANTITY
6545-01-200-8610	BMS MINIMAL CARE DEPOT	SC 6545-8-MQ1	30 JUN 93	

ACT	STOCK NUMBER	ILL NO.	ITEM DESCRIPTION	RMKS	ARC CIIC UI	QTY AUT	QUANTITY ON HAND A I B I C I D I E
	6545-01-200-8610		MEDICAL MATERIEL SET MINIMAL CARE WARD; DEPMEDS	N	SE	1.00	
			M48055				
		L311					
N	6530-01-325-8590	00001	BED ADJ ATTENDANT OPERATED HOSP FOLDING W/O MATTRESS ALUM ALLOY	N	EA	20.00	
			64616 90-0748				
	6530-01-220-7185		CARRIER LITTER WHEELED ATTENDANT OPERATED U/W 6530 00 783 7805	N	EA	1.00	
			64616 5-864				
	6545-00-814-3510		CHEST MEDICAL INSTRUMENT & SUPPLY SET ALUMINUM MOB 30X18X20INCH 81349 MILC0018778	N	EA	7.00	
	6545-00-118-6248	00002	CHEST MEDICAL INSTRUMENT & SUPPLY SET 2 SHELVES 30X18X19 INCHES 98732 PENDING ASSIGNMEN	N	EA	1.00	
	7910-01-177-0672		CLEANER VACUUM ELECTRIC VERTICLE TANK PATBLE 120/230V 50HZ AC 64616 CM 4-198-P	N	EA	1.00	
	7105-00-710-0210	00003	TABLE FOLDING LEGS; LABORATORY 81349 MILT36072 U93814	N	EA	1.00	

11-003

Figure 18, Example - Section II, Component List/Hand Receipt

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.




Figure 19, Example - Section II, Component List/Hand Receipt(Last Page/Signature)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

SIC 6545-8-MG1

6545-01-302-0228
NOT ILLUSTRATED

III-002

Figure 20, Example - Section III, Header Page/ASIOE NSN Illustrations

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

SECTION III - ASSOCIATED SUPPORT ITEMS OF EQUIPMENT (ASIOE)

THE ITEMS LISTED BELOW ARE ASIOE END ITEMS DEDICATED TO THE OPERATION AND/OR MAINTENANCE OF THIS MEDICAL ASSEMBLAGE. ASIOE ARE SEPARATELY DOCUMENTED INTO TOE/THE ARMY AUTHORIZATION DOCUMENTS SYSTEM (TAADS) AND ARE NOT TO BE CONSIDERED COMPONENTS OF THIS MEDICAL ASSEMBLAGE. THEY ARE LISTED IN THIS SECTION FOR INFORMATION PURPOSES ONLY AND DO NOT CONSTITUTE AN ADDITIONAL AUTHORIZATION. TOTAL AUTHORIZATION IS REFLECTED ON THE UNIT'S MTOE.

END ITEM STOCK NUMBER	END ITEM DESCRIPTION	PUBLICATION NUMBER	PUBLICATION DATE
6545-01-299-5619	MMS MINIML CARE DEPYL	SC 6545-8-MG1	30 JUN 93

ACT	STOCK NUMBER	ILL NO.	ITEM DESCRIPTION	RMKS	ARC	CIIC	UI	QTY AUT
	6545-01-302-0228		SINK UNIT SURGICAL SCRUB AND UTENSIL HOSPITAL FIELD: 110V 60C AC 64616 9-0454 T60464	***2	N		EA	1.00

III-003

Figure 21, Example - Section III, Associated Support Item of Equipment Description

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

END ITEM STOCK NUMBER	END ITEM DESCRIPTION	PUBLICATION NUMBER	PUBLICATION DATE
6545-01-299-5619	MMS MINIML CARE DEPOT	SC 6545-8-MG1	30 JUN 93

	00.12.12	SYMPATHOMIMETIC ADRENERGIC AGENTS	
		ALBUTEROL INHALATION AEROSOL 17GM CONTAINER 200 METERED SPRAYS	6505-01-116-9245
	00.28.08	ANALGESICS AND ANTIPYRETICS	
		CODEINE PHOSPHATE AND ACETAMINOPHEN TABLETS USP 100 TABLETS/BT	6505-00-400-2054
	00.68.04	ADRENALS	
		METHYLPREDNISOLONE SODIUM SUCCINATE FOR INJECTION USP 125MG	6505-01-108-0809
04.00		ANTIHISTAMINE DRUGS	
		CHLORPHENIRAMINE MALEATE TABLETS USP 4MG 1000 TABLETS PER BOTTLE	6505-00-299-8610
		BROMPHENIRAMINE MALEATE AND PHENYLPROPANOLAMINE HCL TABS 500/BT	6505-01-206-5977
08.00		ANTI-INFECTIVE AGENTS	
08.12		ANTIBIOTICS	
	08.12.04	ANTIFUNGAL ANTIBIOTICS	
		GRISEOFULVIN TABLETS ULTRAMICROSIZE USP 125 MG 500S	6505-01-030-7982
08.16		ANTITUBERCULARS	
		ISONIAZID TABLETS USP 300 MG 100S	6505-00-132-6904
		RIFAMPIN CAPSULES USP 300MG 100 CAPSULES PER BOTTLE	6505-00-165-6575
08.20		PLASMODICIDES	
		CHLOROQUINE PHOSPHATE TABLETS USP 0.3GM 500 TABLETS PER BOTTLE	6505-00-117-6450
		PRIMAQUINE PHOSPHATE TABLETS USP 26.3 MG 1000 TABLETS PER BOTTLE	6505-00-299-8273
		SULFADOXINE AND PYRIMETHAMINE TABLETS I.S. 25 TABLETS/PACKAGE	6505-01-132-0257
08.24		SULFONAMIDES	
		SULFAMETHOXAZOLE AND TRIMETHOPRIM TABLETS USP 100 TABLETS/BOTTLE	6505-01-016-1470
08.32		ANTITRICHOMONAL AGENTS	
		METRONIDAZOLE TABLETS USP 250MG 250 TABLETS PER BOTTLE	6505-00-890-1840
12.00		AUTONOMIC DRUGS	
12.08		PARASYMPATHOLYTIC CHLORINERGIC BLOCKING AGENTS	
		BENZTROPINE MESYLATE TABLETS USP 2MG 100 TABLETS PER BOTTLE	6505-01-230-8726
12.12		SYMPATHOMIMETIC ADRENERGIC AGENTS	
		EPINEPHRINE INJECTION USP AQUEOUS CARTRIDGE NEEDLE UNIT 1ML 10S	6505-00-133-4449
		PSEUDOEPHEDRINE HYDROCHLORIDE TABLETS USP 60MG 1000 TABLETS/BT	6505-00-858-2366
		TERBUTALINE SULFATE INJECTION USP 1MG/ML 1ML AMPUL 10/PACKAGE	6505-01-013-9941
		TERBUTALINE SULFATE TABLETS USP 5MG 100 TABLETS PER BOTTLE	6505-01-039-2808

Source: https://assist.dla.mil -- Downloaded: 2026-02-26T04:01Z
Check the source to verify that this is the current version before use.

Figure 22, Example - Section IV, American Hospital Formulary Service Therapeutic Classification

END ITEM STOCK NUMBER	END ITEM DESCRIPTION	PUBLICATION NUMBER	PUBLICATION DATE
6545-01-299-5619	MMS MINIML CAME DEPOT	SC 6545-B-M01	30 JUN 93

NSN	ILL NO	PAGE NR
6505 01 318 1565	-	II-024
6510 00 054 7254	-	II-048
6510 00 111 0708	-	II-047
6510 00 116 1311	-	II-052
6510 00 200 7010	-	II-018
6510 00 200 7030	-	II-018
6510 00 201 4000	-	II-022
6510 00 203 5000	-	II-008
6510 00 559 6130	-	II-042
6510 00 582 7992	-	II-016
6510 00 582 7993	-	II-014
6510 00 721 9808	-	II-057
6510 00 724 1017	-	II-016
6510 00 775 5706	-	II-042
6510 00 782 2698	-	II-050
6510 00 782 2699	-	II-052
6510 00 782 2700	-	II-052
6510 00 786 3736	-	II-042
6510 00 889 7021	-	II-054
6510 00 913 7909	-	II-014
6510 00 926 8882	-	II-008
6510 00 935 5800	-	II-016
6510 00 935 5803	-	II-016
6510 00 935 5822	-	II-014
6510 00 935 5823	-	II-014
6510 01 003 7696	-	II-030
6510 01 003 7697	-	II-028
6510 01 050 8722	-	II-014
6510 01 107 7575	-	II-042
6510 01 153 4534	-	II-030
6510 01 271 7671	-	II-042
6515 00 228 7697	-	II-030
6515 00 734 4342	-	II-054
6515 00 828 2462	-	II-056
6515 00 905 1473	-	II-012
6515 00 935 1194	-	II-028
6515 01 030 0164	-	II-020
6515 01 039 4884	-	II-008
6515 01 149 8840	-	II-030
6515 01 149 8841	-	II-030
6515 01 149 8842	-	II-030
6515 01 149 8843	-	II-032
6515 01 187 6606	-	II-004
6515 01 207 8217	-	II-022
6515 01 289 5110	-	II-056
6530 00 075 6636	-	II-050
6530 00 890 1786	-	II-044
6530 00 900 2749	-	II-058
6530 01 047 9608	-	II-026
6530 01 119 0015	-	II-042
6530 01 166 3494	-	II-014
6530 01 166 3499	-	II-044
6530 01 166 9517	-	II-048
6530 01 169 0244	-	II-026
6530 01 220 7186	-	II-002
6530 01 270 1529	-	II-024
6532 01 269 9090	-	II-008
6545 00 914 3510	-	II-002
6545 01 287 2710	-	II-054
6545 01 299 5619	-	II-002
6545 01 302 0228	-	III-002
6645 01 046 8848	-	II-022
7105 00 269 8463	-	II-020
7210 00 119 7358	-	II-006
7210 00 139 5779	-	II-004
7210 00 299 8518	-	II-008
7210 00 299 8520	-	II-006
7240 00 089 3827	-	II-018
7910 01 177 0622	-	II-002
6530 01 325 6590	00001	II-002
6545 00 118 6248	00002	II-004
7105 00 710 0210	00003	II-004
6530 00 771 7025	00004	II-004
6532 00 081 1804	00005	II-004
4510 00 775 6000	00006	II-004
6530 00 115 2093	00007	II-004
6545 00 299 8649	00008	II-004
6515 00 550 7193	00009	II-006
7210 00 119 7356	00010	II-006
4320 00 299 8519	00011	II-006
6532 01 098 8345	00012	II-006
6530 00 794 4050	00013	II-008
6530 00 104 7631	00014	II-014
6530 00 836 8134	00015	II-016
6515 00 116 0531	00016	II-022
6530 01 151 1807	00017	II-026
6515 00 347 8600	00018	II-038
6515 00 754 2834	00019	II-040
6515 00 655 5751	00020	II-040
6515 01 003 2308	00021	II-040
6515 00 663 9801	00022	II-044
6515 01 113 7653	00023	II-048
6532 00 079 7904	00024	II-050
6532 00 079 7902	00025	II-050
6532 00 079 7899	00026	II-050
6532 00 079 7889	00027	II-050
6515 00 935 4088	00028	II-052
6515 00 149 1406	00029	II-058
6515 00 149 1407	00030	II-058
6630 01 119 8575	00031	II-058

V 002

Figure 23, Example - Section V, NSN/Illustration Number/Page Number Cross Reference

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

SECTION VI

NSNS DELETED FROM PREVIOUS PUBLICATION AND APPLICABLE REPLACEMENTS

END ITEM STOCK NUMBER	END ITEM DESCRIPTION	PUBLICATION NUMBER	PUBLICATION DATE
6545-01-299-5619	MMS MINIML CARE DEPOT	SC 6545 B MG1	30 JUN 91

DELETED NSN	REPLACEMENT NSN	REPLACEMENT NSN DESCRIPTION
6505-00-783-6821		
6505-00-890-1657		
6505-00-914-5297		
6505-00-926-8985		
6505-00-932-0682		
6505-01-079-6651		
6515-00-724-4603		
6530-00-226-3641		
6530-01-015-5189		
6532-00-009-2031		
6532-00-863-7484		
0000-00-000-0000		
5340-01-171-5263		
6135-00-935-2587		
6240-00-155-7898		
6505-00-139-4460		
6505-00-139-4548		
6505-00-216-5370		
6505-01-092-0415		

THE ACTIONS REFLECTED IN THIS SECTION ARE CUMULATIVE ACTIONS RESULTING FROM THE ANNUAL UPDATE PROCESS TO THIS SKO. THE DATA SHOWN ABOVE SHOULD BE USED TO BRING YOUR SET, KIT OR OUTFIT AND MANAGEMENT RECORDS UP-TO-DATE.

VI-001

Figure 24, Example - Section VI, NSNs Deleted From Previous Publication and Applicable Replacements

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

BY ORDER OF THE SECRETARY OF THE ARMY:

GORDON R. SULLIVAN
GENERAL, UNITED STATES ARMY
CHIEF OF STAFF

OFFICIAL:

MILTON J. HAMILTON
ADMINISTRATIVE ASSISTANT TO THE
SECRETARY OF THE ARMY

DISTRIBUTION:

TO BE DISTRIBUTED IN ACCORDANCE WITH DA FORM 12-21-E, BLOCK 2002, REQUIREMENTS FOR SC 6545-B-MG1.

Figure 25, Example - Authentication Page

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

INDEX

Abbreviations.....3.2.2.5
Accountability.....3.2.2.7
Acquisition requirements.....6.2
Act (action code) column.....3.2.3.8
Applicability.....1.2
Applicable documents.....2.0
Arc (accounting requirements code) column.....3.2.3.13
Associated support items of equipment (ASIOE).....3.2.2.6.1
Ciic (controlled item inventory code) column.....3.2.3.14
Component listings.....3.2.2.6.4
Concluding material.....3.2.8
Concluding material page numbering.....3.3.3
Content.....3.2
Contents/section III.....3.2.4.1
Contents/section IV.....3.2.5.1
Contents/section V.....3.2.6.1
Contents/section VI.....3.2.7.1
Cover.....3.2.1
Cover and numbering.....3.3.1
Data description:/section V.....3.2.6.2
Data description:/section VI.....3.2.7.2
Data requirements.....6.3
Definitions.....6.5
End item description block.....3.2.3.5
End item stock number block.....3.2.3.4
Examples/figures.....1.3
Explanation and use of columns (see section II).....3.2.2.4
From block (header page only).....3.2.3.1
General.....3.1
General.....3.2.2.1
Government documents.....2.1
Hand receipt number (header page only).....3.2.3.3
Horizontal format.....3.4
Illustration number column.....3.2.3.10
Intended use.....6.1
Item description column.....3.2.3.11
Last page/signature blocks.....3.2.3.18
Marking.....5.2
Non-government publications.....2.2
Notes.....6.0
Order of precedence.....2.3
Other government documents, drawings, and publications.....2.1.2
Packaging.....5.0
Packaging requirements.....5.1
Page numbering.....3.3

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

INDEX

Page size.....3.5
Power requirements.....3.2.2.9
Pub date block.....3.2.3.7
Pub number block.....3.2.3.6
Purpose.....3.2.2.2
Qty aut (quantity authorized) column.....3.2.3.16
Quality assurance provisions.....4.0
Quality assurance provisions.....4.2
Quantity on hand (A thru E) column.....3.2.3.17
Remarks code.....3.2.2.6.3
Reporting of errors and recommending improvements.....3.2.2.10
Requirements.....3.0
Responsibility.....3.2.2.3
Responsibility for compliance.....4.1.1
Responsibility for inspection.....4.1
Rmks (remarks) column.....3.2.3.12
Scope.....1.1
Section I, introduction.....3.2.2
Section I, II, III, IV, V, VI page numbering.....3.3.2
Section II, illustrated component list/hand receipt.....3.2.3
Section III, associated support items of equipment (ASIOE).....3.2.4
Section IV, american hospital formulary service thereapeutic classification.....3.2.5
Section V, NSN/Illustration number/page number cross reference....3.2.6
Section VI, NSNs deleted from previous publications and applicable replacements.....3.2.7
Set.....6.5.2
Sets, kits, outfits (SKO).....6.5.1
Special information.....3.2.2.6
Specifications and standards.....2.1.1
Standard holes.....3.6
Stock number column.....3.2.3.9
Subject term (keyword) listing.....6.6
Supply catalog.....6.5.3
Technical manual acquisition.....6.4
Therapeutic classification index.....3.2.2.6.2
To block (header page only).....3.2.3.2
Transportation information.....3.2.2.8
Ui (unit of issue) column.....3.2.3.15

Custodian:
Army - TM

Preparing Activity:
Army - TM

(Project No. A271)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER: MIL-C-49513(TM)	2. DOCUMENT DATE (YYMMDD):

3. DOCUMENT TITLE
Catalogs, Supply: Sets, Kits, and Outfits

4. NATURE OF CHANGE (Identify paragraph number and include rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)

8. PREPARING ACTIVITY

a. NAME USAMC Materiel Readiness Support Activity	b. TELEPHONE (Include Area Code) (1) Commercial: (606) 293-3469 (2) AUTOVON: 745-3469
ADDRESS (Include Zip Code) ATTN: AMXMD-EP Lexington, KY 40511-5101	IF YOU DO NOT RECEIVE A REPLY WITHIN 45, CONTACT: CALS Evaluation and Integration Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T04:01Z
Check the source to verify that this is the current version before use.