MIL-C-37059B
6 May 1975
SUPERSEDING
MIL-C-37059A
15 July 1974
and
MIL-C-37059
24 May 1974

MILITARY SPECIFICATION

CATHETER-SPECIMEN COLLECTION UNIT, ENDOTRACHEAL, PLASTIC, DISPOSABLE, LUKENS

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a sterile, disposable, specimen container of the Lukens design with catheter, suitable for use in obtaining mucus samples from the respiratory tract in the treatment of upper respiratory diseases.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on the date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water-Resistant, (For Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	Box, Fiberboard.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, Water Resistant (For Carton Sealing).

FSC 6515

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Military

MIL-S-36586	Sterilization Test Strip Set, Bacterial Spore.
MIL-S-37014	Suction Apparatus, Surgical

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center Drawing (DPSC)

No. 24342 CATHETER-SPECIMEN COLLECTING UNIT, ENDOTRACHEAL, PLASTIC, DISPOSABLE, LUKENS

(Copies of specifications, standards, drawings and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

D 412 Standard Method of Tension Testing of Vulcanized Rubber.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa. 19103).

UNITED STATES DEPARTMENT OF HEALTH, EDUCATION AND WELFARE: FOOD AND DRUG ADMINISTRATION

Federal Food, Drug and Cosmetic Act and regulations promulgated thereunder.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the Division of Public Information, Food and Drug Administration 5600 Fishers Lane, Rockville, Md. 20852.)

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (USP).

(Application for copies should be addressed to the Mack Publishing Company, Easton, Pa. 18042.)

3. REQUIREMENTS

3.1 Component parts. The unit shall consist of the following components:

a - container
b - catheter tube
c - suction tube
d - top cap
e - extra cap
f - label.

3.2 Material.

3.2.1 Catheter. The catheter tubing shall be fabricated from clear, flexible, polyvinyl chloride plastic and shall have a matte finish. Shall meet the requirements of the Pharmacopeia of the United States (USP) for class IV plastics.

3.2.2 Suction tubing. The suction tubing shall be fabricated from clear, flexible polyvinyl chloride plastic.

3.2.3 Specimen container. The specimen container shall be fabricated from rigid polystyrene plastic.

3.2.4 Caps. The top cap shall be fabricated from white polyethylene. The extra cap shall be fabricated from natural polyethylene.

3.2.5 Label. Shall be a white, paper, self-adhering pressure sensitive label with a peel off backing. Shall have black printing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

* 3.3 Style, design and dimensions. Shall be a graduated plastic specimen container covered by a white snap-on plastic cap at the top end. The white plastic cap shall contain two circular holes. Into one hole is inserted a plastic, 14 French, whistle tip catheter. Into the other hole is inserted a plastic suction tube, dilated (elongated bubble) at the proximal end to permit rapid attachment to standard suction devices. At the bottom of the container is attached a removable plastic cap for closing the container after specimen collection. The label, approximately 1-3/8 by 3 inches, shall be printed to permit inscription by the user of name, date, room, and doctor. Shall be compatible with standard suction devices such as indicated in MIL-S-37014, for SUCTION APPARATUS, SURGICAL, High Volume, Low Pressure, Mobile; 115 Volt, 60 Hz, AC. Shall be in accordance with DPSC Drawing 24342.

3.4 Sterility. The catheter-specimen collecting units shall be sterile when tested with a biological indicator as specified in 4.4.1.

3.4.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated product).
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.4.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each individual and unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

* 3.5 Breaking strength. The catheter tube shall have a minimum (min.) breaking strength of 10 lbs., (20 inches, min.) and the suction tube shall have a minimum breaking strength of 10 lbs., (20 inches, min.) when tested as specified in 4.4.2.

3.6 Ultimate elongation. The catheter tube and suction tube shall have ultimate elongations of 140%, minimum and 175%, minimum respectively, when tested as specified in 4.4.3.

* 3.7 Freedom from foreign matter. Unit shall be free from foreign matter in vial and lumen of tubes when tested as specified in 4.4.5.

3.8 Identification marking. The specimen container shall be legibly and permanently marked with the manufacturer's name or trademark.

3.9 Workmanship. Workmanship shall be first class throughout. Catheter-specimen collecting units shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the plastics, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as indicated in Table I. The unit of product for examination shall be one catheter-specimen collecting unit.

Table I. Sampling Criteria for Examination

Type of Defect	Inspection level*	AQL (percent defective)
Major	S-2	1.0
Minor	S-4	2.5

*If acceptance level is not exceeded, lot shall be accepted. If rejection number is equalled or exceeded, then the inspection at level II may be performed to determine the final rejection or acceptance.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

* 4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table II.

Table II. Sampling Criteria for Tests

Component (unit of product)	Characteristic	Requirement	Test procedure	Inspection level	AQL
1 Catheter and Suction tubing	Breaking strength	3.5	4.4.2	S-2	1.5
1 Catheter and Suction tubing	Ultimate elongation	3.6	4.4.3	S-2	1.5
1 Catheter	Class IV plastic	3.2.1	4.4.4	USP	2.5
1 End item	Sterility	3.4	4.4.1	4.4.1	0
1 End item	Freedom from foreign matter	3.7	4.4.5	S-2	4.0

4.3 Examination

* 4.3.1 The catheter-specimen collecting units shall be examined to determine compliance with all the requirements contained in this specification. Examination shall be conducted as specified in Tables I and III and in accordance with MIL-STD-105. Examination, however, shall not be restricted to defects listed in Table III. Any units in the sample containing one or more defects shall be rejected (defective) and if the number of defective units in any sample exceeds the acceptance number for that sample, the lot represented by that sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Major	
101	Surface of tubes is tacky.
102	Outside surfaces of catheter tube not smooth and free from burrs or surface imperfections.
103	Hole in catheter tube at distal end not free from burrs, slivers, loose particles or other foreign matter.
104	Tube not free from kinks as indicated by a decrease in the cross sectional area of the lumen of the tubes.
105	Specimen collecting container cracked or broken.
106	Absence of dilatation (elongated bubble) at proximal end of tube as a means of attachment to standard suction devices such as in MIL-S-37014.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects (Cont'd.)

Categories	Defects
Minor	
201	Identification marking not clearly printed.
202	Self adhesive paper label not clearly and legibly printed.
203	Capacity of container less than 20 cc.
204	Extra cap does not make a leak proof fit with container.

* 4.3.2 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table IV. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Table IV. Examination for Preparation for Delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:52Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Sterility test. The catheter-specimen collecting unit shall be tested for sterility in accordance with the method specified in the USP for transfusion and infusion assemblies using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.1.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II, Sampling Criteria for Tests.

4.4.1.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C.
Dry Heat	B. stearothermophilus, or	55° - 65°C.
	B. subtilis var. niger (or globigii)	30° - 35°C.
Gas	B. subtilis var. niger (or globigii)	30° - 35°C.
Radiation	B. pumilus	30° - 35°C.

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.1.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.1 Sterility test and 4.4.1.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.1.1. NOTE: When any inoculated carrier is chosen, product samples must also be tested as required by the USP.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

4.4.1.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Request for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.1.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.2 Breaking Strength. The breaking strength of the catheter and suction tubing shall be determined in accordance with American Society for Testing and Materials (ASTM) Method D-412.

4.4.3 Ultimate elongation. The ultimate elongation of the catheter and suction tubing shall be determined in accordance with method ASTM D-412.

4.4.4 Class IV Plastics. The plastic components of the unit shall be tested for suitability as required by the USP for Class IV plastics.

4.4.5 Freedom from foreign matter. Inside of vial and lumen of tubes shall be viewed from a distance of 14 inches without magnification. Illumination of units shall be twin, 15 watt, cool white fluorescent tubes.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Individual package. Each collecting unit shall be packaged in a sealed peel open type bag of appropriate size and design and be capable of maintaining sterility of contents unless damaged or opened.

* 5.1.1.2 Unit package. Fifty individually packaged collecting units shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

* 5.2.1 Level A. Four unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification.

* 5.2.2 Level B. Four unit packages shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers at the lowest rate to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Individual package. In addition to any commercial markings, each individual package shall be marked in accordance with the requirements of the Federal Food, Drug and Cosmetic Act. The sterility lot (control) number and the legend "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement shall be shown.

5.3.2 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. Sterility lot (control) number, contract or purchase order number and name of contractor shall be shown. The method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat Sterilization) shall be shown following the level of packaging and date of packaging. Example: "A - 5/74 - G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129. Sterility lot (control) number and method of sterilization as specified in 5.3.2 shall be shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

6. NOTES

6.1 Intended use. The catheter-specimen collecting units are intended for use in obtaining mucus samples from the respiratory tract in the treatment of upper respiratory diseases.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6515-00-518-3036	CATHETER-SPECIMEN COLLECTING UNIT, ENDOTRACHEAL, Plastic, Disposable Lukens, 50s.

6.4 Limitations. This specification does not cover all types, classes, grades, styles, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover those items which are normally procured to meet Military requirements.

6.5 The margins of this specification are marked with an asterisk to indicate where changes (additions, modifications, corrections, deletions) from the previous issue were made. This was done as a convenience only and the Government assumes no liability whatsoever for any inaccuracies in these notations. Bidders and contractors are cautioned to evaluate the requirements of this document based on the entire content irrespective of the marginal notations and relationship to the previous issue.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1926

is

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at *reasonable cost* and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE: MIL-C-37059B CATHETER-SPECIMEN COLLECTING UNIT, ENDOTRACHEAL, PLASTIC, DISPOSABLE,

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER LUKENS

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. *RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES*

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:52Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:52Z
Check the source to verify that this is the current version before use.