MILITARY SPECIFICATION

CANNULA, UTERINE, CORROSION-RESISTING STEEL

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a uterine cannula suitable for use in gynecological procedures.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

GG-I-526	Instruments, Dental and Surgical; General Specifications.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

Defense Personnel Support Center (DPSC)

No. 20523	Tip, Uterine, Cannula

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids ore request for proposals shall apply:

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

ANSI Z70.1 (UDC 615.47) American Standard Dimensions of Glass and Metal Luer Tapers for Medical Applications.

(Application for copies should be addressed to the American National Standards Institute (ANSI, 1430 Broadway, New York, N. Y. 10018.)

3. REQUIREMENTS

3.1 General Requirements. The cannulas shall be in accordance with the applicable requirements of GG-I-526, and as specified herein.

3.2 Material. Shall be in accordance with GG-I-526b, as indicated below:

Part Description	Class	Type
Tube of cannula	3	302*, 303*, 304
Remainder of cannula	3 or 8	302*, 303*, 304; 260, 270, 360

*Nickel content may be 8.0% to 12.0%

3.3 Style, design, and dimensions. Shall be a uterine cannula. The proximal end shall be suitable for use with NSN 6515-00-890-1103, Insufflator, Uterotubal, Carbon Dioxide, Automatic Pressure Control and NSN 6515-00-890-1104, Opaque Oil Attachment, Uterotubal Insufflator. Each cannula shall be supplied with five (5) disposable plastic tips (one (1) large and four (4) small) in accordance with Defense Medical Purchase Description for National Stock Number 6515-00-082-2457, Tip Assortment, Uterine Cannula, Plastic, Disposable. Style, design, and dimensions of the cannula shall be in accordance with Defense Personnel Support Center Drawing 20523.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:53Z
Check the source to verify that this is the currentversion before use.

3.3.1 Stopcock. Shall be a one-way stopcock with a female Luer-lock connector at the proximal end and a tube and finger rings attached to opposite end as shown in the drawing. The female Luer-lock connector as shown in the drawing shall be in accordance with Figure 3 of ANSI Standard Z70.1. This Luer connector shall be tested for diameter and depth of cavity as specified in 4.4.1. Stopcock Luer connectors which do not fit the plug gage within the "GO" - "NO GO" limits specified in 4.4.1, shall be considered not acceptable. In addition, when the stopcock Luer connector is fitted onto the plug gage as specified in 4.4.1, there shall be no movement (wobble) of the connector on the gage when the stopcock is pressed lightly from side to side in all directions. The entire Luer-lock connector cavity shall be well plated down to the entrance of the stopcock lumen. The stopcock shall have a single straight bore, and shall be provided with a tapered plug with an open position (bore open to fluid flow) and a closed position. The plug shall have a suitable lever for turning to the different positions.

3.3.1.1 Freedom from foreign matter. When tested as specified in 4.4.2, the Luer cavity shall be substantially free from chips and other foreign matter. In addition, a visual examination of the bore of the stopcock shall reveal no presence of foreign matter, chips, burrs, or occlusion of the lumen.

3.3.2 Fluid tightness. When tested as specified in 4.4.3, the entire cannula shall not show leakage at any point exceeding the formation of a falling drop.

3.4 Finish. Finish shall be mirror.

3.5 Identification marking. The cannula shall be permanently and legibly marked with the name of the manufacturer or registered trademark.

3.6 Hardness. Types 302, 303 and 304 shall have a Rockwell hardness of B75 to B85.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the metals, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I. The unit of product for examination purposes shall be one cannula.

Table I. Sampling for Examination

Type of Defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II. Where feasible, the same sample shall be used for the determination of two or more test characteristics. Unit of product, for test purposes, shall be one cannula.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:53Z
Check the source to verify that this is the current version before use.

Table II. Sampling for Tests

Component	Characteristic	Requirement	Test procedure	Insp level	AQL (percent defective
Cannula	Plating	GG-I-526	GG-I-526b	S-2	**
	Salt-spray test	GG-I-526	GG-I-526	S-2	**
	Boil test	GG-I-526	GG-I-526	S-2	**
	Hardness	3.6	GG-I-526	S-2	**
	Diameter and depth	3.3.1	4.4.1	II	1.0
	Freedom from foreign matter	3.3.1.1	4.4.2	S-4	1.5
	Fluid tightness	3.3.2	4.4.3	S-4	1.5
	Dimensions	3.3	--	S-2	4.0
	Composition analysis	3.2	GG-I-526	*	**

* Sufficient amount from each lot of each metal used for the construction of the cannulas shall be selected for composition analysis.
** Acceptance number shall be zero (0) for all sample sizes.

4.3 Examination. The cannulas shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with MIL-STD-105 and the following classification of defects. Examination shall include, but shall not be limited to the defects listed in Table III.

Table III. Classification of Defects

Categories	Defect
Critical	None defined.
Major	
101	Style and design not as specified.
102	Instrument not constructed as specified.
103	Instrument not free from sharp edges, rough areas, burrs, cracks, grind marks, or overlaps.
104	Stopcock does not operate smoothly.
Minor	
201	Finish not as specified.
202	Instrument not permanently and legibly marked with either the name or trademark of the contractor.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Diameter and depth of Luer cavity. The Luer cavity shall be tested with the male Luer taper gage of Figure 5 of ANSI Standard Z70.1. (This gage shall be as shown in the ANSI Standard, except that it may be constructed with a suitable fluid path to permit performance of the fluid tightness tests). The stopcock, shall be slipped onto this gage with very light pressure. The fit of the stopcock, on this gage shall be tested by means of the feeler gage in accordance with Figures 5 and 6 of ANSI Standard Z70.1.

4.4.2 Freedom from foreign matter test. The Luer cavity shall be examined under 7X magnification, and good illumination, for chips and other foreign matter. In addition, a cotton swab tightly wrapped on the end of a small wooden applicator shall be vigorously turned down into the bottom of cavity, withdrawn, and examined for chips and other foreign matter under 7X magnification and good illumination. The tube and stopcock shall be tested for blockage by inserting a wire through the lumen of the tube and the bore of the stopcock.

4.4.3 Fluid tightness. The female Luer connector of the stopcock shall be attached by means of a plug gage of Figure 5, of ANSI Standard Z70.1 (except that gage shall have a fluid path) to a water pressure source having a pressure gage. The water shall be run through the stopcock and tube so as to fill any cavity space and the lumen of the tube. The lumen of the tube shall be sealed by a piece of solid rubber slipped onto the point, and a pressure of 15 pounds per square inch applied for 10 seconds.

4.4.4 Hardness. Shall be determined as specified in GG-I-526b.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. Each cannula, as specified, constitutes one unit of issue.

5.1.2 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

5.3 Marking. Marking shall be as specified in MIL-STD-129.

6. NOTES

6.1 Intended use. The cannulas are intended for for use in the treatment of tubal occlusions.

6.2 Ordering data. Purchasers should exercise any desired options offered herein, and the procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6515-00-890-1105	CANNULA, UTERINE, Corrosion-Resisting Steel.

It does not include all types or sizes as indicated by the title nor all types which are commercially available.

6.4 Supersession data. This specification includes requirements of Purchase Description No. 3, dated 17 March 1969 for FSN 6515-890-1105 CANNULA, UTERINE, Corrosion-Resisting Steel.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1845

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304



OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37065

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:53Z
Check the source to verify that this is the current version before use.