MILITARY SPECIFICATION

CASE, MEDICAL INSTRUMENT AND SUPPLY SET NYLON, NONRIGID, NO. 6

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers an empty nylon case intended to hold instruments and supplies listed hereinafter, and be suitable for use in minor surgical procedures in the field.

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on the date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

V-T-285	Thread, Polyester
DDD-L-20	Label; for Clothing, Equipage, and Tentage, (General Use)
PPP-B-566	Boxes, Folding, Paperboard
PPP-B-585	Boxes, Wood, Wirebound
PPP-B-601	Boxes, Cleated-Plywood
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner
PPP-B-636	Boxes, Shipping, Fiberboard
PPP-B-676	Boxes, Setup

Military

MIL-T-5038	Tape, Textile and Webbing, Textile, Reinforcing, Nylon
MIL-F-21840	Fastener Tapes, Hook & Pile, Synthetic
MIL-C-43128	Cloth, Plain Weave, Nylon, Water Repellent, OG-106

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

FED-STD-751	Stitches, Seams, and Stitching

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes
MIL-STD-129	Marking for Shipment and Storage
MS 16686	Medical Services Symbol

DRAWING

Defense Personnel Support Center (DPSC)

10091	Case, Medical Instrument and Supply Set, Nylon, Nonrigid, No. 6

2.2 Other publication. The following document forms a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids shall apply.

THE COLOR ASSOCIATION OF THE UNITED STATES, INC.

DEPARTMENT OF DEFENSE
Standard Shades for Sewing Threads

(Application for copies should be addressed to the Color Association of the United States, Inc., 200 Madison Avenue, New York, N. Y. 10016.)

3. REQUIREMENTS

3.1 Component parts. The case shall consist of a double layer body piece, four flaps, two pockets, two series of loops, and hook and pile tape closures.

3.2 Materials.

3.2.1 Basic material. The basic material shall be cloth, plain weave, nylon, water repellent, Olive Green 106 conforming to MIL-C-43128.

3.2.2 Tape, nylon. The tape used for binding and forming the loops shall be 3/4 inch, Olive Drab 7 nylon tape conforming to type III of MIL-T-5038.

3.2.3 Thread polyester. The thread for all stitching shall conform to type I, class I, sub-class B of V-T-285. The thread shall be size F, dyed OG-106, except the colorfastness requirements and test methods shall be limited to laundering and weathering only.

3.2.4 Fasteners. The nylon tape fastener consisting of a hook and pile tape shall conform to type I, class 1, one inch width for the hook tape and two inch width for the pile tape of MIL-F-21840. The color shall be OG-106. The colorfastness requirements and test methods shall be limited to laundering only.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

3.3 Construction. Construction shall be as shown on Drawing 10091 and as specified herein. Each body piece, plus flap nos. 2 and 4, shall be cut in one piece and consist of two layers (thickness) of nylon cloth specified in 3.2.1. Flap nos. 1 and 3 shall be cut from a single layer (thickness) of cloth as specified in 3.2.1. On the inner surface of flap nos. 2 and 4 a pocket shall be constructed measuring approximately 3-1/4 inches in width and extending the length of the flap. The pocket shall be formed from a single layer (thickness) of the cloth specified in 3.2.1. Two lengths of retaining loops, made of nylon tape specified in 3.2.2, shall be stitched (bartacked) securely to the inner layer (one thickness) of the body piece, to form 9 loops on each side of the case. The retaining loop tape shall be located 2 inches in from each side of the case. The loops shall be suitable for accepting and retaining the instruments shown in table I. The tape between each pair of loops shall be bartacked, and the tape ends shall be turned under and caught by the extreme bartacks. All bartacks shall be 5/8 inch long and shall contain a minimum of 28 stitches. After the loops have been stitched to one layer of nylon cloth, this layer shall be stitched to the other layer, and the pockets and flaps attached. All raw edges shall be bound with 3/4 inch tape specified in 3.2.2, using seam type BSa-1.

3.3.1 Stitches, seams, and stitchings. Stitch, seam, and stitching types shall conform to FED-STD-751. Unless otherwise specified there shall be not less than 8 stitches per inch, using stitch type 301. All ends of stitching shall be securely back-stitched for a length of not less than 3/4 inch, except where anchorage is provided by another seam or stitching.

3.3.2 Attach fasteners. Attach nylon tape fasteners specified in 3.2.4 in the location shown on Drawing 10091. The hook and pile tape shall be aligned so that after folding, the case may be neatly and securely closed.

3.3.3 Illustrated interior. The inner surface of the body piece shall present a series of legible and permanent outlines of the items which the case is intended to accommodate (Item Nos. 1, 2, 3, 4, 5, 8, 9, and 14 of Table I), as shown on Drawing 10091. The outline marking shall be in accordance with the type IV method specified in DDD-L-20.

3.4 Intended contents. Each finished case shall be suitable for holding the items shown in table I, in the quantities shown.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Table I. List of contents - Surgical Instrument Set, Minor Surgery

Item No.	National Stock No.	Item Identification	Unit	Quantity
1	6515-00-299-8736	Holder, Suture Needle, Hegar-Mayo, 6 inch.	Each	1
2	6515-00-333-3600	Forceps, Dressing, Straight, 5-1/2 inch.	Each	1
3	6515-00-334-6800	Forceps, Hemostatic, Straight, Kelly 5-1/2 inch.	Each	2
4	6515-00-337-9900	Forceps, Tissue, Tweezer Type, Straight, 5-1/2 inch.	Each	1
5	6515-00-344-7800	Handle, Surgical Knife, Detachable Blade, No. 3.	Each	1
6	6515-00-352-4500	Needle, Suture, Surgeons, Regular, 3/8 Circle, Cutting Edge, Size 12, 6s.	Pkg.	1
7	6515-00-352-4540	Needle, Suture, Surgeons, Regular, 3/8 Circle, Cutting Edge, Size 16, 6s.	Pkg.	1
8	6515-00-356-9100	Probe, General Operating, 5 inch.	Each	1
9	6515-00-365-1820	Scissors, General Surgical, Straight, 5-1/2 inch.	Each	1
10	6515-00-531-7884	Suture, Nonabsorbable, Surgical, Silk, Braided, Size 0, 3s.	Pkg.	2
11	6515-00-660-0010	Blade, Surgical Knife, Detachable, Carbon Steel, No. 11, 6s.	Pkg.	2
12	6515-00-660-0011	Blade, Surgical Knife, Detachable, Carbon Steel, No. 10, 6s.	Pkg.	2
13	6515-00-754-2812	Suture, Nonabsorbable, Surgical Silk, Braided, Size 00, 17 Strands, 12s.	Pkg.	6/12
14	6515-00-812-7000	Thermometer, Clinical, Veterinary.	Each	1

3.5 Marking for identification. Each case shall be marked by stenciling or printing with black waterproof paint or ink as shown on Drawing 10091. Symbol shall be in accordance with type I of MS16686.

3.6 Workmanship. Workmanship shall be first class throughout. Complete cases shall be free from defects which detract from their appearance or may impair their suitability for their intended use. The cases shall conform to the quality of product established by this specification. The occurrence of defects shall not exceed the limits set by the applicable acceptable quality levels.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the nylon cloth, webbing, tape, thread, hardware, and fasteners, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables II and III (see 4.2.1 and 4.2.2) and 4.3.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in table II. Unit of product for examination purposes shall be one case.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Table II. Sampling for Visual Defects

	Inspection level	AQL (defects per 100 units)
Major	II	2.5
Total (Major and Minor combined)	II	6.5

4.2.2 Examination. The case shall be examined for defects including, but not limited to, those listed in table III.

Table III. Classification of Defects

Major

101	Any component part missing.
102	Any required operation omitted or improperly performed.
103	Any hole, cut, tear, mend, patch or unauthorized splice.
104	Needle chews that may develop into a hole.
105	Any weakening defect such as smash, multiple float, or loose slub, that may develop into a hole.
106	One or more spots or stains of an obviously permanent nature.
107	Any seam twisted, puckered, affecting appearance or serviceability.
108	Any open seam - 1/4 inch or more.
109	Raw edges.
110	Tape fastener improperly aligned.
111	Tape fastener insecurely attached.
112	Markings missing, incorrect, or illegible.

Minor

201	One or more removable spots or stains clearly noticeable.
202	Thread ends not trimmed.
203	Any part of case pleated, twisted, or caught in any unrelated row of stitching.
204	Seam type other than specified.
205	Stitch type other than specified.
206	Less than 8 stitches per inch.
207	Broken or skipped stitches, not amounting to an open seam, occurring twice or more in the same row of stitching. Open seam less than 1/4 inch.
208	Ends of stitching backstitched less than 3/4 inch.
209	Outlines of components not properly illustrated as specified in Drawing.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

4.2.2.1 Examination for dimensional defects. During this examination, any measurement deviating from the normal dimensions and tolerances specified in the drawing shall be scored as a size measurement defect. The sample unit for this examination shall be one completely fabricated case. The lot size shall be expressed in units of one case. The sample size shall be selected in accordance with inspection level S-3 of MIL-STD-105. The acceptable quality level shall be 4.0 (one class) defects per 100 units.

4.3 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging. Packaging shall be Level A or C as specified (see 6.2).

5.1.1 Unit of issue. One case, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
Not applicable	12 units	288 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units shall be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit preparation. Each case shall be neatly folded and secured in the closed position as illustrated on Drawing.

5.1.3.2 Intermediate package. Intermediate package shall be a box of appropriate size constructed in accordance with PPP-B-566 or PPP-B-676, except that commercial colors shall be acceptable. Closure shall be adequate to prevent spilling of contents under normal handling.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing. Packing shall be Level A, B or C as specified (see 6.2).

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant, grade V3c. Special requirements of applicable table shall apply for PPP-B-636 box.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure of wood boxes shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II of PPP-B-636. Special requirements of applicable table shall apply for PPP-B-636 box.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Interior packages. Each intermediate package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Intended use. The case covered by this specification is a component of National Stock No. 6545-00-957-7650, Surgical Instrument Set, Minor Surgery, and is used as the container for the items listed in 3.4.

6.2 Ordering data. Procurement documents should specify the following:

(a) Title, number, and date of this specification.
(b) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

NSN	Item Identification
6545-00-913-6525	CASE, MEDICAL INSTRUMENT AND SUPPLY SET, Nylon, Nonrigid, No. 6

6.4 This specification does not include all types, classes, grades, sizes, etc., of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the types, etc., which are normally procured to meet military requirements.

6.5 Supersession data. This specification includes the requirements of Defense Medical Purchase Description No. 6, dated 12 January 1971 and Modification No. 1, dated 31 January 1973.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6545-0015

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1975-603-117/4518

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-C-37066

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.