MILITARY SPECIFICATION

GUIDES, CIRCUMCISION, PLASTIC, DISPOSABLE

This specification has been approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers two types of sterile, disposable, plastic circumcision guides for use on male infants.

1.2 Classification. The circumcision guides shall consist of the following types and sizes (see 6.2):

Type I	Size (in centimeters)
Regular	1.3

Type II	Size (in centimeters)
Very small	1.1
Small	1.2
Regular	1.3
Large	1.5
Very large	1.7

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposals, form a part of the specification to the extent specified herein:

SPECIFICATIONS

Federal

MMM-A-250	Adhesive, Water Resistant (for Sealing Fiberboard Boxes).
PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.
PPP-B-676	Box, Set-Up.
PPP-T-76	Tape, Pressure-Sensitive Adhesive Paper, (for Carton Sealing).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Military

MIL-L-10547	Liners, Case and Sheet overwrap, Water-Vaporproof, Waterproof, Flexible.
MIL-S-36586	Sterilization Strip Test, Bacterial Spore.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

Drawing No. 23246 Guide Circumcision, Plastic.

(Copies of specifications, standards, and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply:

THE UNITED STATES PHARMACOPEIAL CONVENTION, INC.

The Pharmacopeia of the United States of America (U.S.P.)

(Application for copies shall be addressed to the Mack Publishing Company, Easton, Pennsylvania 18042).

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and General Regulations for its Enforcement.

(Application for copies should be addressed to the Food and Drug Administration, U.S. Department of Health, Education and Welfare, 5600 Fishers Lane, Rockville, Maryland 20852.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

1. Records of examinations and tests performed by or for the contractor 4.1

3. REQUIREMENTS

3.1 Material.

3.1.1 Plastic guide. The plastic guide shall be fabricated of clear butyrate plastic (Eastman #460E-37038M).

3.1.2 Ligature. The ligature shall be made of linen.

3.2 Style, design and construction. Shall be shaped generally like a spherical sector to fit over the glans penis and under the prepuce. A break-away handle shall be attached at opposite sides of the short diameter. A shoulder shall be present on the exterior near the short diameter so that the ligature is maintained in position over the prepuce. Dimensions shall be as shown for the 1.3 cm size in Defense Personnel Support Center (DPSC) Drawing No. 23246. The linen ligature shall be 12 to 13 inches long.

3.3 Sterility. The circumcision guide and linen ligature shall be sterile when tested with a biological indicator as specified in 4.4.1.

3.3.1 Certificate of sterility. A certificate of sterility, stating that samples from each sterilizer lot have been tested and found to be sterile, shall be submitted to the procuring activity. The certificate of sterility shall include:

a. Item identification.
b. National stock number.
c. Contract or purchase order number.
d. Sterility lot or control number.
e. Date of sterilization.
f. Quantity of product in sterilizer lot.
g. Type of sterilization (gas, radiation, heat, steam, etc.).
h. Form of biological indicator (i.e., inoculated paper spore strip or carrier, inoculated product, or inoculated simulated
i. Test organism(s) and population density used.
j. Test medium, incubation temperature, incubation time, and dates of tests.
k. Number of test units.
l. Statement that samples comply with sterility tests.
m. Signature of head of laboratory or a designated qualified alternate.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

When inspection is made at destination, the certificate of sterility shall be submitted by the contractor to the Directorate of Medical Materiel, ATTN: DPSC-ATQ, Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, Pa., 19101, and a copy forwarded with each shipment to the consignee. When inspection is made at source, certificate of sterility shall be furnished to the cognizant Government Quality Assurance Representative for submission to the procuring activity.

3.3.2 Sterility lot. A sterility lot is that single quantity of product subjected to the same manufacturing operation and simultaneously sterilized in the same sterilizer chamber. Each unit package shall be clearly marked with a sterility lot (control) number and date of sterilization as specified in 5.3.1. Each sterility lot shall be traceable to the pertinent manufacturing lot(s).

3.4 Animal implantation. The plastic guide and linen ligature shall show no adverse tissue reaction upon gross and microscopic examination, when tested as specified in 4.4.2.

3.4.1 Certificate of animal implantation test. A certificate listing all lot or control numbers shall state that representative samples from each lot or control have been tested and have passed the animal implantation test. This certificate shall include the specific tests performed and the test results. The record of the animal implantation test and its results shall be certified by the pathologist who performed or supervised performance of the animal implantation test. For destination inspection, certificates shall be submitted by the contractor directly to Defense Personnel Support Center, Directorate of Medical Materiel, Technical Operations Division, 2800 South 20th Street, Philadelphia, PA 19101, Attention: Quality Assurance Branch, and one copy of each shall be forwarded with each shipment to the consignee. For source inspection, certificates shall be furnished to the procuring activity via the government inspector.

3.5 Packaging. Each unit package for Type I shall contain 125 circumcision guides. Each unit package for Type II shall contain 128 circumcision guides of the following sizes and quantities:

Assorted Sizes (in centimeters)		Quantity
Very small	1.1	8
Small	1.2	28
Regular	1.3	82
Large	1.5	8
Very large	1.7	2

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:53Z
Check the source to verify that this is the current version before use.

3.6 Instructions. Each unit package shall be supplied with 125 or 128 copies of special information and instructions for parents. The information should include as a minimum, the following:

"a. The plastic rim usually drops off in 5 to 8 days following circumcision.
b. No special dressing is required and the baby can be bathed just as though he had not been circumcised.
c. Notify your doctor if healing does not proceed as described, if any unusual swelling is observed, or if the plastic ring is not off within 10 to 12 days."

In addition, each box shall contain instructions for the person performing the circumcision. These instructions shall include anatomical illustrations and correct methods of insertion, ligation, and excision.

3.7 Workmanship. Workmanship shall be first class throughout. The complete unit shall be free from defects which detract from its appearance or impair its serviceability. The complete unit shall show smooth outer and inner surfaces and be free from burrs, chips, soils, slivers, or other foreign matter.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 20260-21T09:53Z
Check the source to verify that this is the current version before use.

4.2 Sampling. The unit of product shall be, for examination purposes, a plastic circumcision guide with the accompanying linen ligature.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels shall be as indicated in Table I.

Table I. Sampling for Examination

Type of defect	Inspection level	AQL (percent defective)*
Major	S-2	1.0
Minor	S-1	1.5

* If acceptance number is not exceeded, lot shall be accepted. If rejection number is equalled or exceeded, inspection at Level II may be performed to determine final rejection or acceptance.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in Table II below:

Table II. Sampling Criteria for Tests

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Unit of product	Sterility	3.3	4.4.1	S-2	*
	Animal implantation	3.4	4.4.2	S-1	*

* Acceptance number shall be zero for all sample sizes.

4.2.3 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination. The circumcision guide shall be examined to determine compliance with all requirements contained in this specification. Examination shall be conducted as specified in Tables I and III. Examination, however, shall not be restricted to defects listed in Table III. Any circumcision guide in the sample containing one or more defects shall be rejected and if the number of defective circumcision guides in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

Table III. Classification of Defects

Categories	Defects
Major	
101	Linen ligature not present.
102	Burrs or molding seams present at contact surface.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Table III. Classification of Defects (cont'd)

Categories	Defects
Minor	
201	Plastic guide and linen ligature not in inner plastic envelopes.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Sterility test. The circumcision guide and linen ligature shall be tested for sterility in accordance with the method specified in the USP for nonparenteral devices using inoculated product, inoculated simulated product, or paper spore strip (carrier). Product containing a biological indicator shall be marked or labeled so that it clearly differentiates from product intended for distribution. The incubation medium and minimum time in each case shall be soybean casein digest for 7 days for inoculated product, inoculated simulated product, or paper spore strip (carrier). The incubation temperature shall be as indicated in paragraph 4.4.1.1. The minimum number of test units, regardless of lot size, shall be 10, except for inoculated carrier which shall be 20 for Type I Sterilization and 30 for Type II Sterilization including inoculated carrier and product test units. The inspection level shall be as indicated in Table II, Sampling Criteria for Tests.

4.4.1.1 Inoculated product or simulated product. For each type of sterilization, the organism used shall be as listed below:

Type of Sterilization	Organism	Incubation Temperature
Steam	B. stearothermophilus	55° - 65°C
Dry Heat	B. stearothermophilus, or	55° - 65°C
	B. subtilis var. niger (or globigii)	30° - 35°C
Gas	B. subtilis var. niger (or globigii)	30° - 35°C
Radiation	B. pumilus	30° - 35°C

The number of viable cells shall be sufficient to provide assurance that the product has been sterilized.

4.4.1.2 Inoculated carrier. Inoculated carriers (paper spore strips) if used shall be in accordance with MIL-S-36586. Other paper carriers shall be inoculated to be of equivalent resistance. The method of incubation and number of test units shall be as indicated in 4.4.1 Sterility test and 4.4.1.1 Inoculated product or simulated product. The method of sterilization will control the organism in the spore strip to be tested for as shown in 4.4.1.1. NOTE: When an inoculated carrier is chosen, product samples must also be tested as required by the USP.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:53Z
Check the source to verify that this is the currentversion before use.

4.4.1.3 Alternate method. Suppliers conducting sterility tests or designating lots in a manner other than that indicated herein shall submit proof that the alternate method provides equal assurance concerning the sterility of the product. However, all sterility tests shall be conducted using biological indicators. Requests for approval of any alternate method of testing for sterility shall be made to the Contracting Officer, and approval obtained before employment of the alternate method on a government contract.

4.4.1.4 Product testing. Testing for sterility of the product after receipt by the government shall be without a biological indicator and in accordance with the USP.

4.4.2 Animal implantation test. Samples of the plastic guide and linen ligature shall be tested in accordance with the USP for Implantation Test, Plastic Containers, for class VI plastics, except after "72 hours" add "and not to exceed 144 hours"; and delete "Examine macroscopically the area of the tissue surrounding the center portion of each implant strip. Use a magnifying lens if necessary," and substitute: "The muscle tissue surrounding the implants shall be examined, both grossly and microscopically, by a pathologist certified by the American Board of Pathology or the American College of Veterinary Pathologists."

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One package containing 128 guides of assorted sizes along with 128 copies of instructions or one package containing 125 guides of size 1.3 cm along with 125 copies of instructions, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	4 units	8 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

PHILADELPHIA, PA. 19101

5.1.3 Level A.

5.1.3.1 Immediate container. Each guide shall be packaged in a sealed peel open type envelope of an appropriate size and capable of maintaining sterility of contents unless the envelope is damaged or opened. The envelope shall be designed with tab to facilitate opening.

5.1.3.2 Unit package. Unit package shall be a snug fitting box constructed in accordance with PPP-B-676 or PPP-B-636, type CF, class domestic. Commercial colors are acceptable on unit packages and the closure shall be secured to prevent spilling of contents under normal handling.

5.1.3.3 Intermediate package. Intermediate package shall be a box of appropriate size and design constructed in accordance with PPP-B-636, type CF, class weather-resistant. Boxes, shall be closed by taping the top and bottom flaps with 2 inch wide tape conforming to PPP-T-76. Closing of bottom flap with adhesive conforming to MMM-A-250 or metal stitches is permitted.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Closure and strapping shall be as specified in the appendix of the applicable box specification. Special requirements as specified in PPP-B-636 shall apply for fiberboard boxes.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraph of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method I as specified in PPP-B-636. Special requirements as specified in PPP-B-636 shall apply for fiberboard boxes.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Immediate container. Each immediate container shall be commercially identified. Marking shall include lot (control) number, size, name and address of manufacturer and the word "STERILE". In addition, marking shall include the legend: "STERILITY GUARANTEED UNLESS PACKAGE IS DAMAGED OR OPENED" or similar statement. Marking shall also comply with the requirements of the Federal Food, Drug, and Cosmetic Act.

5.3.2 Unit and Intermediate packages. Each unit and each intermediate package shall be marked incordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable box is fabricated of water-resistant material. Lot (control) number, contract or purchase order number and name of contractor shall be shown. In addition, the method of sterilization (G-Gas, R-Radiation, C-Cold, S-Steam, D-Dry Heat sterilization) shall be shown following the level of packaging and date of packaging. Example: "A-7/74-G". Final sterilization shall be on or after indicated packaging date.

5.3.3 Exterior container. Each exterior container shall be marked in accordance with MIL-STD-129. Lot (control) number, shall be shown. In addition, marking shall include the method of sterilization as specified in 5.3.2.

6. NOTES

6.1 Intended use. The circumcision guides covered by this specification are intended for guidance in the circumcision of infants.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

6.2 Ordering data. Procurement documents should specify the following:

(a) Number and date of this specification.
(b) Type and size required (see 1.2 and 3.5).
(c) Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.3 This specification covers the following circumcision guides that appear in the Federal Supply Catalog:

6515-00-364-7867	GUIDE, CIRCUMCISION, Plastic, Disposable, 1.3 cm, 125s
6515-00-757-5907	GUIDE, CIRCUMCISION, Plastic, Disposable, Assorted Sizes, 128s

It does not include all types or sizes as indicated by the title nor all types commercially available.

6.4 Supersession. This specification includes requirements of Purchase Description No. 1 for NSN 6515-00-364-7867 GUIDE, CIRCUMCISION, Plastic, Disposable, 1.3 cm, 125s, dated 1 May 1974.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6515-1839

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:53Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-G-37061

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 SUPERSEDES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:53Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026l0-21T09:53Z
Check the source to verify that this is the currentversion before use.