MIL-J-37076
2 August 1974

MILITARY SPECIFICATION

JAR, SCREW CAP, Plastic, Pathological Specimen, 1 Pint

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 This specification covers a one-pint plastic jar suitable for containing, transporting and storing fluid, semi-solid and solid laboratory specimens of a questionable pathological nature.

2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATION

Federal

| | |
|---|---|
| PPP-B-636 | Box, Fiberboard. |

STANDARDS

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Material. Material shall be as hereinafter specified and suitable for the purpose intended.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

3.2 Jar. Jar shall consist of a container and cap, which shall mate to effect a secure seal by screw action. When tested as specified in 4.4.1, the seal shall be leakproof and there shall be no fluid outside the jar or between the walls of the container and cap after 1/2 hour: there shall be no evidence of degredation due to chemical attack of the container or cap after 12 hours.

3.2.1 Container. Container shall be a transparent, rigid, polystyrene cylinder with one open end. The container may be slightly tapered. Bottom shall be flat or slightly convex and may have a raised circle near the edge. Thickness of wall and bottom of container shall be uniform with a minimum thickness of .040 inch. The container shall have a height of 4 $\pm$ 1/4 inches. Container shall rest on a flat surface without rocking.

3.2.2 Cap. Cap shall be fabricated of suitable plastic or metal and shall have a gasket which shall mate with the top edge of the container. Side of cap shall provide gripping areas.

3.3 Workmanship. Workmanship shall be first class throughout. Jar shall be free from defects which detract from their appearance or may impair their serviceability. Surface of container shall be smooth; bottom of container may have a mold mark.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Table I and MIL-STD-105.

Table I For examination

| | Inspection level | AQL (percent defective) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 1.0 |
| Minor defects | II | 2.5 |
| For dimensional examination | S-2 | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Table II and MIL-STD-105.

Table II For tests

| Unit of product | Characteristic | Requirement | Test procedure | Inspection level | AQL |
|---|---|---|---|---|---|
| Jar | Leakage and chemical attack | 3.2 | 4.4.1 | S-3 | 2.5 |

4.3 Examination. The jar shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the following table:

Table III Classification of defects *

| Categories | Defects |
|---|---|
| Major | |
| 101 | Component part missing. |
| 102 | Jar not free of cracks. |
| 103 | Gasket not secure. |
| 104 | Surface not smooth (except for area of mold mark). |
| Minor | |
| 201 | Jar not free of scratches. |
| 202 | Directions, if present, not complete, correct or legible. |
| 203 | Container not free of rocking when placed on a flat surface. |

* Examination shall not be restricted to the classified possible defects listed above.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

4.3.2 Dimensional examination. The jar shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Leakage. Fill 1/3 of the sample jars approximately half-full with distilled water, 1/3 of the sample jars approximately half-full with 70% methyl alcohol solution, and the remaining 1/3 of the sample jars approximately half-full with formaldehyde. The caps shall be secured in place with 24 inch-pounds of torque, the jars inverted, allowed to stand for at least 1/2 hour and then shaken vigorously for 15 seconds.

Allow the jars to stand for at least 12 hours in an inverted position.

5. PREPARATION FOR DELIVERY

5.1 Unit of issue. One jar, as specified, constitutes one unit of issue.

5.1.1 Packing quantities. The number of units of issue specified below shall be packed in each exterior container for the required level of protection specified in the procurement document.

5.1.1.1 Packing variation permitted. If the required number of units to be shipped is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.2 Packing. Packing shall be Level A, B or C (see 6.1).

5.2.1 Level A.

5.2.1.1 Exterior container. One hundred (100) units shall be packed in accordance with PPP-B-636, type CF, class weather-resistant. Box design shall include partitions, so arranged that each jar is contained in an individual cell. When jars are stacked in two or more tiers, a pad shall be interposed between each tier. Closure of box shall be in accordance with appendix of PPP-B-636. In addition, waterproofing shall be as specified in 30.4. Special requirements of applicable table shall apply for PPP-B-636 box.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

5.2.1.2 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. One hundred (100) units shall be packed in a box constructed in accordance with PPP-B-636, type CF, class domestic. Box design shall include partitions, so arranged that each jar is contained in an individual cell. When jars are stacked in two or more tiers, a pad shall be interposed between each tier. Closure of box shall be in accordance with method II of appendix to box specification. Special requirements of applicable table shall apply for PPP-B-636 box.

5.3 Marking.

5.3.1 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. National Stock Number (NSN).
c. Selection of applicable level of packing (see 5.2).

6.2 This specification covers the following item as listed in the Federal Supply Catalog:

| National Stock No. | Item Identification |
|---|---|
| 6530-00-985-7241 | JAR, SCREW CAP, Plastic, Pathological Specimen, 1 Pint |

6.3 This specification does not include all types, classes, grades or sizes of the commodity indicated by the title of this specification, or which are commercially available, but is intended to cover the type normally procured to meet military requirements.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6530-1322

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. *Fold on lines on reverse side, staple in corner, and send to* preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-J-37076

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304




OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
ATTN: DPSC-ATT
2800 South2 20th Street
Philadelphia, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.