MILITARY SPECIFICATION

SCALE, DIAL INDICATING, Egg, Platform

This limited coordination Military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information, but it has not been approved for promulgation as a coordinated document. It is subject to modification. However, pending its promulgation as a coordinated Military specification, it may be used in procurement.

1. SCOPE

1.1 This specification covers one type of dial indicating platform scale, suitable for weighing eggs.

2. APPLICABLE DOCUMENTS

2.1 Specifications standards and drawing. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case and Sheet, Overwrap; Water-Vaporproof, or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6670

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

DRAWING

Defense Personnel Support Center (DPSC)

DPSC Dwg. No. 10163 Scale, Dial Indicating, Egg, Platform.

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

NATIONAL BUREAU OF STANDARDS (NBS)

Handbook 77 Design and Test of Standards of Mass.
Vol. III
No. 3

(Application for copies should be addressed to the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20401.)

3. REQUIREMENTS

3.1 Material. Material shall be of the highest quality throughout, and suitable in all respects for the purpose intended.

3.2 Style and design. The scale shall be in strict accordance with Defense Personnel Support Center Drawing 10163.

3.2.1 Mechanism. The scale shall be operated by means of a spring mechanism without counterpoise weights.

3.2.2 Load receiver. The load receiver shall be of the floating platform type.

3.2.3 Dial. The scale shall have a circular dial with a range of at least 0 to 50 pounds (lb.) avoirdupois in 2-ounce (oz.) intervals.

3.2.4 Locking device. Manipulation of locking device shall lock the dial at the weight being indicated, so that weight may be read after the egg crate is removed from the load receiver.

3.2.5 Adjusting screw. Calibration of the dial shall be accomplished by means of the adjusting screw. Dial shall read the same as the weight on the load receiver when tested as specified as 4.4.1.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

3.3 Accuracy. Dial shall read the same as the weight on the load receiver ± 8 oz.

3.4 Finish. The finish shall be baked enamel

3.5 Identification marking. Each scale shall be permanently and legibly marked with the name or registered trademark of the manufacturer.

3.6 Workmanship. Workmanship shall be first class throughout. Scale shall be free from defects listed in Table III or other defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual, or auditory, investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with Table I and MIL-STD-105. Unit of product shall be one scale

Table I. For examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Minor defects	II	2.5
For dimensional examination	S-1	2.5

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with Table II and MIL-STD-105. Unit of product shall be one scale, and acceptance number shall be zero.

Table II. For tests

Characteristic	Requirement	Test Procedure	Inspection level
Precision	3.2.5	4.4.1	S-1
Accuracy	3.3	4.4.2	S-1

4.3 Examination. The scale shall be examined to determine compliance with all requirements contained in this specification. Examination shall be in accordance with, but shall not be restricted to, defects listed in the following table:

Table III. Classification of defects

Category	Defects
Major	
101	Scale not free of sharp edges.
102	Surfaces not free of pits and burs.
103	Enameled surface not free of bare spots.
104	Numbers on dial move when pressure is exerted on the top of the load receiver and while the locking device is engaged.
105	Numbers on dial do not move when pressure is exerted on the top of the load review and while the locking device is not engaged.
Minor	
201	Identification marking not legible or permanent.

4.3.1 Dimensional examination. The scale shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements. Class C weights traceable to National Bureau of Standards standards shall be used for testing.

4.4.1 Precision. Weights weighing at least 45 pounds shall be placed in the center of the load receiver. Locking device shall be disengaged. Dial shall be calibrated by means of the adjusting screw to read the same as the weight on the load receiver. Locking device shall be engaged, then disengaged and the dial read. If the same scale is used for testing in accordance with 4.4.2, testing in accordance with 4.4.2 shall be done first.

4.4.2 Accuracy. Weights weighing at least 45 lbs. shall be placed in the center of the load receiver. Locking device shall be disengaged and the dial read.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One scale, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table IV shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table IV. Packaging Quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	6 units

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each scale shall be wrapped in kraft paper or shall be inserted into a plastic bag and the kraft wrap or plastic bag secured. Each scale shall be further packaged in a box of appropriate size constructed in accordance with PPP-B-636, type CF, class domestic. The box shall be designed with suitable interior fitting or shall contain sufficient cushioning material to insure against shifting of or damage to the contents. Closure of the box shall conform to method I as specified in PPP-B-636.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type, PPP-B-621, class 2, or PPP-B-636, class weather-resistant. Special requirements as specified in Table II of PPP-B-636 shall apply for fiberboard boxes.

5.2.1.2 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4.2 of PPP-B-636.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Fiberboard box shall be closed in accordance with method II as specified in PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements specified in Table I of PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

5.3 Marking.

5.3.1 Unit package. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. National Stock Number (NSN).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification covers the following item listed in the Federal Supply Catalog:

National Stock No.	Item Identification
6670-00-810-7780	SCALE, DIAL INDICATING, Egg, Platform

6.3 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6670-A129

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
2800 South 20th Street
ATTN: DPSC-ATT
Philadelphia, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No 22 R255

INSTRUCTIONS The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government Fold on lines on reverse side, staple in corner, and send to preparing activity Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements Attach any pertinent data which may be of use in improving this document If there are additional papers, attach to form and place both in an envelope addressed to preparing activity

DOCUMENT IDENTIFIER AND TITLE

MIL-S-37071 (DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER \

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1 HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A GIVE PARAGRAPH NUMBER AND WORDING

B RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2 COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3 IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4 REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO

DATE

DD FORM 1 JAN 72 1426

REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:54Z
Check the source to verify that this is the current version before use.