MILITARY SPECIFICATION

STAND, DISTILLING APPARATUS, LABORATORY

This limited coordination Military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information, but it has not been approved for promulgation as a coordinated document. It is subject to modification. However, pending its promulgation as a coordinated Military specification, it may be used in procurement.

1. SCOPE

1.1 This specification covers a metal stand suitable for supporting a laboratory distilling apparatus at a height which will allow a 13-inch clearance between the water draw-off hose nipple and the supporting countertop (see 6.2).

2. APPLICABLE DOCUMENTS

2.1 Specifications, standards and drawing. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein.

SPECIFICATIONS

Federal

PPP-B-585	Boxes, Wood, Wirebound.
PPP-B-601	Boxes, Wood, Cleated-Plywood.
PPP-B-621	Boxes, Wood, Nailed and Lock-Corner.
PPP-B-636	Box, Fiberboard.

Military

MIL-L-10547	Liners, Case, and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

Military

MIL-STD-105	Sampling Proceeures and Tables for Inspection by Attributes.
MIL-STD-129	Marking for Shipment and Storage.

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24338 Stand, Distilling Apparatus, Laboratory

(Copies of specifications, standards and drawing required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

3. REQUIREMENTS

3.1 Materials. All metal parts, except screws and nuts, shall be fabricated from 0.0747 inch (No. 14 U.S. Standard Gauge) thick steel. Commercial tolerances shall apply.

3.2 Design and construction. The components of the stand shall be constructed in accordance with DPSC Drawing No. 24338, and shall consist of:

a. Three angle tie bars.
b. Two stand ends with predrilled holes to accommodate levelers not required to be supplied with subject item (see 6.3).
c. Screws and nuts for proper assembly of the components of the stand.

The assembled stand shall be capable of supporting a 150-pound load.

3.3 Dimensions. The dimensions of the angle bars and ends shall be in accordance with 3.1 and DPSC Drawing No. 24338. A tolerance of ± 1/32 shall apply to the dimensions with the exception of the thickness.

3.4 Finish. The angle bars and ends shall be thoroughly cleaned to remove foreign particles such as dirt, grease and metal chips. After cleaning, the components shall have an applied finish resistant to corrosion. The manufacturer's normal commercial finish shall be acceptable. The screws and nuts shall be suitably resistant to corrosion.

3.5 Instructions. Each unit shall be supplied with printed instructions for the proper assembly of the components of the stand.

3.6 Workmanship. Workmanship shall be first class throughout. The stand shall be free from burrs, sharp edges, deformation, corrosion and other defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:55Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and Table I.

Table I Sampling for examination

	Inspection level	AQL (percent defective)
For visual examination		
Major defects	II	1.0
Minor defects	II	2.5
For dimensional examination	S-1	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105.

4.3 Examination. The stand shall be examined to determine compliance with all requirements contained in this specification.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

4.3.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in Table II. Examination shall not be restricted to the classified possible defects listed in the table.

Table II Classification of defects

Categories	Defects
Major	
101	Components of stand not free of deformation.
102	Components of stand not free of burrs or sharp edges.
103	Applied finish on components of stand not free of imbedded foreign matter, chips, corrosion or discontinuities.
104	Applied finish on components not free of tackiness.
105	Threads on screws or nuts not free of damage or discontinuities.
Minor	
201	Finish on components of stand not uniformly applied.
202	Components of stand not free of removable foreign matter.

4.3.2 Dimensional examination. The stand shall be examined for defects in examination. Any dimension not within specified limits shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

5. PREPARATION FOR DELIVERY

5.1 Packaging.

5.1.1 Unit of issue. One stand, as specified, constitutes one unit of issue.

5.1.2 Packaging quantities. The number of units of issue indicated in Table III shall be packaged in each unit, intermediate and exterior container, as applicable, for the required level of protection specified in the procurement document.

Table III Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	Not required	5 units

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

5.1.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

5.1.3 Level A.

5.1.3.1 Unit package. Each unit shall be disassembled, and the disassembled components shall be wrapped in kraft paper. The wrapped components shall be further packaged in an snug-fitting box constructed in accordance with PPP-B-636, type CF, class domestic. The box shall contain sufficient cushioning material to insure against shifting of or damage to the contents. Closure of the box shall conform to method I as specified in PPP-B-636.

5.1.4 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A.

5.2.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 3, style 3; PPP-B-601, overseas type; PPP-B-621, class 2; or PPP-B-636, class weather-resistant. Fiberboard boxes shall conform to special requirements specified in PPP-B-636.

5.2.1.2 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes.

5.2.1.3 Closure. Closure of wood boxes shall be in accordance with appendix of applicable box specification. Closure of each fiberboard box shall be as specified in the appendix, and waterproofing shall conform with 30.4 of PPP-B-636.

5.2.1.4 Strapping. Strapping, when required, shall be in accordance with appendix of applicable box specification.

NOTE: Strapping shall not be required for shipments forwarded to a receiving activity within the continental limits of the United States for storage and redistribution.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:55Z
Check the source to verify that this is the current version before use.

5.2.2 Level B.

5.2.2.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-585, class 1, style 3; PPP-B-601, domestic type; PPP-B-621, class 1, or PPP-B-636, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636. In addition, fiberboard boxes shall conform to the special requirements specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking.

5.3.1 Unit packages. Each unit package shall be marked in accordance with MIL-STD-129. When labels are utilized, waterproofing shall be required only when applicable carton is fabricated of water-resistant material. Contract or purchase order number and name of contractor shall be shown.

5.3.2 Exterior container. Exterior container shall be marked in accordance with MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. National Stock Number (NSN).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).

6.2 This specification covers the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6640-00-110-9413	STAND, DISTILLING APPARATUS, LABORATORY

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:55Z
Check the source to verify that this is the current version before use.

The stand is intended for use with the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6640-00-110-9412	DISTILLING APPARATUS, LABORATORY

6.3 For bench mounting of NSN 6640-00-110-9412, the four levelers supplied with this item are intended to be removed and transferred to subject item (NSN 6640-00-110-9413) in accordance with DPSC Drawing No. 24338.

6.4 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type normally procured to meet military requirements.

Review activities:
Army - MD
Navy - MS
Air Force - 03

Custodian activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0841

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.

☆ U. S. GOVERNMENT PRINTING OFFICE: 1974-603-109/1326

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:55Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE
MIL-S-37073 (DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A
☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?
   A. GIVE PARAGRAPH NUMBER AND WORDING.

   B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?
   ☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF 1 JAN 68 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:55Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
ATTN: DPSC-ATT
2800 South 20th Street
Philadelphia, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:55Z
Check the source to verify that this is the current version before use.