MILITARY SPECIFICATION

TEST KIT, SYPHILIS DETECTION

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a test kit suitable for use in performing the Rapid Plasma Reagin Tests (RPR CARD TEST) for the presence of syphilis and other treponematoses by means of specifically prepared disposable cards and antigen. The kit shall be for use in laboratories where numerous venous blood specimens are run and where mechanical rotation is available. (See 6.2)

2. APPLICABLE DOCUMENTS

2.1 The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

PPP-B-636	Box, Fiberboard.
PPP-C-186	Containers, Packaging and Packing for Drugs, Chemicals, and Pharmaceuticals. (See 6.3)

STANDARDS

Military

MIL-STD-105	Sampling Procedures and Tables for Inspection by Attributes.

(Copies of specifications and standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

2.2 Other publications. The following [illegible] a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

U.S. DEPARTMENT OF HEALTH, EDUCATION, AND WELFARE,
FOOD AND DRUG ADMINISTRATION

Federal Food, Drug, and Cosmetic Act and
Regulations Promulgated Thereunder.

(Application for copies should be addressed to the Food and Drug Administration, Washington, DC 20204.)

3. REQUIREMENTS

3.1 Material shall be as specified hereinafter for the specified component. The kit shall consist of the following components necessary for performing 500 Rapid Plasma Reagin (RPR) Circle Card Tests.

3.1.1 Components. Each kit shall contain the following components:

(a) 50 Brewer Diagnostic Cards with 10, 18-mm printed circles each, plus Reading Guide Information (See 3.2.1 and 3.2.1.1, respectively.)

(b) 500 (minimum) dispensing tubes. (See 3.2.2)

(c) 1 Plastic needle hub with needle attached. (See 3.2.3)

(d) 3 Ampuls, 3-ml size, RPR Card antigen suspension. (See 3.2.4)

(e) 1 Flexible plastic dispensing vial for antigen. (See 5.1.1.4)

(f) 1 Directions for Use pamphlet. (See 3.2.5)

(g) Venereal Disease Research Laboratory RPR "approval" note. (See 3.2.4.2)

3.2 Description of components.

3.2.1 Brewer diagnostic card. Shall be plastic-coated board, approximately 3 inches wide and 5 inches long, with ten (10), printed, 18 mm internal diameter circles. Fifty (50) such diagnostic cards shall be furnished in each kit.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the current version before use.

3.2.1.1 Reading guide information. The information shall illustrate the control for a reactive test; a reactive (minimal) test; and a non-reactive test. One reading guide containing the information shall be furnished with each kit. If the reading guide card is not supplied, the information on the card shall be contained in the directions for use pamphlet.

3.2.2 Dispensing tubes. The dispensing tubes shall be plastic with a controlled inner bore which will cause delivery of 0.05 ml of plasma or serum, when compressed. The tubes shall be closed at one end so as to provide a suitable stirring rod. A box containing at least 500 dispensing tubes shall be furnished with each kit.

3.2.3 Needle and needle hub. Shall be a 20 gauge, 3/4 inch, burr free, square-cut needle capable of dispensing 58 to 62 drops per ml. The needle shall be attached to a plastic needle hub and shall be shielded with a suitable polyethylene cover. One needle unit shall be furnished with each kit.

3.2.4 Antigen. Each kit shall be furnished with three (3) ampuls, each containing approximately 3 ml of antigen charcoal suspension, suitable for use with the RPR Card Test.

3.2.4.1 Conformity to standards. The reactivity of each lot of RPR Card antigen suspension with a standard control serum shall be compared to a standard pattern of reactivity for conformity.

3.2.4.2 Approval of Venereal Disease Research Laboratory. A sample of each lot of RPR Card antigen suspension used in the Syphilis Detection Test Kit, shall be submitted to the Venereal Disease Research Laboratory, Communicable Disease Center, Atlanta, Georgia 30322, for testing and approval before release of the product. Each kit shall contain one (1) insert stating that the Venereal Disease Research Laboratory of the United States Public Health Service has tested the lot number designation on the face of the kit and found said lot number of antigen suspension to reproduce reactivity of its standard test reagents. If the insert is not supplied, the information on the insert shall be on the outside of the unit package and on the 9 ml package RPR Card antigen suspension containing the three 3 ml size ampuls.

3.2.4.3 Potency period. The antigen shall have a potency period (expiration dating period) of 12 months from the date of manufacture. The date of manufacture is defined as the date of filling of the final containers, or the date of manufacturer's or contractor's final quality approval of the final containers, which shall be not later than one month after the date of filling.

3.2.4.4 Ingredients. The chemicals, including sodium chloride, and the distilled water entering into the finished antigen shall comply with the Reagent Control section of the Manual of Tests for Syphilis, 1969, U.S. Department of Health, Education and Welfare, Public Health Service. The end item shall be approved and released by the Center for Disease Control, Atlanta, Georgia 30333. In addition, the manufacturer's proprietary [illegible] shall be acceptable.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the current version before use.

3.2.4.5 Delivery. Not less than 10 months of the expiration dating period (potency period) shall remain at time of delivery to the Government.

3.2.5 Directions for use pamphlet. One (1) pamphlet which describes and illustrates complete performance of the RPR Card Test shall be furnished as a component of each kit.

3.3 Workmanship. Workmanship shall be first class throughout. The kit and its components shall be free from defects which detract from their appearance or may impair their serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 A company supplying ingredient(s) to the contractor will not be considered an acceptable facility for the performance of any inspection requirements specified herein.

4.2 Lot. For purposes of this specification, a lot, batch, or control is that single, uniform, and homogeneous quantity of antigen suspension produced from one formulation, subjected to the same compounding and manufacturing operation, and filled into final containers.

4.2.1 Inspection. Inspection shall be conducted, as necessary, to determine compliance with the requirements of this specification.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the currentversion before use.

4.3 Sampling. Sampling shall be conducted in accordance with the procedures set forth in MIL-STD-105.

For visual examination	Inspection level	AQL (percent defective	Unit of product
Major	S-2	1.0	Unit Package
Minor	II	2.5	Unit Package

4.3.1 Classification of defects. Examination will be conducted in accordance with the following classification of defects:

Categories and defects*

Major
- 101 50 Brewer Diagnostic Cards not included with each kit.
- 102 Reading Guide Information not included with each kit.
- 103 Plastic needle-hub with needle attached not included with each kit.
- 104 Less than 500 dispensing tubes in each kit.
- 105 Three ampules of 3 ml size and an RPR Card antigen suspension not included with each kit.
- 106 One flexible plastic dispensing vial for antigen not included with each kit.
- 107 One directions for use pamphlet not included with each kit.
- 108 Venereal disease research laboratory RPR "approval" note not included with each kit.

Minor
- 201 Exterior of kit not free from contents or foreign matter.

*Inspection is not restricted to the classified possible defects listed above.

4.4 Tests.

4.4.1 Tests shall be conducted in accordance with the requirements set forth in this document.

4.5 Examination of preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of section 5 for packaging, packing and marking. Defects shall be scored as specified in Table I. Sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the currentversion before use.

Table 1. Examination of preparation for delivery

Examination	Defect
Containers	Not as specified.
Quantity	Quantity in interior packages and shipping containers not as specified.
Materials	Any component missing or damaged.
Workmanship	Inadequate application of components such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, Incorrect, Illegible, improper size, location, sequence, or method of application.

5. PREPARATION FOR DELIVERY

5.1 Preparation for delivery shall be as specified herein and shall be in accordance with all applicable requirements of PPP-C-186, and as specified herein.

5.1.1 Immediate containers. Shall be as follows:

5.1.1.1 Antigen. Each immediate container for the antigen shall comply with the following classification:

GROUP A CLASS 1 TYPE a STYLE 7 GRADE 2

CLOSURE K

5.1.1.2 Dispensing tubes. Shall be furnished in a set-up, folding carton of suitable size to hold 500 dispensing tubes.

5.1.1.3 Needle. The needle-hub component shall be suitably protected by a polyethylene shield.

5.1.1.4 Dispensing bottle for the antigen. Shall be a non-rigid ophthalmic type bottle having a threaded finish and an ophthalmic type dispensing tip inserted in the neck of the bottle. The dispensing tip shall fit the plastic needle hub.

Closure. Shall be a rigid, polystyrene screwcap with an integral plug conforming to the orifice of the bottle.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the currentversion before use.

5.1.1.5 Diagnostic cards. Shall be wrapped in a transparent polyvinyl chloride film.

5.2 Labeling. Labeling shall be in accordance with the requirements of the Federal Food, Drug, and Cosmetic Act, and shall include the information required below.

5.2.1 Immediate container. Each immediate container label for the Antigen, Capillary tubes and Dispensing bottle shall bear the following information:

(a) label information in accordance with commercial practice.

5.2.2 Unit packages. Each unit package label shall bear the following:

(a) labeling information in accordance with commercial practice.

(b) the National Stock Number.

(c) the expiration date.

NOTE: The National Stock Number shall not be preceded by any prefix.

Strip labeling may be used.

Label adhesion test shall not apply.

5.3 Packaging. Packaging shall be Level A or C as specified (see 6.1).

5.3.1 Unit of issue. One test kit, as specified, constitutes one unit of issue.

5.3.2 Packaging quantities. The number of units of issue indicated in Table II shall be packaged in each unit, intermediate, and exterior container, as applicable, for the required level of protection specified in the procurement document:

Table II. Packaging quantities

Unit package	Intermediate package	Exterior container
1 unit	10 units	20 units

5.3.2.1 Packing variation permitted. If the required number of units in the entire shipment is less than the number of units specified to be overpacked in an exterior container, such units may be packed in an exterior container of suitable size and design, acceptable to a common carrier, which shall insure safe delivery to destination.

Source: https://assist.dla.mil -- Downloaded: 202610-21T09:57Z
Check the source to verify that this is the current version before use.

5.4 Packing. Packing shall be Level A or C as specified (see 6.1).

5.4.1 Level A.

5.4.1.1 Exterior container. Exterior container shall be designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant, grade V3c. Closure, waterproofing, and strapping shall be as specified in the appendix of the box specification.

5.4.2 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size, and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.5 Marking.

5.5.1 Intermediate package. Each intermediate package shall be marked in accordance with paragraph 5.5.3 of PPP-C-186. In addition, Type II shelf-life markings apply as specified in PPP-C-186. Marking shall include the legends:

"STORE BETWEEN 2° - 8° C. (35° - 46° F.)."
and
"DO NOT FREEZE."

5.5.2 Exterior container. Each exterior container shall be marked in accordance with paragraph 5.5.4 of PPP-C-186. In addition, Type II shelf-life markings apply as specified in PPP-C-186. Marking shall include the legends:

"STORE BETWEEN 2° - 8° C. (35° - 46° F.)."
and
"DO NOT FREEZE."

NOTE: The unrefrigerated shipping time for this item shall be not more than 168 hours (7 days).

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

(a) Title, number and date of this specification.
(b) the National Stock No.
(c) Selection of applicable levels of packaging and packing (see 5.3 and 5.4).

6.2 This specification covers the following item:

National Stock Number	Item Identification
6505-00-159-5012	TEST KIT, SYPHILIS DETECTION, Brewer type, 500 tests.

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the current version before use.

6.3 [illegible] refers to PPP-C-186 refer to Interim Federal Specification PPP-C-001868 which is mandatory for use until superseded by a revision.

6.4 This specification does not cover all types, classes, grades, or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the types which are normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DH

Project No. 6505-1315
cs

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the current version before use.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS. The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable [illegible] at reasonable cost and minimum delay. It will sometimes enhance use of the document. Both contractors and government activities, as manufacturers, vendors who are prospective suppliers of the product are invited to submit comments to the Government. Fold on lines as required, staple or glue, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced documents or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-T-37096

NAME OF ORGANIZATION AND ADDRESS	CONTRACT NUMBER
	MATERIAL PROCURED UNDER A ☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRE INTERPRETATION IN PROCUREMENT USE

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*	TELEPHONE NO
	DATE

DD FORM 1 JAN 72 1426 REPLACES EDITION OF [illegible] WHICH MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

FOLD

DEFENSE PERSONNEL SUPPORT CENTER
2800 South 20th Street
Philadelphia, Pa. 19101

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S. MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

HEADQUARTERS, DEFENSE PERSONNEL SUPPORT CENTER
ATTN: DIRECTORATE OF MEDICAL MATERIEL, CODE ATT
2800 SOUTH 20TH STREET
PHILADELPHIA, PA. 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 202640-21T09:57Z
Check the source to verify that this is the current version before use.