MILITARY SPECIFICATION

# WATER BATH, ELECTRIC, Precision, 6 Rack, 115 volt, 50-60 Hz, AC

This limited coordination Military specification has been prepared by the Defense Personnel Support Center based upon currently available technical information, but it has not been approved for promulgation as a coordinated Military specification. It is subject to modification. However, pending its promulgation as a coordinated document, it may be used in procurement.

## 1. SCOPE

1.1 This specification covers an electrically heated, thermoregulator-controlled water bath with constant water level, suitable for use in laboratory testing; it has a temperature range from slightly above ambient to 60°Centigrade (°C.) when the bath is uncovered, and to 100°C. when the bath is covered (see 6.4).

## 2. APPLICABLE DOCUMENTS

2.1 Specifications and standards. The following documents, of the issue in effect on date of invitation for bids or request for proposals, form a part of this specification to the extent specified herein:

SPECIFICATIONS

Federal

| | |
|---|---|
| GG-T-336 | Thermometer, Self-Indicating, Liquid-in-Glass, Chemical. |
| PPP-B-601 | Boxes, Wood, Cleated-Plywood. |
| PPP-B-621 | Boxes, Wood, Nailed and Lock-Corner. |
| PPP-B-636 | Box, Fiberboard. |

Military

| | |
|---|---|
| MIL-M-7298 | Manual, Technical: Commercial Equipment. |
| MIL-L-10547 | Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible. |



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

STANDARDS

Federal

| | |
|---|---|
| FED-STD-66 | Steel, Chemical Composition and Hardenability. |

Military

| | |
|---|---|
| MIL-STD-105 | Sampling Procedures and Tables for Inspection by Attributes. |
| MIL-STD-129 | Marking for Shipment and Storage. |

(Copies of specifications and standards required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposals shall apply.

AMERICAN NATIONAL STANDARDS INSTITUTE (ANSI)

| | |
|---|---|
| C73.11 | Dimensions of Caps, Plugs and Receptacles, 125 volts, 15 amperes, 2-Pole, 3-Wire, Grounding Type, General Purpose. |

(Application for copies should be addressed to the American National Standards Institute, 1430 Broadway, New York, NY 10018.)

UNDERWRITERS' LABORATORIES, INC. (UL)

| | |
|---|---|
| 499 | Standard for Heating Appliances, Electric. |

(Application for copies should be addressed to the Underwriters' Laboratories, Inc., 207 East Ohio Street, Chicago, IL 60611; 333 Pfingsten Road, Northbrook, IL 60062; 1285 Walt Whitman Road, Melville, L.I., NY 11746; or 1655 Scott Boulevard, Santa Clara, CA 95050.)

3. REQUIREMENTS

3.1 Fire and casualty hazards.

3.1.1 The contractor shall submit proof, prior to delivery, that the items to



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

be delivered under the contract conform to the applicable requirements of UL 499. Proof shall consist of one of the following:

a. UL label or UL listing of the product, provided the contractor agrees to have Underwriters' Laboratories, Inc. forward a copy of the UL report directly to the contracting officer, upon the latter's request.

b. Laboratory report from a qualified independent laboratory. If an independent laboratory is used, it must be adequately equipped and competent to determine compliance with UL standards. The qualification of such laboratory shall be determined by the Government which reserves the right to require a facilities report and to conduct a facilities survey. Any laboratory report submitted shall include actual test methods, data taken and results of tests. Detailed findings of examinations necessary to determine compliance shall also be included. The test report shall also have a section that describes the test sample in such a way that the report can be used to insure identity of production items with the test sample. This section shall be supplemented with construction information such as materials used, spacings between electrical components, types of bushings, photographs, wiring diagrams, sketches and parts list, all appropriately identified.

3.1.2 Compliance with above paragraph does not excuse compliance with all other requirements of the contract and specifications.

3.2 Materials. Materials shall conform to the applicable specifications and requirements as specified herein. Materials not specifically designated shall be of high commercial quality throughout, and shall be entirely suitable for the purpose intended.

3.2.1 Corrosion-resisting steel. Chemical composition of corrosion-resisting steel shall comply with FED-STD-66, AISI 300 series.

3.2.2 Insulation. Insulation shall be of nonhygroscopic, non-packing, non-powdering material having a thermal conductivity coefficient not exceeding 0.27 BTU/hr/sq.ft./$^{o}$F./in. of thickness.

3.3 Design and construction. The water bath shall be capable of operating on 115 volt, 50/60 Hz, AC, 1000 watt supply lines. Water bath shall be of double-walled construction with insulation between walls. Exterior walls shall be of corrosion-resisting steel .025 inch thick No. 24 United States Standard (USS) gage. Exterior dimensions shall be: length, 15 $\pm$ 1/4 inches, width, 14 $\pm$ 1/4 inches, and height, 16-3/4 $\pm$ 1/4 inches. Nominal interior (water chamber) dimensions shall be: length, 13 inches, width, 12 inches and depth, 7 inches. Design shall be such that



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

the water chamber accommodates six Wassermann-type laboratory test tube racks that can hold twenty 13-mm. test tubes, each rack having overall dimensions as follows: length, 10 inches, width, 2-1/16 inches and height, 4 inches (see 6.2). The bottom and sides of water chamber shall be fabricated of corrosion-resisting steel .019 inch thick No. 26 USS gage. For purposes of this specification, the portion which holds the test tube racks will be defined as water chamber and the entire unit as water bath. The water chamber and outside body shall be of spot-welded construction. Water chamber shall be soldered watertight. Water bath shall have a removable base fabricated of galvanized steel .060 inch thick No. 16 USS gage. The bottom of the water bath shall be supplied with projecting feet. The water bath shall consist essentially of components specified in 3.3.1 through 3.3.15 and other components necessary for the water bath to meet the requirements of this specification.

3.3.1 Cover. Cover shall be fabricated of corrosion-resisting steel .025 inch thick No. 24 USS gage and shall be the gable type. Cover shall have a plastic handle on the top and two cut-outs on one side. Cut-outs shall provide space for the thermoregulator and thermometer holder with thermometer. In order to prevent the cover from sliding off or into the water chamber, the bottom of cover shall protrude into the water chamber and shall have a flange on sides which rest on the top of water bath. The design and dimensions shall be such that the cover can be easily set in place and can be easily removed.

3.3.2 Diffuser platform. Diffuser platform shall be fabricated of corrosion-resisting steel expanded and flattened 1/2-inch diamond shape No. 20-22 USS gage. Shall be polished on both sides with a suitable commercial finish. All sides shall be bent in order to raise the platform 1-1/4 inches above the bottom of water chamber. Platform shall support the test tube racks and protect the stirring bar and heater.

3.3.3 Stirring bar. Stirring shall be achieved with a magnetic egg bar sealed in molded teflon, located in magnetic field at the bottom center of water chamber on centering monel disk and surrounded with flow director. Nominal length of the stirring bar shall be 2-1/2 inches.

3.3.4 Centering disk. Centering disk for stirring bar shall be fabricated of monel .018 inch thick No. 26 USS gage. Shall be a nominal 3 inches in diameter.

3.3.5 Flow director. Flow director shall be cylindrical and shall be fabricated of corrosion-resisting steel .038 inch thick No. 20 USS gage. Nominal dimensions shall be: height, 1-1/4 inches and inside diameter, 3-3/8 inches. Shall have holes 1/8-inch in diameter (nominal). Holes shall be so placed that the hot water is distributed evenly throughout the water chamber.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

3.3.6 Heating element. Heating element shall be a copper-sheathed immersion type with copper gaskets, washers and locknuts for watertight mounting. Shall be located at the bottom of water chamber. Shall be rated 120 volt, 1000 watts.

3.3.7 Stirrer motor. Shall be induction type, non-sparking, with high temperature insulation and lubrication. Shall operate at 120 volts, 50/60 cycles, 35 watts nominal, and shall turn at 580 revolutions per minute (r.p.m.). Motor shall drive a magnet located below bottom of water chamber and fan located below motor. The fan shall drive away hot air from motor, through the vents in the sides and in the back of water bath.

3.3.8 Thermoregulator. Thermoregulator shall be immersion type, mercury sealed in glass, connected to an electronic relay switch assembly. Thermoregulator shall be graduated in 1°C. increments with total range of -10° to +110°C. There shall be a knurled ring near the top of the stem. Temperature settings shall be made by rotation of the ring.

3.3.9 Thermometer. The thermometer shall be a partial immersion mercury-in-glass type, and shall have nominal dimensions of length, 10 inches and diameter, 1/4 inches. As a minimum the thermometer shall include a 0-100°C. range with 1° graduations. The thermometer shall have markings to clearly indicate the proper depth of immersion for the thermometer when it is inserted in the water bath. The thermometer shall be capable of being secured and positioned in the thermometer holder of 3.3.10. When tested as specified in 4.4.1, the thermometer shall have an accuracy of $\pm$ 0.5°C.

3.3.10 Thermometer holder. Thermometer holder shall be fabricated of corrosion-resisting steel .038 inch thick No. 20 USS gage. Shall have a hole in which a rubber ring shall be inserted to hold the thermometer safely.

3.3.11 Thermoregulator bracket. Shall be fabricated of corrosion-resisting steel .038 inch thick No. 20 USS gage. Shall have hole which has rubber ring inserted in it and which accommodates the thermoregulator. Shall be securely fastened in the back of water bath.

3.3.12 Water level regulator and drain valve. Constant water level control and drain cock shall be fabricated of nickel-plated brass 1/8 National Taper Pipe (N.T.P.) size.

3.3.13 Control panel housing. Control housing shall be fabricated of plastic and shall be located on front side of water bath. Shall have nameplate, heater switch and pilot light.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

3.3.13.1 Nameplate. Nameplate shall include marking specified in 3.6.

3.3.13.2 Heater switch. Heater switch shall be "On-Off" line type, single pole, 15 amperes, 125 volts, AC.

3.3.13.3 Pilot light. Pilot light shall light when heating element is energized and shall not light when heating element is not energized.

3.3.14 Motor switch. Motor switch shall be "On-Off" line type, single pole, 15 amperes, 125 volts, AC. Shall be located on right side from control panel housing in front of water bath.

3.3.15 Power cord. A three-conductor power cord with a non-detachable plug shall be supplied. A strain relief bushing shall be provided where the power cord enters the water bath unit. The plug shall be a three-blade male plug that is suitable for use with receptacles described in ANSI C73.11.

3.4 Performance.

3.4.1 Dielectric strength. There shall be no electrical breakdown when the water bath is tested as specified in 4.4.2.

3.4.2 Temperature stability. The indicated temperature shall be constant within $\pm$ 0.05°C. (highest temperature shall not exceed lowest temperature by more than 0.1°C.), when tested as specified in 4.4.3.

3.4.3 Heating time. The water bath shall reach a constant temperature (as defined in 3.4.2) within 2-1/4 hours, when tested as specified in 4.4.4.

3.4.4 Operating temperature range. Water in the uncovered water bath shall reach a temperature of at least 60°C. and water in the covered water bath shall reach the temperature of boiling water when tested as specified in 4.4.5.

3.5 Identification markings. Each bath shall have a nameplate, which shall be permanently and legibly marked with the following information: manufacturer's name or registered trademark, contract or purchase order number, National Stock Number, "U.S.", the rated voltage, frequency and wattage, and serial number. Markings secured by adhesive shall not be used.

3.6 Service data. The contractor shall furnish, with each unit, two copies each of a technical manual which contains complete instructions for installation, operation and maintenance, and lists of component parts. The manual shall comply



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

with the requirements of MIL-M-7298. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed technical manual, for review and comment by the Directorate of Medical Materiel, Defense Personnel Support Center. The preliminary manuals shall be submitted sufficiently early to permit adequate review by the Directorate of Medical Materiel, Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of any required revisions and corrections in the final manual without delaying delivery under the contract or purchase order. In addition, the contractor shall furnish 15 copies of the approved technical manual to the contracting officer for distribution. Should the supplier have furnished acceptable manuals for the identical item within two years preceding the date of contract and propose to furnish manuals identical with those previously accepted, the requirements for review of drafts and submission of 15 additional copies may be waived upon request to the contracting officer. Waiver of submission of 15 additional copies shall be at the option of the Government.

3.7 Workmanship. Workmanship shall be first class throughout. Soldering and welding shall be free from cracks and porosity. The water bath shall be free from sharp edges, pits, burrs, cracks, dents and other defects which detract from its appearance or may impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

4.1.3 Certificate of quality. Certificate of quality, supplied by the manufacturer of the steel or thermometer, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results and the specification requirements.

4.2 Sampling.

4.2.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptance quality levels (AQL) shall be as indicated in Table I. Unit of product shall be the water bath.

Table I Sampling for examination

| | Inspection level | AQL (percent defective) |
|---|---|---|
| For visual examination | | |
| Major defects | II | 1.0 |
| Minor defects | II | 2.5 |
| For dimensional examination | S-2 | 2.5 |

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and Table II. Unit of product shall be the water bath or thermometer as applicable.

Table II Sampling for tests

| Characteristic | Requirement | Test procedure | Inspection level | AQL (percent defective) |
|---|---|---|---|---|
| Thermometer accuracy | 3.3.9 | 4.4.1 | S-1 | * |
| Dielectric strength | 3.4.1 | 4.4.2 | 100% | - |
| Temperature stability | 3.4.2 | 4.4.3 | S-2 | * |
| Heating time | 3.4.3 | 4.4.4 | S-2 | * |
| Operating temperature range | 3.4.4 | 4.4.5 | S-1 | * |

* Acceptance number shall be zero.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

4.3 Examination. The bath shall be examined to determine compliance with all requirements contained in this specification.

4.3.1 Classification of defects. The water bath shall be examined for defects, including, but not limited to, those listed in Table III.

Table III Classification of defects

| Categories | Defects |
|---|---|
| Major | |
| 101 | Unit not complete with all component parts. |
| 102 | Unit not free of cracks, tears or distortion. |
| 103 | Soldering or welding not free of cracks or porosity. |
| 104 | Unit not free of wobbling or rocking when placed on a flat and level surface. |
| 105 | Unit not free of sharp edges. |
| 106 | Metal parts not free of corrosion. |
| 107 | Power cord not adequately secured within unit. |
| 108 | Power cord not free of cuts, tears or abrasion. |
| 109 | Blades of power cord plug not properly aligned. |
| Minor | |
| 201 | Unit not free of minor dents. |
| 202 | Metal surfaces not free of pits and burrs. |
| 203 | Metal surfaces not free of scratches or chipping which detract from appearance. |
| 204 | Identification marking not complete, legible or permanent. |

4.3.2 Dimensional examination. The bath shall be examined for defects in dimensions. Any dimension not within tolerances specified herein shall be classified as a defect.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Thermometer accuracy. The thermometer shall be tested against a standard thermometer at 37 $\pm$ 2°C., 56 $\pm$ 2°C. and 100°C. The standard thermometer shall be certified by the National Bureau of Standards.

4.4.2 Dielectric strength. With the power cord plug disconnected and the on-off switches in the "Off" position, the water bath shall be subjected to a 60 Hz, AC potential, gradually increasing to 1240 volts, applied between the current-carrying parts and the metal enclosure, and maintained at maximum value for a period of 1 minute.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

4.4.3 Temperature stability. The thermoregulator shall be set so that the water in the bath attains the following temperatures: 37 ± 1°C. and 56 ± 1°C. Cover shall be in place and water shall be at constant level setting. Location of sensing portion of the temperature-detecting device shall be in accordance with service data submitted by the contractor. The following procedure shall be followed for 37°C. and 56°C. settings: After a stable temperature has been reached, temperature readings shall be recorded continuously until the pilot light has been energized 15 times. Maximum time interval between recorded readings shall be 10 seconds. Temperature-detecting device shall have a response time of 1 second (maximum) and shall indicate temperature changes of 0.01°C. (see 6.3). Response time shall be defined as the interval of time required for the temperature-detecting device to adjust to a change of 0.01°C.

4.4.4 Heating time. The water bath shall be filled with water 20°C. ± 1° and the cover put in place. The thermoregulator shall be set at the temperature of boiling water and heating system shall be turned on. The time for the water to reach boiling temperature shall be monitored and recorded.

4.4.5 Operating temperature range. Water and thermometer shall be placed in uncovered water bath in accordance with service data submitted by contractor. Thermoregulator shall be set at 60°C. and the heating system energized and allowed to heat until a constant temperature is reached (see 3.4.2). Thermoregulator shall be reset to the temperature of boiling water and the cover shall be put in place. Water bath shall be allowed to continue to heat until a constant temperature is reached (see 3.4.2). Ambient temperature shall be 20° ± 1°C.

## 5. PREPARATION FOR DELIVERY

### 5.1 Packaging.

5.1.1 Unit of issue. One water bath, as specified, constitutes one unit of issue.

5.1.2 Level A.

5.1.2.1 Unit preparation. The thermoregulator and thermometer shall be packaged in a plastic or spiral-wound chipboard container of suitable size. Container design shall include suitable interior fittings or shall contain sufficient cushioning material to insure against shifting of or damage to the contents. The thermoregulator and thermometer shall be secured within the water bath chamber. The water bath shall be wrapped in kraft paper and cushioned with corrugated fiberboard or other suitable material to insure against marring, scratching or damage to the unit.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

5.1.3 Level C. Units shall be packaged in standard commercial containers of the size and kind commonly used, which will afford the degree of protection required for shipment and use of the product for its intended purpose.

5.2 Packing.

5.2.1 Level A. Each unit, prepared as specified in 5.1.2.1, shall be packed in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-601, overseas type; PPP-B-621, class 2, or PPP-B-636, grade V3c. Closure and strapping shall be as specified in the appendix of the applicable box specification.

5.2.1.1 Case liner. Each Level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall conform to applicable paragraphs of appendix thereto. Case liner shall not be required for fiberboard boxes. Each fiberboard box shall be waterproofed in accordance with 30.4 of PPP-B-636.

5.2.2 Level B. Each unit, prepared as specified in 5.1.2.1, shall be packed in an exterior container designed for a type 3 load and constructed in accordance with PPP-B-601, domestic type; PPP-B-621, class 1; or PPP-B-636, type CF, class domestic. Closure shall be as specified in the appendix of the applicable box specification. Closure of fiberboard boxes shall conform to method II as specified in PPP-B-636.

5.2.3 Level C. The subject commodity shall be packed in substantial commercial containers of the type, size and kind commonly used for the purpose, so constructed as to insure acceptance and safe delivery by common or other carriers, at the lowest rate, to point of delivery called for in the contract or purchase order.

5.3 Marking. Marking shall be as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Procurement documents should specify the following:

a. Title, number and date of this specification.
b. National Stock Number (NSN).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

6.2 The test tube racks mentioned in 3.3 are listed in the Federal Supply Catalog as follows:

| National Stock No. | Item Identification |
|---|---|
| 6640-00-442-6450 | RACK, TEST TUBE, LABORATORY, Wassermann, Corrosion-Resisting Steel, 20-Tube |

Test tube racks are not required to be supplied with the water bath.

6.3 A temperature-detecting device which meets the response time and sensitivity requirements of 4.4.3 is Hewlett Packard, Model 2801-A, Quartz Thermometer.

6.4 The water bath covered by this specification is listed in the Federal Supply Catalog as follows:

| National Stock No. | Item Identification |
|---|---|
| 6640-00-765-0599 | WATER BATH, ELECTRIC, Precision, 6-Rack, 115 volt, 50-60 Hz, AC |

6.5 This specification does not cover all types, classes, grades or sizes of the commodity indicated by the title of this specification, or those which are commercially available, but is intended to cover the type normally procured to meet Military requirements.

Custodians:
Army - MD
Navy - MS
Air Force - 03

Review activities:
Army - MD
Navy - MS
Air Force - 03

Preparing activity:
Defense Supply Agency - DM

Project No. 6640-0908

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current DODISS.



Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

FOLD

POSTAGE AND FEES PAID
DEFENSE SUPPLY AGENCY
DoD - 304

U.S.MAIL

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

Commander
Defense Personnel Support Center
ATTN: DPSC-ATT
2800 South 20th Street
Philadelphia, PA 19101

FOLD

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

OMB Approval
No. 22-R255

INSTRUCTIONS: The purpose of this form is to solicit beneficial comments which will help achieve procurement of suitable products at reasonable cost and minimum delay, or will otherwise enhance use of the document. DoD contractors, government activities, or manufacturers/ vendors who are prospective suppliers of the product are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity.

DOCUMENT IDENTIFIER AND TITLE

MIL-W-37090 (DSA-DM)

NAME OF ORGANIZATION AND ADDRESS

CONTRACT NUMBER

MATERIAL PROCURED UNDER A

☐ DIRECT GOVERNMENT CONTRACT ☐ SUBCONTRACT

1. HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE?

A. GIVE PARAGRAPH NUMBER AND WORDING.

B. RECOMMENDATIONS FOR CORRECTING THE DEFICIENCIES

2. COMMENTS ON ANY DOCUMENT REQUIREMENT CONSIDERED TOO RIGID

3. IS THE DOCUMENT RESTRICTIVE?

☐ YES ☐ NO *(If "Yes", in what way?)*

4. REMARKS

SUBMITTED BY *(Printed or typed name and address - Optional)*

TELEPHONE NO.

DATE

**DD FORM 1 JAN 72 1426** REPLACES EDITION OF 1 JAN 66 WHICH MAY BE USED

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T09:57Z
Check the source to verify that this is the current version before use.