INCH-POUND
A-A-59953
15 May 2014

COMMERCIAL ITEM DESCRIPTION

INDICATOR, TEMPERATURE, THERMOCOUPLE

The General Services Administration has authorized the use of this commercial item description for all federal agencies.

1. SCOPE. This commercial item description (CID) describes a Indicator, Temperature, Thermocouple hereinafter referred to as a Thermometer Indicator used to measure electrical values or temperatures. The Thermometer Indicator consists of a high brightness digital display and associated keys for an intuitive user interface.

2. SALIENT CHARACTERISTICS. The following specifications are considered minimum Air Force requirements.

2.1 Digital display. The Thermometer Indicator shall have a digital display that is easily readable and displays temperature in Deg. C or Deg. F

2.2 Thermocouple types. The Thermometer Indicator shall accept six thermocouple types: J, K, T, R, S, and E.

2.3 Inputs. The Thermometer Indicator shall have dual inputs with wide measurement ranges.

2.4 Displays. The Thermometer Indicator shall display T1, T2, and T1-T2 simultaneously with back-light and includes Hold, Min/Max, Relative and Timer / Date Stamp functions.

2.5 Outputs. The Thermometer Indicator shall output real time measurement data with time stamp through the industry standard TIA-232-F (RS232) serial.

2.6 Case. The Thermometer Indicator shall have a hard carrying case.

2.7 Thermocouple Probes. The Thermometer Indicator shall come equipped with two Type K thermocouple probes.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any data that may improve this document should be sent to: WR-ALC - 404 SCMS/GUEEA, 235 Byron Street Suite 19A, Robins AFB, GA 31098-1611. Since contact information can change, you may want to verify the currency of this address information using the ASSIST Online database at https://assist.dla.mil/ .

AMSC N/A FSC 6685

Distribution Statement A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:31Z
Check the source to verify that this is the current version before use.

2.8 Dimensions. The Thermometer Indicator shall weigh 20 ounces or less with dimensions of 2 by 8 by 4 inches or less.

2.9 Operating conditions. The Thermometer Indicator shall have an operating temperature range and condition: 0º to 50º C at 0 to 85% relative humidity (R.H.) (non-condensing).

2.10 Range. The Thermometer Indicator shall have a range of – 200º C to 1760º C by use of the allowable type thermocouples with a resolution of 0.1 Deg. C for the lower part of the range and 1.0 Deg. C for the higher part of the range for the specific type thermocouples.

2.11 Accuracy. The Thermometer Indicator shall be accurate to 0.1% of reading plus 0.7º C for the lower part of the range and 0.1% of reading plus 1.4º C for the higher part of the range for each type thermocouple.

3. REGULATORY REQUIREMENTS.

3.1 Recycled, recovered, or environmentally preferable materials. Recycled, recovered, or environmentally preferable materials should be used to the maximum extent possible provided that the material meets or exceeds the operational and maintenance requirements, and promotes economically advantageous life cycle costs. The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with 23.403 of the Federal Acquisition Regulation (FAR). However, used, rebuilt, or refurbished items shall not be provided.

3.2 Green Procurement Program. Green Procurement Program (GPP) is a mandatory federal acquisition program that focuses on the purchase and use of environmentally preferable products and services. GPP requirements apply to all acquisitions using appropriated funds, including services and new requirements. FAR 23.404(b) applies and states the GPP requires 100% of EPA designated product purchase that are included in the Comprehensive Procurement Guidelines list that contains recovered materials, unless the item cannot be acquired: a) competitively within a reasonable timeframe; b) meet appropriate performance standards, or c) at a reasonable price. The prime contractor is responsible for ensuring that all subcontractors comply with this requirement.

4. PRODUCT CONFORMANCE PROVISIONS

4.1 Product conformance. The contractor shall certify that the product offered meets the requirements of this description and that the product conforms to the producer's own drawings, specifications, standards, and quality assurance practices and is the same product offered for sale in the commercial marketplace. Small businesses that cannot offer products with established commercial market acceptability may compete by submission of bid samples. (The government reserves the rights to require proof of such conformance before first delivery and thereafter as may be otherwise provided under the provisions of the contract).

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:31Z
Check the source to verify that this is the current version before use.

4.2 Operation and Maintenance Manuals. One copy of the commercial operation and maintenance manual(s), including calibration procedures and adjustments, as applicable, shall be shipped with the equipment. One additional copy of the operation and maintenance manual shall be shipped to the cognizant engineer at this address; Tiffany Morris, 404 SCMS/GUEEA, 235 Byron Street, Ste 19A, Robins AFB, GA 31098-1670.

4.3 Warranty. The Thermometer Indicator and related parts shall carry a one year, parts and labor, warranty from the date of purchase.

5. PACKAGING.

5.1 Preservation, packing, and marking shall be as specified in the contract or order.

6. NOTES.

6.1 Source of documents.

6.1.1 Department of Defense and Federal documents may be obtained at https://assist.dla.mil/ or from the Document Automation and Production Service, Bldg 4D (DPM-DODSSP), 700 Robbins Avenue, Philadelphia PA 19111-5094.

6.1.2 TIA documents may be obtained at http://tiaonline.org/ or from TIA, Inc., 1320 N. Courthouse Rd., Suite 200 Arlington, VA 22201

6.1.3 FAR may be obtained from the Superintendent of Documents, P.O. Box 371954, Pittsburgh, PA 15250-7954. Electronic copies of the FAR may be obtained from www.acquisition.gov/far/.

6.2 National Stock Number (NSN). The following NSN is associated with this CID:

6685-01-557-7514

6.3 Key Words.

Digital Display
High brightness
Relative Humidity
User Interface

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:31Z
Check the source to verify that this is the current version before use.

Custodians:
Air Force - 84

Preparing Activity:
Air Force - 84

Reviewers:
Air Force - 99

Agent:
Air Force - 99

(Project 6685-2014-002)

NOTE: The activities listed above were interested in this document as of the date of this document. Since organizations and responsibilities can change, you should verify the currency of the information above using the ASSIST Online database at https://assist.dla.mil/ .

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:31Z
Check the source to verify that this is the current version before use.