

REVIEW ACTIVITIES: ARMY - MU, MD
NAVY - MS
AIR FORCE - 03

USER ACTIVITIES: ARMY - ME, AS,
NAVY - OS
DSA - DM

FED. SUP CLASS
6640



MS36382-1



MS36382-2

PART NUMBER	DIMENSIONAL DATA mm		
	HEIGHT		OD body
	w/o cover	w/cover	
MS36382-1	26 ± 3.0	30 ± 3.0	50 ± 3.0
MS36382-2	93 ± 3.0	98 ± 3.0	60 ± 3.0

(A)

MATERIAL: Dish shall be of clear glass.

DESIGN: Dish and cover shall be of heavy molded glass.

Cover shall be of flat plate approximately 7 mm in thickness and shall have a recessed circular groove approximately 0.5 mm in depth extending around its periphery which shall be ground to fit the rim of the dish.

Dish shall be cylindrical in shape and the base shall be flat so as to provide a secure resting surface. The top edge shall be ground to permit a perfect sealing of body and cover.

NOTES:
1. For design feature purposes, this standard takes precedence over procurement document referenced herein.
2. Referenced document shall be of the issue in effect on date of invitations for bid.

APPROVED 1 DEC 1967 REVISED (A) 17 AUG 1970

This military standard is approved by the Department of Defense and is mandatory on all activities. Selection for all new engineering and design applications and for [illegible] use shall be made from this document.

P.A. MU Other Cust MS 03	TITLE DISH BIOLOGICAL PREPARATION stender design		MILITARY STANDARD MS 36382
Procurement Specification MIL-D-2207	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 SEP 63 672 (Coordinated) EDITION OF 1 SEP 57 MAY BE USED.

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:39Z
Check the source to verify that this is the current version before use.