A-A-53291
7 December 1987

COMMERCIAL ITEM DESCRIPTION

RETRACTOR PERIOSTEAL, (Prichard No. 3)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a double-ended, Prichard No. 3, Periosteal Retractor suitable for use in retracting oral tissues during oral or periodontal surgery.

Salient characteristics:

General requirements. The retractor shall be in accordance with the applicable requirements of GG-I-526 with the options and additional requirements specified herein.

Materials.

Operative ends. The operative ends shall be fabricated of class 4, martensitic stainless steel, type 440A, 440B, or 440C.

Handle. The handle shall be plated brass. Brass shall conform to class B copper alloy number 330, as specified in ASTM B 135 for chemical composition only. The brass handles shall be chromium over nickel plated, as specified for class 8 material in GG-I-526.

Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C52 to C55 for type 440A; C54 to C58 for type 440B; and C56 to C60 for type 440C. The handles and shanks shall be of the proper temper and shall have a hardness suitable for the intended purpose. Hardness shall be determined as specified herein.

Design and dimensions. The retractor shall be a double-ended, Prichard No. 3, Periosteal Retractor suitable for use in retracting oral tissues during oral or periodontal surgery. The design and dimensions of the retractor shall conform to the latest revision of Defense Personnel Support Center (DPSC) Drawing No. 24436.

AMSC N/A FSC 6520

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Construction. The retractor shall be of three-piece construction (one operative end at each end and a handle). Knurled or serrated portions of the handle shall provide adequate finger grip during use. The operative ends shall be permanently secured to the handle.

Finish. The spoon-shaped operative end shall have a fine honed finish on the inside face and a mirror or bright polished finish on the back face. The flat, spatula-type operative end shall be mirror or bright polished with smoothly rounded peripheral edges. Handles shall have a mirror or bright finish, except for the serrated or knurled portion. The serrated or knurled portion of the handle shall provide adequate finger grip during use. The interface of the shanks and handle shall be smoothly blended with no crevices evident.

Corrosion resistance. The copper sulfate test requirements of GG-I-526 shall apply to the operative ends. The boil test requirements of GG-I-526 shall apply to the entire instrument.

Identification marking. Each instrument shall be permanently and legibly marked with the instrument identification which shall appropriately identify the instrument. All marking shall be on the smooth portion of the handle.

Workmanship. In addition to the requirements of GG-I-526, the operative ends shall be free of discoloration, deep pits, coarse grind marks, cracks and burrs. The serrated or knurled portion of the handle shall be free of barbs. The entire instrument shall be free of defects which detract from its appearance or impair its serviceability, proper functioning, or intended use.

Unit. Each (EA). One retractor, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Quality conformance inspection. The examinations and tests required to assure conformance of those items or lots of items to be offered for acceptance are classified as specified in tables I, II, and III.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105. The inspection levels and acceptable quality levels (AQL's) shall be as indicated in table I. The unit of product for examination purposes shall be one instrument.

TABLE I. Sampling for examination.

Type defect	Inspection level	AQL (percent defective)
Major	II	1.0
Minor	II	4.0

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as indicated in table II. The unit of product for test purposes shall be one instrument.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TABLE III. Sampling for tests.

Component	Characteristic	Inspection level	AQL
Entire instrument	Composition	S-1 *	**
Operative end	Hardness	S-2	1.5
Operative end	Copper sulfate test	S-2	1.0
Entire instrument	Boil test	S-2	1.0

*A sufficient amount of each lot of metal used in the construction of the retractors may be selected for composition analysis in lieu of sample retractors, provided lot identity has been maintained and can be demonstrated to the Government.

**Acceptance number shall be zero for all sample sizes.

Examination. The retractors shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with the following classification of defects. Examination shall include, but shall not be limited to, the defects listed in table III.

TABLE IV. Classification of defects.

Categories	Defects
Major	
101	Design not as specified for the applicable style number of the instrument.
102	Finish of operative end not as specified.
Minor	
151	Finish of handle and shank not as specified.
152	Retractor not free from dirt, oil, polishing compound, and foreign matter.
	Identification marking not present, not properly located, illegible, or incomplete.
	Handle serrations or knurling missing, not uniform, or not complete.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Tests.

Hardness. Hardness shall be determined in accordance with American Society for Testing and Materials (ASTM) Designation E18 (Rockwell) or E92 (Vickers or Diamond Pyramid).

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch/pound units, providing they fall within the tolerance specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerance specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing, shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document.

Stock Number	Item Identification
6520-01-093-0166	RETRACTOR, PERIOSTEAL, Prichard No. 3: Doubled-ended

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be as specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

AGENT:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

PROJECT NO. 6520-2044

49066

Location: NCR Disk 2

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6. REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.