A-A-53293A
22 July 1988
SUPERSEDES
A-A-0053293(DM)
7 December 1987

COMMERCIAL ITEM DESCRIPTION

CONTROL STRIPS, VISUAL URINALYSIS

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a reactive control strip that is reconstituted to provide positive reactions with visual urinary test strips.

Salient characteristics.

Shall be a control strip with six (6) impregnated pads (reactive reagent areas) that, when placed in solution, shall yield positive results when tested with all standard urinary test strips and color charts (excluding those for phenol ketones).

Shall have reactive reagents formulated to be reconstituted in 12 milliliters (ml) of distilled water (in accordance with manufacturer's instructions) to provide a solution that will give expected test results for bilirubin, blood, glucose, ketones, nitrite, pH, protein, and urobilinogen.

Shall have control strip contain six impregnated pads, of which the ingredients are:

Protein - bovine serum albumin.
Glucose.
Ketones - aluminum acetylacetonate.
Bilirubin - crystalline bilirubin.
Blood - bovine hemoglobin.
Nitrite - sodium nitrite.
Urobilinogen - p-phenylenediamine.

AMSC N/A FSC 6550

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Shall be supplied with a package insert that includes the following data:

Instructions for preparing a reconstituted solution.
Expected test results for bilirubin, blood, glucose, ketones, nitrite, pH, protein, and urobilinogen.
Instructions for proper storage of a control strip.

Workmanship. The control strip and package insert shall be free from defects which detract from their appearance or impair their serviceability.

Unit. Bottle (BT). One bottle containing twenty-five control strips, together with a package insert, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Inspection and testing. Inspection and testing shall be conducted in accordance with commercial practice to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Sampling for inspection and testing. Sampling for inspection and testing shall be in accordance with commercial practice.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U. S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its contractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6550-01-093-4010

Ordering data. Expiration dating period required, intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

A-A-53293A

MILITARY INTERESTS:

Custodians:
Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:
DoD - MB

Agent:
DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

PHS
FDA-MPQAS

Project No. 6550-0077

Location: Disk #12

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.