A-A-53298A
7 July 1989
SUPERSEDING
A-A-0053298(DM)
5 February 1988

COMMERCIAL ITEM DESCRIPTION

SCOOP, DUCT, COMMON (Mayo)

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers a Mayo Common bile scoop for removing gallstones from the common bile duct.

Salient characteristics.

General. Shall be a Mayo common duct scoop designed to remove gallstones from the common bile duct.

Material. The shaft shall be fabricated from silver plated malleable copper. The handle and distal tip shall be fabricated from silver plated brass.

Style and design. The scoop shall be in accordance with Defense Personnel Support Center Drawing Number 23427.

Finish. The scoop shall have a mirror finish of silver plating of 0.001 inch minimum thickness. Its surface shall be smooth and free from blisters, porosity and edge build-up. The finish shall conform to the finish performance requirements.

Corrosion resistance. The scoop shall meet the boil test requirements specified in ASTM Designation F1089 for Corrosion Testing of Surgical Instruments.

Finish performance. The silver plating of the stem shall exhibit no flaking separation or any visible indication of loss of adherence after being subjected to a 90 degree bend with a 1/4 inch radius.

Identification marking and labeling. The scoop shall be marked with the name or registered trademark of the manufacture and the word "stainless". If the gallstone scoop is of foreign manufacture, the instrument shall bear the name of the country of manufacture.

AMSC/NA FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:17Z
Check the source to verify that this is the current version before use.

Workmanship. The gallstone scoop shall be free from defects which detract from its appearance or impair its serviceability.

PRODUCT QUALITY LEVEL

Visual and dimensional examination. Shall be examined to verify that the design, construction and workmanship are in accordance with applicable requirements.

Materials. The supplier shall be responsible for performing tests to assure that materials are in accordance with requirements specified.

Unit. Each (EA). One gallstone scoop, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribe requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for visual examination shall be in accordance with MIL-STD-105. The unit of product for examination shall be one gallstone scoop. Sampling shall be level II with an AQL of 1.0.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the currentversion before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified herein. The unit of product for test purposes shall be one gallstone scoop. AQL shall be 1.0 percent defective. Inspection level shall be S-2.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder, In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:17Z
Check the source to verify that this is the currentversion before use.

Recovered materials. The offeror/contractor is encouraged to use recovered materials in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

NSN	ITEM IDENTIFICATION
6515-01-139-8171	SCOOP, DUCT, COMMON, MAYO, SMALL
6515-01-139-8172	SCOOP, DUCT, COMMON, MAYO, MEDIUM
6515-01-139-8173	SCOOP, DUCT, COMMON, MAYO, LARGE

Ordering data Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-4556

☆U.S. GOVERNMENT PRINTING OFFICE: 1989 - 604-033/13781

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.