A-A-0053299(DM)
30 November 1987

INTERIM COMMERCIAL ITEM DESCRIPTION

BAG, BIOHAZARD DISPOSAL

The General Services Administration has authorized the use of this Interim Commercial Item Description which was developed by the Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information.

This Interim Commercial Item Description covers a heavy duty polypropylene disposable bag suitable for autoclaving biologically contaminated material prior to disposal.

Salient characteristics:

The biohazard disposal bag shall be a heavy-duty, puncture-resistant, polypropylene bag capable of withstanding autoclave temperatures up to 300°F (148°C) without softening, melting, or breaking. The bag shall be of tubular construction to prevent splitting after autoclaving and shall have a single or double bottom heat seal. Seams running along the length of the bag are not acceptable.

The bag shall have the following nominal dimensions:

18 inches (45.72 cm) wide
26 inches (66.04 cm) long
0.002 inch (2.0 mils) thick

The bag shall be labeled in a fluorescent orange or orange-red color to signify a biohazard. A biological hazard warning symbol in accordance with OSHA regulation 1910.145 shall be used on the bag. The symbol design shall be a fluorescent orange or orange-red and approximately 4 inches square, including a black background.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

The symbol shall be located within an orange or orange-red rectangle a minimum of 9 inches by 11 inches in size, and containing the following minimum words and letter heights:

Word	LETTER HEIGHT
Caution	1/4 inch
Biohazard bag	5/8 inch
For autoclaving and disposing of hazardous Contaminated materials	5/32 inch 1/2 inch

Within the rectangle, a light colored space approximately 6 inches by 1-1/2 inches shall be provided for writing identification and instructions.

As an alternate, an all orange-red bag shall be acceptable, providing the above symbol, wording, and writing space requirements are met. The color shall not be faded, inconsistent, or incomplete.

Each package of biohazard bags shall be supplied with enough wire reinforced ties to secure each bag.

The bag shall not be ripped, torn, or otherwise unusable, the bag shall be clean.

Workmanship. The biohazard disposal bag shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing 200 biohazard disposal bags, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for sampling purposes shall be two biohazard disposal bags as specified in this description.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred units)
For visual examination		
Major defects	II	1.0
Total defects (Major and minor combined)	II	4.0
For dimensional examination	S-1	2.5

Examination. The biohazard disposal bag shall be examined to determine compliance with all requirements contained in this specification.

Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects below.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

TABLE II. Categories and Defects.

Categories	Defects
Major	
101	Bag is not a heavy-duty, puncture resistant bag, 2.0 mils thick.
102	Bag not of tubular construction.
103	Bag not heat sealed on bottom.
104	Bag not 18 inches by 26 inches.
105	Bag not labeled in accordance with OSHA regulation 1910.145.
106	Bag is ripped, torn, or otherwise unusable.
107	Wire reinforced bag tie not supplied.
Minor	
201	Symbol letter heights not correct.
202	Bag is not clean.

Tests.

Autoclave Test.

Filled bag shall be closed with wire reinforced tie and autoclaved at 300°F (148°C) minimum. After normal autoclaving is completed, bag shall be emptied and inspected for damage. Any part of the bag softening, melting or breaking shall constitute a major defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:17Z
Check the source to verify that this is the current version before use.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6530-01-107-5799

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1934

PGC 58336

Location: CID/BAG/H13-25

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.