INCH-POUND

A-A-53299B
31 January 1994
SUPERSEDING
A-A-53299A
30 May 1989

COMMERCIAL ITEM DESCRIPTION

BAG, STERILIZATION-BIOHAZARD

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers heavy duty polypropylene disposable bags suitable for autoclaving biologically contaminated material prior to disposal.

Salient characteristics:

The biohazard disposal bag shall be a heavy-duty, puncture-resistant, polypropylene bag capable of withstanding autoclave temperatures up to 300°F (148°C) without softening, melting, or breaking. The bag shall be of tubular construction to prevent splitting after autoclaving and shall have a single or double bottom heat seal. The bottom heat seal shall not allow separation, tearing, or leakage of contents. Seams running along the length of the bag are not acceptable.

The bags shall be supplied in the following sizes:

Size 1 - 18 inches wide x 26 inches long.

Size 2 - 24 inches wide x 36 inches long.

The minimum thickness shall be 0.002 inch.

AMSC N\A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

The bag shall be labeled in a fluorescent orange or orange-red color to signify a biohazard. A biological hazard warning symbol in accordance with OSHA regulation 1910.145 shall be used on the bag. The symbol design shall be fluorescent orange or orange-red and approximately 4 inches square, including a black background.

The symbol shall be located within an orange or orange-red rectangle a minimum of 9 inches by 11 inches in size, and containing the following minimum words and letter heights:

WORD	LETTER HEIGHT
Caution	1/4 inch
Biohazard bag	5/8 inch
For autoclaving and disposing of hazardous	5/32 inch
Contaminated materials	1/2 inch

Within the rectangle, a light colored space approximately 6 inches by 1½ inches shall be provided for writing identification and instructions.

As a alternate, an all orange-red bag shall be acceptable, providing the above symbol, wording, and writing space requirements are met. The color shall not be faded, inconsistent, or incomplete.

Each package of biohazard bags shall be supplied with enough wire reinforced ties to secure each bag.

The bag shall not be ripped, torn, or otherwise unusable, the bag shall be clean.

Workmanship. The biohazard disposal bag shall be free from defects which detract from its appearance or impair its serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Unit. Package (PG). One package containing 200 biohazard disposal bags of one size only, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Source: https://assist.dla.mil-- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be two biohazard disposal bags as specified in this description.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 with an AQL of 2.5 (percent defective) and an inspection level of S-1. The unit of product for test purposes shall be one biohazard bag.

First article. Where a first article sample is required, all tests shall be performed on that unit.

Acceptance testing. When a first article sample has not been required, one unit shall be subjected to the performance test specified herein. If a test article sample has successfully passed all tests, performance test shall not be required at acceptance.

Tests.

Autoclave Test. Filled bag shall be closed with wire reinforced tie and autoclaved at 300°F (148°C) minimum. After normal autoclaving is completed, bag shall be emptied and inspected for damage. Any part of the bag softening, melting or breaking shall constitute a defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

NOTE: The following National Stock Numbers are covered by this document.

NSN	ITEM IDENTIFICATION	SIZE
6530-01-107-5799	BAG, STERILIZATION-BIOHAZARD	1
6530-01-107-5798	BAG, STERILIZATION-BIOHAZARD	2

MILITARY INTERESTS:

Custodians:

Air Force-03
Army -MD
Navy-MS

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

FDA-MPQAS
USPHS
VA-OSS

Project Number: 6530-2427

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3, and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6, and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I. RECOMMEND A CHANGE:	1. DOCUMENT NUMBER A-A-53299B	2. DOCUMENT DATE (YYMMDD) 31 January 1994

3. DOCUMENT TITLE
BAG, STERILIZATION-BIOHAZARD

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME DEFENSE PERSONNEL SUPPORT CENTER ATTN: DPSC-MSE	b. TELEPHONE (Include Area Code) (1) Commercial: AC 215-737-2870 (2) AUTOVON: AV 444-2870
c. ADDRESS (Include Zip Code) 2800 SO. 20TH STREET PHILADELPHIA, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT: Defense Quality and Standardization Office 5203 Leesburg Pike, Suite 1403, Falls Church, VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340

DD Form 1426, OCT 89 — Previous editions are obsolete. — 9829/6

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:17Z
Check the source to verify that this is the current version before use.