A-A-53316
26 February 1988

# COMMERCIAL ITEM DESCRIPTION

## BOOTH, AUDIOMETRIC EXAMINATION

The General Services Administration has authorized the use of this commercial item description.

This commercial item description covers an audiometric examination booth suitable for use in the testing, calibration, and recording of aural acuity in medical treatment facilities.

Salient characteristics:

The audiometric examination booth shall be a prefabricated testing room, of one size and type as indicated below, suitable for use in the testing, calibration, and recording of aural acuity in medical treatment facilities.

The booth shall be so designed that its unassembled components can be readily handled and stored without special handling equipment and can be stored in minimum space, and that it can be disassembled and re-erected at least three times without adversely affecting its sound-excluding capabilities, and that the erected room (booth) can be moved from one location to another by means of a fork lift without disassembling.

The audiometric examination booth shall be of such design as to provide a room (booth) meeting the sound exclusion requirements as follows:

a. Sound Reduction Qualities shall be in accordance with ASTM Standard C423-77:

| Octave Bands, Hz | Sound Absorption Coefficient |
|---|---|
| 250 | 0.43 |
| 500 | 0.78 |
| 1000 | 0.96 |
| 2000 | 0.95 |
| 4000 | 0.95 |
| NRC | 0.90 |

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

b. Noise reduction shall be in accordance with ASTM Standard E596:

| Octave Bands, Hz | 125 | 250 | 500 | 1K | 2K | 4K | 8K | NIC |
|---|---|---|---|---|---|---|---|---|
| Noise Reduction (db) | 23 | 33 | 43 | 51 | 54 | 54 | 53 | 46 |

The booth shall have the following dimensions:

| | |
|---|---|
| Inside | 38 to 40 in. wide x 36 in. long x 78 in. high. |
| Outside | 46 to 48 in. wide x 42 to 44 in. long x 86 to 90 in. high. |
| Weight | 1500 lbs max. |

The booth shall have one door with net clear openings of 30 to 33 in. by 73 to 76 in. Door shall be left or right hand swing, as specified by requisitioner, and magnetically sealed. Door shall swing outward.

The floor shall float on properly loaded vibration isolators.

There shall be one acoustic glass window not less than 24 x 30 inches, mounted in the center of one side panel, the side panel away from the door hinge.

The following accessories shall be provided:

Floor mat: Shall be provided with a carpet.

Jack panel: A completely wired jack panel shall be provided. It shall include as a minimum the following connections:

Size 3-wire phone jacks
One cinch jones 303 connector
One cinch jones 304 connector

The jack panel shall not degradate performance of the booth.

Lighting and electrical: Shall be provided with at least one incandescent light fixture and one duplex outlet.

Ventilation System: Shall be provided with a ventilation system with at least a 100 CFM rate. Shall have intake and exhaust silencers.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

It is essential that the requisitioner communicate to the DPSC his choices on 3 options:

a. Whether door shall be hinged left or right.

b. Whether booth shall be installed by the manufacturer.

c. Whether 110 VAC/60 Hz or 220 vac/50Hz shall be the power requirements.

The following permanent marking shall be displayed on the booth doors:

This booth must be located in sufficient quiet area so that the following sound pressures within the booth are not exceeded:

| Test tone frequency (Hz) | 125 | 250 | 500 | 1000 | 2000 | 4000 | 8000 |
|---|---|---|---|---|---|---|---|
| Sound Pressure Levels (db) | 34.5 | 23.0 | 21.5 | 29.5 | 34.5 | 42.0 | 45.0 |

Workmanship. The audiometric examination booth shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One audiometric examination booth, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one audiometric booth.

Defects.

Does not assemble into a structurally sound and integral booth.

Window not acoustically sealed.

Door does not open and close smoothly and/or acoustically seal.

Jack panel circuit and circuit insulation defective.

Identification marking not complete, permanent, and legible.

Finish not free of chips, cracks, scratches, and other appearance defects.

Any component not secured in place.

Sampling for tests. Sampling for tests (dimensional) shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and inspection level of S-4. The unit of product for test purposes shall be one audiometric booth.

Tests. Reports from a recognized and approved laboratory shall be supplied and include the following information:

1. The audiometric examination booth has been tested in accordance with ASTM Standards E596, E90 and E336 and meets specified noise reduction values.

2. The audiometric examination booth has been tested in accordance with ASTM Standard C423-77 and meets specified sound absorption coefficients.

Testing shall have been performed within five years of date of solicitation.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6515-00-065-9600

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

A-A-53316

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3964

PGC

Location: CID/BOOTH/H17-12



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

**INSTRUCTIONS:** In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

*NOTE.* This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

*(Fold along this line)*

*(Fold along this line)*

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

# STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

*(See Instructions – Reverse Side)*

| 1 DOCUMENT NUMBER | 2 DOCUMENT TITLE |
|---|---|
| | |

| 3a. NAME OF SUBMITTING ORGANIZATION | 4 TYPE OF ORGANIZATION *(Mark one)* |
|---|---|
| | ☐ VENDOR |
| b ADDRESS *(Street, City, State ZIP Code)* | ☐ USER |
| | ☐ MANUFACTURER |
| | ☐ OTHER *(Specify)* _____ |

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

| 7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional | b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional |
|---|---|
| c MAILING ADDRESS *(Street City State, ZIP Code)* – Optional | 8 DATE OF SUBMISSION *(YYMMDD)* |
| | |

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.