A-A-53317
28 December 1987

COMMERCIAL ITEM DESCRIPTION

CENTRIFUGE, LABORATORY

The General Services Administration has authorized the use of this commercial item description.

This Commercial Item Description covers a table model variable speed clinical centrifuge with a protective guard bowl and a four place horizontal head for 15 milliliter (ml) tubes.

Salient Characteristics:

Shall be a table model centrifuge for small volume centrifuging on 115 Volt, 50/60 Hz AC supply lines in military facilities having limited space. Shall have maximum overall dimensions of 18 long by 18 wide x 18 high (inches).

Shall be fabricated of stainless steel or aluminum and consists of a base, guard bowl, hinged cover, universal type motor, speed control device and line cord.

Shall be supplied with a detachable four (4) place horizontal head for 15 ml tubes, and a corrosion-resistant steel shield with rubber cushions for each trunnion ring.

The base shall support or contain the guard bowl, and shall house the motor and speed control device. The bottom of the base shall be provided with flexible rubber feet for cushioning and stability.

The guard bowl shall provide a chamber for the detachable head and tubes that are inserted for horizontal centrifuging. The interior of the bowl shall be finished with a durable baked-on enamel or stainless steel. The cover shall be hinged to the unit. The center of the cover shall have a cutout (hole) which allows for safe use of a speed counter device (which is not required to be supplied).

The motor shall be a universal speed type that operates on 115 volt, 50/60 Hz, AC supply lines. The head shall be capable of being secured to the motor shaft or a motor shaft adapter.

AMSC N/A FSC 6640

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:18Z
Check the source to verify that this is the current version before use.

The speed control device shall be an autotransformer or rheostat which provides for six or seven steps of speed variation. The speed selection or variation shall be through use of an accessible selector knob or handle control that includes an "off" position.

Shall be capable of obtaining at the tips of the tubes the following relative centrifugal forces (RFC).

A maximum of 400 gravities at the lowest speed setting.
A minimum of 1200 gravities at the top speed setting.

The line cord shall be a three-wire electrical cord that terminates with a non-detachable 3-blade male plug that is suitable for use with receptacles specified in American National Standards Institute (ANSI) C73-11 (Dimensions of Plugs and Receptacles - General Purpose, 125 Volts, 15 Amperes, 2-Pole, 3-Wire, Grounding Type). The exposed length of the line cord shall not be less than six feet.

Workmanship. Centrifuge shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One centrifuge, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contact or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Sampling for tests. Sampling for tests shall be conducted in accordance with commercial practice. The unit of product for test purposes shall be one centrifuge.

Source: https://assist.dla.mil -- Downloaded: 202640-26T09:18Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements.

Operational test. An operational test shall be performed on all centrifuges to determine that the centrifuges and their controls are capable of proper operation.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric Products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No. 376, and all other requirements of this document are met. If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch-pound units, a request should be made to the contracting officer to determine if the product is acceptable. The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:18Z
Check the source to verify that this is the current version before use.

A-A-53317

NOTE: The following National Stock Number is covered by this document:

Stock Number	Item Identification
6640-00-494-3822	CENTRIFUGE, LABORATORY

Ordering data. Intermediate/exterior package quantities, labeling and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITY:

VA-OSS
PHS
FDA-MPQAS

PGC 56469

Project No. 6640-1412

Location: NCR DISK # 12

Source: https://assist.dla.mil -- Downloaded: 202610-26T09:18Z
Check the source to verify that this is the current version before use.