A-A-53318
26 February 1988

COMMERCIAL ITEM DESCRIPTION

MARKER SET, X-RAY FILM IDENTIFICATION

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal Specification GG-M-125.

This Commercial Item Description covers an x-ray film marker set consisting of letters, numerals, holders, and a storage box, suitable for use in identifying x-ray films.

Salient characteristics:

The x-ray film identification marker set shall consist of the following:

1. Letters and numerals.

   a. Fifteen sets of alphabet, A through Z.
   b. Twenty each of letters R and L.
   c. Forty complete sets of numerals, 0 through 9.
   d. Twenty-five each of letters A through G, J, L through P, R through V, and Y.

2. Two channel aluminum mounting holder. (2 required).

3. Compartmented storage box.

Letters and numerals (characters)

The letters and numerals shall be flat and gothic style. The letters shall be capital letters. All characters shall be 1/4 inch ± 1/64 inch high. Height measurements shall be determined by measuring across the face of the characters.

The letters and numerals shall be 0.039 inch ± .005 inch thick and of a width suitably proportionate to their height. They shall be stamped or molded of lead. The total tin, arsenic, antimony, and copper content of the lead shall not exceed 13 percent.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

The letters and numerals shall be completely and securely enclosed in formed or molded transparent plastic of the non-warping and slow burning type. The use of adhesives to bond the plastic to the characters is not acceptable. The plastic shall not split, crack, or separate from the letters or numerals when the character is bent approximately 15° or dropped.

All characters shall be vertical within 1/64 inch from absolute vertical. From an established line, approximately 3/16 inch from the bottom of the plastic, all numerals shall be within 1/32 inch from absolute horizontal.

The overall size of the plastic/lead letter or numeral combination shall be 0.625 inch ± 0.010 inch high, 5/16 inch ± 1/32 inch wide and 0.060 inch ± 0.010 thick. The letters and numerals shall be mounted within 1/32 of the centerline of the plastic.

The characters shall be clean and free of grease, oil, dirt, and other foreign matter.

Two channel aluminum mounting holder

The channels of the holder shall be designed to hold the plastic-backed characters and to permit them to slide into place easily.

The holder shall be a nominal 4-1/2 inches long and shall not be more than 1-7/16 inches wide.

The holder shall be fabricated of sheet aluminum, or as an aluminum extrusion, a minimum of 0.018 inch thick. There shall be two channels, each arranged to hold the characters. The double channel holder shall be one piece construction or shall consist of two single channel holders spot welded together.

Any spot welds shall be sound and free from pits, deformities, or corrosive areas.

The holder shall be free of sharp corners, sharp edges, cracks, or deformities.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Compartmented storage box

The storage box shall be fabricated of wood, metal, or plastic and shall have a hinged cover. There shall be 36 compartments for letters and numerals and a space for two double-channel holders. When closed the nominal overall dimensions of the box shall be as follows:

Length - 9-1/2 inches
Width - 12-1/2 inches
Height - 2 inches

The compartments shall be approximately 1-3/8 inch by 1-3/8 inch. The holder area shall have a width of 2-1/4 inches minimum and shall be large enough to accommodate the holder. Construction of the compartments shall be sturdy, and shall be securely fastened, to prevent accidental removal of partitions.

The box shall have a hinged lid lined with sponge rubber or polyurethane foam, cemented in place, and so built that when the lid is closed, it will not be possible for lead letters or numerals to move to adjacent compartments when the box is turned upside down or shaken in any position.

The lid shall open approximately 105° and shall be so designed that it will remain open when placed in that position. The hinges shall be sturdy and firmly mounted on either the short or the long side of the box. A spring-action catch shall be provided to close the lid.

The storage box shall be free from burrs, slivers, sharp edges, or deformities.

The box shall be clean, and free of grease, oil, dirt, or other foreign material.

Workmanship. The x-ray film identification marker set shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Set (SE). One marker set, as specified, constitutes one unit.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one storage box, one holder, five letters and five numerals as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Defects.

Unit of product - Letters and numerals

Characters not flat, Gothic style. Letters not capital.

Characters not completely and securely enclosed in formed or molded transparent plastic.

Mounted letters and numerals do not slide easily into appropriate holder.

Mounted letters and numerals not centered and vertical in plastic backing.

Unit of product - Holders

Holder not of design specified.

Holder not fabricated of aluminum, a minimum of 0.018 inch thick.

Holder not free of sharp corners, sharp edges, cracks, or deformities.

Spotwelds on holder not sound and not free from pits, deformities, or corrosive area.

Unit of product - Storage box

Storage box not compartmented as specified.

Letters and numerals move to adjacent compartments when storage box is turned upside down and shaken.

Lid not lined with sponge rubber or polyurethane foam.

Lid does not remain open at approximately 105°.

Spring-action lid catch not provided.

Box not free from burrs, slivers, sharp edges, or deformities.

Box not clean, and free of grease, oil dirt, or other foreign material.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-3. The unit of product for tests purposes shall be one storage box, one holder, five letters and five numerals.

Tests.

Bend test

Five letters and five numerals shall be selected at random for each set selected. Each end of the plastic encased character shall be bent approximately 15°. The plastic shall not split, crack or separate from the character and there shall be no other damage to the enclosed character or the plastic enclosure. Any damage to the character or plastic shall constitute a major defect.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-00-608-0620

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

A-A-53318

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6525-1223

PGC 49608

Location: CID/MARKER/H13-28

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b Recommended Wording

c Reason/Rationale for Recommendation

6 REMARKS

7a NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street City, State ZIP Code)* – Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil - Downloaded: 2026-01-26T09:18Z
Check the source to verify that this is the current version before use.