A-A-53324
26 February 1988

COMMERCIAL ITEM DESCRIPTION

HOLDER, RADIOGRAPHIC FILM CASSETTE

The General Services Administration has authorized the use of this Commercial Item Description in preference to Federal specification GG-H-590.

This Commercial Item Description covers a wall-mounted, vertical type cassette holder suitable for holding standard radiographic film cassettes and grids.

Salient characteristics:

The radiographic film cassette holder shall be a vertical type holder designed for wall mounting and suitable for holding standard radiographic film cassettes from 5 by 7 inches to 14 by 17 inches, inclusive, and a standard radiographic 14 by 17 inch stationary grid.

The holder shall have approximate dimensions of 38 inches high, 16 inches wide, and 4 inches deep. The holder shall consist of at least the following:

a. Mounting base
b. Carriage tracks
c. Cassette carriage assembly
d. Adjustable cassette adapter
e. Chinrest
f. Height indicator

The mounting base shall consist of one or more castings or rigid sheet metal plates with suitably placed holes to permit its being fastened to a wall.

The carriage tracks shall be fabricated from corrosion-resisting metal. They shall be smooth, parallel, and of sufficient length to permit a range of motion of at least 15 inches with a 14 inch by 17 inch cassette mounted in the cassette carriage with its long axis vertical.

AMSC N/A FSC 6525

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

The cassette carriage shall be reinforced with steel and shall move freely and smoothly in a vertical direction. There shall be a locking device to permit locking the cassette carriage at any point in its vertical travel. The cassette carriage shall be adjustable to hold and lock in place any standard type cassette, with the cassette in either the horizontal or vertical position. The cassette carriage shall be marked to facilitate centering of cassettes. The front of the cassette carriage shall be open so that the cassette is always in view as an aid to centering.

The adjustable cassette adapter shall be designed so that any standard size cassette may be used with a standard 14 by 17 inch grid. The adapter shall be fabricated of aluminum or sheet steel. It shall have a clamping device for holding the cassette, and a lock device, such as a knurled thumbscrew, to lock the cassette in place.

The chinrest shall be shaped to the contour of the chin so as to facilitate positioning of the patient in chest radiography. It shall be permanently fastened to the top edge of the cassette carriage. The chinrest shall be made of cast aluminum or other suitable material, shall be highly polished, and shall have no sharp corners or edges.

The height indicator shall be graduated in inches and shall be constructed so as to become an integral part of the cassette holder. It shall be interchangeable for right or left side mounting, and shall be adjustable. The indicator shall be fabricated of sheet steel or other suitable material.

The holder shall be degreased, primed, and finished with smooth enamel in the manufacturer's standard color. Trimming shall be chrome or other suitable corrosion-resistant material. The finish shall not fade, crack, blister, or peel under normal conditions of use.

Any welding shall be durable, sound in every detail, free of cracks, porosity, and other defects, and in accordance with the latest practices recommended by The American Welding Society.

The holder shall be free of burrs and sharp edges. It shall be clean and free of oil.

All holder material shall be high grade throughout and of sufficient strength to maintain operational alignment for the life of the equipment.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Instructions shall be furnished by the manufacturer for installation, operation, and maintenance.

Workmanship. The radiographic film cassette holder shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Each (EA). One radiographic film cassette holder, as specified, constitutes one unit.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 with an AQl of 1.0 (percent defective) and an inspection level of II. The unit of product for examination purposes shall be one film cassette holder as specified in this description.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Defects.

Holder not complete with all specified component parts.

Cassette carriage fails to slide smoothly and freely over the entire length of its travel.

Locks on either cassette carriage or adjustable cassette adapter fail to fix or release at any set position.

Adjustable cassette adapter fails to slide smoothly and freely over the length of its travel.

Channels in upper and lower horizontal members of carriage assembly, including cassette adapter, not properly located and aligned to accommodate cassette and grid.

Component parts not free of cracks and burrs or sharpness at edges.

Plating not free of blistered, excessively scratched or unplated areas.

Cassette carriage adjustment not within range required to hold any standard type cassette in vertical or horizontal position.

Height indicator scale not present or not graduated as specified.

Height indicator scale not adapted for right-side or left-side mounting.

Cassette carriage not marked to facilitate centering cassettes.

Finish not smooth, adherent and free of chipped, cracked, blistered, or bare areas.

Holder not clean and free of oil.

Sampling for tests. Sampling for tests (dimensional) test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-1. The unit of product for tests purposes shall be one radiographic film cassette holder.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Number is covered by this document:

6525-00-299-8300

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1219

PGC 49480

Location: CID/HOLDER/H13-23

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.