A-A-53326
26 February 1988

COMMERCIAL ITEM DESCRIPTION

ENVELOPE, STERILIZATION

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a see-through type laminated plastic and paper surgical instrument sterilization envelope with sterilization indicators.

Salient characteristics:

The sterilization envelope shall be a continuous roll of sterilization wrap suitable for wrapping surgical instruments or other items for sterilization purposes. The envelope shall be supplied in the following approximate sizes:

Size 1 - 3 inches wide by 100 ft. long (7.6 cm by 30.5m) roll

Size 2 - 6 inches wide by 100 ft. long (15.2 cm by 30.5m) roll

Size 3 - 9 inches wide by 100 ft. long (22.9 cm by 30.5m) roll

The sterilization envelope shall be composed of transparent laminated film and mylar with surgical kraft paper.

The envelope shall be suitable for steam or gas sterilization. Each sterilization envelope shall contain continuous or discontinuous, but frequently repeated, chemical indicator marks printed on the outside and inside of the envelope. These indicators shall provide visual indication that exposure to the sterilization process (steam or gas) has taken place. The steam indicator shall not change fully when exposed in the ethylene oxide process. The ethylene oxide indicator shall not change fully when exposed in the steam sterilization process.

The envelope shall be suitable for heat sealing. It shall be easy to open aseptically when it is sealed correctly.

AMSC N/A FSC 6530

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil-- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Both the paper and plastic surface of the sterilization envelope shall be sufficiently permeable to permit contents to be sterilized via steam and ethylene oxide, and sufficiently impermeable after sterilization so that contamination from the atmosphere shall not enter the closed envelope.

Both the paper and plastic surfaces of the envelope shall be strong enough to minimize wrapper perforation. The film surface shall be clear for easy product identification.

The sterilization envelope shall be free of tears, cuts, holes, perforations, grease, dirt, or foreign matter.

Workmanship. The sterilization envelope shall be free from defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing ten packaged envelopes, as specified, constitutes one unit. The package shall be a sealed peel-open type container capable of maintaining sterility of contents unless the package is damaged or opened.

Quality Assurance Provisions.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, with an AQL of .65 (percent defective) and an inspection level of S-4. The unit of product for examination purposes shall be one 12 inch length of the envelope roll as specified in this description.

Defects.

Envelope not fabricated from laminated film and Mylar with surgical Kraft paper suitable for heat sealing.

Envelope does not have sterilization indicator markings.

Envelope not free of tears, cuts, holes, or perforations.

Film front not clear enough for easy product identification.

Envelope not free of grease, dirt, or foreign material.

Sampling for tests. Sampling for all tests including dimensional test, shall be conducted in accordance with MIL-STD-105, with an AQL of 2.5 (percent defective) and inspection level of S-1 for the dimensional test and seal strength test, and with an AQL of zero (percent defective) and an inspection level of S-1 for the indicator tests. The unit of product for test purposes shall be one twelve inch length of the envelope roll.

Tests.

Steam indicator test. The steam indicator of the envelope shall change to the degree required by the envelope instructions when exposed to either a 3-1/2 minute at 272°F pre-vacuum cycle or a 250°F/20 minute gravity cycle. No test failure shall be permitted.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Ethylene Oxide Indicator Test. The ETO indicator of the envelope shall change to the degree required by the envelope instructions when exposed to a standard ETO cycle. No test failure shall be permitted.

Seal strength test. A test sample 1.0 inch wide by 2-1/2 inches long shall be cut from the envelope. The manufacturer's seal shall lie along one of the 1 inch edges. One of the flaps shall be lightly folded back against the seal to form a test specimen 1.0 inch wide by approximately 4-1/2 inches long. A tensile force of 350 grams shall be applied to the 1 inch ends of the specimen. The seal shall not fully part. If the seal fully parts, it shall constitute a defect.

Contractor certification. The contractor shall certify that the product offered meets the salient characteristics of this description and conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The Government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Regulatory requirements.

Federal Food, Drug and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Public Law 94-580, as amended, and as implemented by Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

NOTE: The following National Stock Numbers are covered by this document:

6530-01-125-3220 (Size 1)
6530-01-123-8011 (Size 2)
6530-01-125-3219 (Size 3)

Ordering data. Intermediate/exterior package quantities, labeling, and marking must be specified in the contract and/or order.

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

PREPARING ACTIVITY:

DoD - MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6530-1953

PGCs 06206
58335

Location: CID/ENVELOPE/H13-25

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS. In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE. This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
ATTN: DPSC-ATTS
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426

PREVIOUS EDITION IS OBSOLETE.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:18Z
Check the source to verify that this is the current version before use.