MILITARY SPECIFICATION

DOSIMETER, THERMOLUMINESCENT, DT-526/PD

This specification is approved for use by the Naval Electronic Systems Command, Department of the Navy, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers the design, production, and tests for a thermoluminescent dosimeter DT-526/PD, hereinafter referred to as the dosimeter. The radiation sensitive element of th dosimeter is a luminescent solid that stores part of the energy received from exposure to gamma radiation. The storage is due to trapping of electrons, which have been freed by the radiation, as imperfections in the solid. When the phosphor is heated, the stored energy is emitted as luminescent light. The light output is proportional to the accumulated radiation dose.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. Unless otherwise specified, the following specifications and standards of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DoDISS) specified in the solicitation form a part of this specification to the extent specified herein.

SPECIFICATIONS

MILITARY

MIL-E-16400	Electronic, Interior Communication and Navigation Equipment, Naval Ship And Shore: General Specification For
MIL-E-17555	Electronic And Electrical Equipment, Accessories, And Repair Parts: Packaging And Packing Of

STANDARDS

MILITARY

MIL-STD-105	Sampling Procedures And Tables For Inspection By Attributes
MIL-STD-108	Definitions Of And Basic Requirements For Enclosures For Electric And Electronic Equipment
MIL-STD-109	Quality Assurance Terms And Definitions
MIL-STD-167-1	Mechanical Vibrations Of Shipboard Equipment (Type I - Environmental and Type II - Internally Excited)
MIL-STD-810	Environmental Test Methods And Engineering Guidelines

2.1.2 Other Government publications. The following other Government publications form a part of this specification to the extent specified herein.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Commander, Naval Electronic Systems Command (ELEX-8111), Washington, DC 20363, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document, or by letter.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

NATIONAL BUREAU OF STANDARDS (NBS)

NBS Special Publication 250	Calibration And Test Services

NAVAL ELECTRONIC SYSTEMS COMMAND (NAVELEX)

NAVELEX 0967-LP-456-6010	Thermoluminescent Dosimeter (TLD) Reader (Radiac Computer-Indicator (CP-1112/PD)

(Copies of specifications, standards, and publications required by manufacturers in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)

2.1.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence.

3. REQUIREMENTS

3.1 General requirements. The dosimeter shall conform to the requirements of MIL-E-16400, except as hereinafter specified. The dosimeter shall be two-way interchangeable, at all levels down to component parts, mechanically and electrically with Navy type dosimeters. The basic equipment description and requirements for operation of the dosimeter shall be as specified in NAVELEX 0967-LP-456-6010.

3.1.1 First article. When specified, a sample shall be subjected to first article inspection (see 4.3 and 6.3).

3.2 Environmental. Environmental requirements shall be as specified in 3.2.1 through 3.2.3.

3.2.1 Temperature range and thermal design. Temperature range of the dosimeter shall be Range 4 in accordance with MIL-E-16400. Thermal design shall be in accordance with MIL-E-16400.

3.2.2 Humidity. The dosimeter shall be designed to conform to the humidity requirements specified in MIL-E-16400. Tropicalization is not required.

3.2.3 Vibration. The dosimeter shall be capable of withstanding the Type I vibration test of MIL-STD-167-1.

3.3 Characteristics. Characteristics of the dosimeter shall be as specified in 3.3.1 through 3.3.9.

3.3.1 Dosimeter composition. The dosimeter shall consist of a shielded case, dosimeter element, cap with accident dosimeter, and alligator clip. Parts shall be two-way interchangeable with existing dosimeters in Navy use.

3.3.1.1 Weight. Weight of the dosimeter shall not exceed 56 grams (g) (2 ounces (oz)).

3.3.1.2 Size. The case with cap in place shall not exceed 22.9 cubic centimeters (1.4 cubic inches). The inside diameter of the case shall be not more than 0.050 centimeters (cm) (0.020 inch (in.)) greater than the outside diameter of the dosimeter element.

3.3.2 Case, cap, and alligator clip. The case, cap, and alligator clip shall be in accordance with the requirements of 3.3.2.1 through 3.3.2.5.

3.3.2.1 Accessibility. The dosimeter shall be tamper proof insofar as consistent with design, that is, the dosimeter shall not be easily opened by the wearer. A special tool (see 6.4), a pentagonal wrench installed on CP-1112/PD Thermoluminescent Dosimeter (TLD) Readers as specified in NAVELEX 0967-LP-456-6010, shall be used for opening the case. A cap catch equivalent to the catch used on the dosimeter shall be used to ensure that the dosimeter is tamper proof.

3.3.2.2 Finish. The exterior finish of the case with cap in place shall be smooth and void of flares and crevices insofar as consistent with design. The case shall be black and the cap shall be gray.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

3.3.2.3 Attachment. The case shall have a belt loop (loop material to be same as case material) to attach the dosimeter to a 3.81 cm (1.500 in.) belt. Removal of the dosimeter without unbuckling the belt shall be impossible. The dosimeter shall be provided with a double-jawed alligator type metal clip. When clipped to a cotton shirt pocket, a clip shall withstand a force of 2.7 kilograms (kg) (6 pounds (lbs)) without sufficient slippage to disengage the clip from the cloth or from the case.

3.3.2.4 Case breakage. The case shall withstand the case breakage test specified in 4.5.5.

3.3.2.5 Shield. The shield shall be configured so that the dosimeter will conform to the requirements specified in 3.3.5.3.2.

3.3.3 Accident dosimeter. The accident dosimeter shall be incorporated in the same manner as in dosimeters in Navy use. Radiation from the indium after exposure shall be capable of being measured with a Geiger-Mueller type detector, without removal of the accident dosimeter from the cap or detector case. The accident dosimeter shall consist of two indium discs 0.9525 cm (0.375 in.) in diameter by .0254 cm (0.01 in.) ±0.00254 cm (0.001 in.) thick and separated by two sulfur discs weighing 0.3720937 g ±0.014 g (0.013125 oz ±0.000525 oz) each. A bulk quantity of flowers of sulfur shall be pressed at 45369 kg (10,000 lbs) per disc surface, or greater pressure, to form each sulfur disc.

3.3.4 Dosimeter element. The radiation dosimeter element shall consist of manganese activated calcium fluoride and shall be glass enclosed. The dosimeter element shall be designed to be compatible with and capable of being read on a CP-1112/PD computer-indicator with the accuracy and erasing requirements specified in 3.3.5.3 and 3.3.5.4. The detector glow curve indicated on the CP-1112/PD computer-indicator shall exhibit a prominent single peak which occurs no later than 11 seconds from the initiation of a read cycle. Smaller secondary peaks occurring before the prominent single peak shall be not greater than one-tenth the peak level of the prominent peak.

3.3.4.1 Dosimeter fading and self-irradiation. After irradiation, the fading or growth from 48 hours to 1 year shall not exceed ±5 percent. Dosimeters shall be free of trace contaminants so that a dose of not more than 0.05 milliroentgens (mR) per day shall register.

3.3.4.2 Dosimeter lead. The dosimeter leads protruding from the glass enclosed bulb shall be gold plated. The plating shall be smooth and continuous. Lead diameter shall be not less than 0.09398 cm (0.037 in.).

3.3.5 Dosimeter operation. The dosimeter shall have the operating characteristics specified in 3.3.5.1 through 3.3.5.4.

3.3.5.1 Range. The dosimeter shall cover the range from 1 mR to 5000 roentgens (R).

3.3.5.2 Minimum dose. The minimum detectable dose shall be at least 1 mR.

3.3.5.3 Accuracy. Accuracy of the dosage indication over the range of 5 mR to 1500 R shall be ±10 percent or 5 mR, whichever is greater, and ±20 percent over the range of 1500 R to 5000 R.

3.3.5.3.1 Accuracy at cobalt 60 and cesium 137 energies. The dosimeter shall measure true gamma dose from an unshielded cobalt 60 source and an unshielded cesium source within ±10 percent or 5 mR, whichever is greater, over the range of 5 mR to 1500 R and ±20 percent over the range of 1500 R to 5000 R. The dosimeter shall be read in a time period of between 1 hour and 2 hours after exposure. This accuracy shall be determined by sources as specified in a and b:

a. Calibrated by NBS
b. Calibrated by the contractor or by a Government activity using Victoreen R meters, or equivalent, which have, in turn, been calibrated by NBS

3.3.5.3.2 Energy response. The dosimeter shall indicate dose received within ±20 percent of the true R dose, or 10 mR, whichever is greater, when exposed to X or gamma radiation over the entire energy range of 80 thousand electron volts (KEV) to 1.3 million electron volts (MEV) and read in the CP-1112/PD TLD Reader between 1 hour and 2 hours after exposure. At energies below 80 KEV, response shall uniformly decrease and be less than 50 percent at 50 KEV. A constant potential X-ray machine with the appropriate tube voltage and filtration, in accordance with normal NBS procedures, shall be used for calibration tests in the low enery region. The X-ray machine filtration specified in TABLE I is for heavily filtered X-rays (see NBS Technique H-F, NBS 250).

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

TABLE I. Heavily filtered X-rays.

Step	Kilovolts constant potential	Equivalent inherent filter, aluminum mm 1/ (in.)	Filter added mm 1/ (in.)				Approximate effective energy (KEV)
			Lead	Tin	Copper	Aluminum	
A	50	4 (0.16)	0.12 (0.0048)	0	0	0	40
B	100	4 (0.16)	0.53 (0.0212)	0	0	0	70
C	150	4 (0.16)	0	1.5 (0.060)	4 (0.16)	0	120
D	200	4 (0.16)	0.7 (0.028)	4 (0.16)	0.6 (0.024)	0	170
E	250	4 (0.16)	2.7 (0.108)	1 (0.040)	0.6 (0.024)	0	215

3.3.5.4 Erasing capability. After readout, a dosimeter element, which has received a dose up to 2 R, shall have a residual of less than 1.0 mR and be ready for reuse without further heating or processing. At exposures greater than 2 R, additional readout cycles with a CP-1112/PD TLD Reader may be required for reuse.

3.3.6 Drop test. The dosimeter shall withstand the drop test specified in 4.5.4.

3.3.7 Dosimeter reusability. The dosimeter shall be designed to be reusable for a minimum of 300 times. A reuse is a single readout cycle with a CP-1112/PD TLD Reader set to give 6 amperes (A) for 20 seconds.

3.3.8 Degree of enclosure. The dosimeter shall be watertight in accordance with MIL-STD-108.

3.3.9 Identification of case. The nomenclature and serial number shall be engraved on the case.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Government verification. All quality assurance operations performed by the contractor will be subject to Government verification at any time. Verification will consist of, but is not limited to, a) surveillance of the operations to determine that practices, methods, and procedures of the written quality program are being properly applied, b) Government product inspection to measure quality of the product to be offered for acceptance, c) Government inspection of delivered products to assure compliance with all inspection requirements of this specification. Failure of the contractor to promptly correct deficiencies discovered by him or of which he is notified shall be cause for suspension of acceptance until corrective action has been taken or until conformance of the product to prescribed criteria has been demonstrated.

4.1.2 Quality assurance terms and definitions. Quality assurance terms used in this specification shall be as defined in MIL-STD-109

4.2 Classification of inspections. The inspection requirements specified herein are classified as specified in a through c:

a. First article inspection (see 4.3)
b. Quality conformance inspection (see 4.4)
   1. Production inspection (see 4.4.1)
   2. Production control inspection (see 4.4.2)
c. Inspection of preparation for delivery (see 4.6)

Source: https://assist.dla.mil - Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

4.3 First article inspection. Unless otherwise specified (see 6.2), 1500 units shall be required for first article inspection. First article inspection shall consist of all examination and testing necessary to determine compliance with the requirements of this specification. First article inspection shall include the tests specified in TABLE II.

TABLE II. First article inspection.

Examination or test	Requirement paragraph	Test paragraph
Part compatibility	3.1	4.5
Finish	3.3.2.2	4.5
Marking	3.1, 3.3.9	4.5
Dosimeter lead	3.3.4.2	4.5
Performance tests		4.5
Range and sensitivity	3.3.5.1, 3.3.5.2	4.5
Accuracy	3.3.5.3	4.5
Dosimeter fading and self-irradiation	3.3.4.1	4.5
Accuracy at cobalt 60 and cesium 137 energies	3.3.5.3.1	4.5
Energy response	3.3 5.3.2	4.5
Surface examination		
Weight	3.3.1.1	4.5
Size	3.3.1.2	4.5
Erasing capability	3.3.5.4	4.5.7
Humidity	3.2.2	4.5.2
Temperature and thermal design	3.2.1	4.5, 4.5.1
Enclosure	3.3.8	4.5
Vibration	3.2.3	4.5.3
Drop and case breakage	3.3.2.4, 3.3.6	4.5.4, 4.5.5
Clip test	3.3.2.3	4.5.6
Reusability	3.3.7	4.5.8

4.4 Quality conformance inspection. Quality conformance inspection shall be as specified in 4.4.1 through 4.4.2.1.

4.4.1 Production inspection. Production inspection shall be made on every dosimeter offered for delivery. Production inspection shall comprise such examination and testing which will prove the workmanship and reveal the omissions and errors of the production process, such as functional and performance tests at a limited number of points, tests which detect deviations from design, tests of adjustment, and tests which detect hidden defects of material. Production inspection shall include the tests specified in TABLE III. The accuracy test shall be made with the dosimeter exposed to 80 mR and 800 mR gamma dose of unshielded cobalt 60 or unshielded cesium 137. During the accuracy test, any dosimeter which exhibits a glow curve with a peak occurring later than 11 seconds from the initiation of its read cycle, shall be rejected. The average exposure reading of each lot of dosimeters submitted shall be within ±5 percent of the corrected exposure reading, based on Government-furnished standard dosimeters (see 6.5) after the lot has been subjected to the same exposure to the same source.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

TABLE III. Production inspection.

Examination or test	Requirement paragraph	Test paragraph
Surface examination		
Weight	3.3.1.1	4.5
Size	3.3.1.2	4.5
Part compatibility	3.1	4.5
Finish	3.3.2.2	4.5
Marking	3.1, 3.3.9	4.5
Dosimeter lead	3.3.4.2	4.5
Performance test		
Accuracy at cobalt 60 and cesium 137 energies	3.3.5.3.1	4.4.1, 4.5

4.4.2 Production control inspection. Production control inspection, including sampling, shall conform to TABLE IV and to the procedures for special inspection levels of MIL-STD-105. The inspection level shall be S-3 with an acceptable quality level (AQL) of 6.5 for single sampling plan. The inspection shall be normal, tightened, or reduced, as indicated by the inspection condition (except for the dosimeter reuse tests). This test shall be tightened inspection only. Each test dosimeter shall be exposed to a minimum of 800 mR radiation and cycled for at least 300 times without degradation below the requirements specified in 3.3.5.3. Production control inspection shall be performed on dosimeters that have passed production inspection. The dosimeter shall satisfactorily conform to the requirements of production control inspection prior to release for shipment. The dosimeters used for this test shall be marked for test and segregated from the other dosimeters.

TABLE IV. Production control inspection.

Examination or test	Requirement paragraph	Test paragraph
Enclosure test	3.3.8	4.5
Clip test	3.3.2.3	4.5.6
Performance tests		
Range	3.3.5.1	4.5
Energy response	3.3.5.3.2	4.5
Erasing capability	3.3.5.4	4.5.7
Temperature and thermal design	3.2.1	4.5, 4.5.1
Humidity	3.2.2	4.5.2
Drop and case breakage	3.3.2.4, 3.3.6	4.5.4, 4.5.5
Vibration	3.2.3	4.5.3
Reusability	3.3.7	4.4.2, 4.5.8

4.4.2.1 Rejected lots. If an inspection lot is rejected, the contractor may rework the lot to correct the defects, or screen out the defective units, and resubmit the lot for reinspection. Resubmitted lots shall be inspected using tightened inspection. Such lots shall be separate from new lots and shall be clearly identified as reinspected lots.

4.5 Test methods. Examinations and tests shall be in accordance with the applicable test methods of MIL-E-16400, unless otherwise specified in 4.5.1 through 4.5.10.

4.5.1 Temperature tests. The dosimeter shall be capable of satisfactory operation over the temperature range specified in 3.2.1, and shall conform to the accuracy requirements of 3.3.5.3.1 over the entire cycle.

4.5.2 Humidity test. The humidity test shall be conducted in accordance with Method 507.2, Procedure III of MIL-STD-810 to determine conformance to the requirements specified in 3.2.2.

4.5.3 Vibration test. The dosimeter shall be vibration tested to determine conformance to the requirements specified in 3.2.3.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

4.5.4 Drop test. The dosimeter shall be capable of withstanding six drops from a height of 1.83 meters (6 feet) onto a hardwood floor, two drops on each end and two drops on the side.

4.5.5 Case breakage test. The case (less cap) shall be placed on an unyielding surface. An unyielding weight of 5.625 kg (12.5 lbs) shall be allowed to free fall 30 cm (12 in.) onto the case. Three drops shall be performed on each test sample. Visual inspection shall show that no pieces departed the test sample.

4.5.6 Clip test. The detector clip shall conform to the requirements of 3.3.2.3 after fully opening and closing the jaws 1000 times.

4.5.7 Residual exposure. Residual exposure of sample dosimeters shall be checked after the dosimeters have been exposed to 1.5 R, 5 R, 50 R, 500 R, and 5000 R. Residual exposure values shall be:

Exposure	Residual exposure
1.5 R	≤1 mR
5 R	≤5 mR
50 R	≤30 mR
500 R	≤300 mR
5000 R	≤3 mR

No dosimeter shall be exposed to an accumulative dose greater than 5000 R. After bake-out of the residual exposure, no dosimeter shall indicate greater than 15 mR.

4.5.8 Dosimeter reusability test. Dosimeters shall be randomly selected from each production lot in accordance with MIL-STD-105 inspection level S-3. Each of these units shall be exposed to at least 500 mR and read with a CP-1112/PD TLD Reader set to give 6 A for 20 seconds. This shall be done ten times for each test unit and the results determined. The test units shall then either be placed on an automatic cycling fixture which has been demonstrated to produce the same results as the CP-1112/PD TLD Reader for its heating cycle specified herein, or shall be manually subjected to the heat cycle of the CP-1112/PD. This shall occur once every 5-minute period for a total of 280 times. If the automatic cycling fixture is used, the test units shall be removed from the fixture. All test units shall then be exposed to at least 500 mR and read as specified herein on a CP-1112/PD TLD Reader. The average reading shall be determined for each group of 10 readings. The response of each dosimeter shall not fall outside ±10 percent of actual exposure based on the Government-furnished standard dosimeter (see 6.5). Accept or reject criteria shall be as specified in MIL-STD-105, inspection level S-3, AQL 6.5, tightened inspection only.

4.5.9 Failure. Failure shall be any malfunction or parameter deviation that prevents the equipment from performing within the operational requirements specified herein. All failures shall be considered relevant unless determined by the procuring activity to be caused by a condition external to the equipment under test which is not a test requirement and not encountered in service.

4.5.10 Final restoration and disposition of tested equipment. Upon completion of reuse testing, the tested equipment shall be returned to production for reinspection and reconditioning as required, and may be delivered as part of the contract quantity if it conforms to all terms and conditions of the contract for acceptance.

4.6 Inspection of preparation for delivery. Inspection shall be performed to ensure conformance with the requirements of Section 5.

5. PACKAGING

(The preparation for delivery requirements specified herein apply only for direct Government procurements. Preparation for delivery requirements of referenced documents listed in Section 2 do not apply unless specifically stated in the contract. Preparation for delivery requirements for products procured by contractors shall be specified in the individual order.

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

5.1 Preservation, packaging, packing, and marking. Unless otherwise specified herein, preparation for delivery shall be in accordance with the applicable levels of preservation, packaging, packing, and marking specified in MIL-E-17555 (see 6.2).

6 NOTES

6.1 Intended use. The dosimeter is intended for shipboard use and at shore facilities.

6.2 Ordering data. Procurement documents should specify:

a. Title, number, and date of this specification
b. Number of first article samples to be submitted if other than specified in 4.3
c. Levels of preservation, packaging, packing, and marking (see 5.1)

6.3 First article. When a first article is required it shall be tested and approved under the appropriate provisions of 7-104.55 of the Defense Acquisition Regulation. The first article should be a first production item. The first article should consist of 1500 units. The contracting officer should include specific instructions in all procurement instruments, regarding arrangements for examinations, tests, and approval of the first article.

6.4 Special tools. Special tools are defined as those tools not listed in the National Supply Catalog (copies of this catalog may be consulted in the office of the Defense Contract Administration Service).

6.5 Government-furnished equipment (GFE). The contracting officer should arrange to furnish the GFE listed in 4.4.1 and 4.5.8.

Preparing activity:
NAVY-EC

(Project 6665-N453(EC))

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document*(s)* or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEPARTMENT OF THE NAVY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12503 WASHINGTON D C

POSTAGE WILL BE PAID BY THE DEPARTMENT OF THE NAVY

Commander
Naval Electronic Systems Command
(ELEX 8111)
Washington, DC 20363

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1. DOCUMENT NUMBER	2 DOCUMENT TITLE

3a NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
b ADDRESS *(Street, City, State, ZIP Code)*	☐ USER
	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* — Optional	b WORK TELEPHONE NUMBER *(Include Area Code)* — Optional
c MAILING ADDRESS *(Street, City, State, ZIP Code)* — Optional	8 DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 IOUS EDIT ... 'S OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-20T20:40Z
Check the source to verify that this is the current version before use.