INCH-POUND

A-A-55000
15 February 1995

COMMERCIAL ITEM DESCRIPTION

CATHETERIZATION SET, CARDIOVASCULAR, Radial Artery

The General Services Administration has authorized the use of this Commercial Item Description.

This Commercial Item Description covers a sterile, disposable radiopaque teflon catheter set suitable for radial artery catheterization.

Salient characteristics:

Material. The needle cannula and guide wire shall be fabricated from Class 3 austenitic corrosion-resisting steel of the composition listed in accordance with the American Society for Testing and Materials (ASTM) F899.

General. The catheter set shall be a sterile, disposable, 20 gauge by 1-3/4 inches (4.45 cm), radiopaque, polytetrafluoroethylene catheter over a 22 gauge, thin wall needle with integral wire guide suitable for radial artery catheterization. The distal tip of the spring wire guide shall be soft, serving as an obturator for blood flow restriction. Shall have an extra short needle bevel with precision fit between the bevel heel and catheter tip and a polytetrafluoroethylene reinforcing sleeve at the catheter-hub junction to minimize kinking. The proximal portion of the catheter shall have an indented circumferential ring for suture attachment of the catheter to the skin. Shall have an actuating lever for advancing the spring wire guide through the needle lumen. The feed tube and introducer needle hub shall be of clear plastic for easy visualization of blood flashback. Tne catheter hub shall be in accordance with the International Standards Organization (ISO) 594.

AMSC N/A FSC 6515

DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Needle. Except as specified herein the needles shall be in accordance with the International Standards Organization (ISO) 7864. The heel of the needle bevel shall extend beyond the catheter tip when the needle is fully inserted into the catheter. The needle cannula shall make a free-sliding fit within the catheter. Needle point shall be well defined.

The needle hub shall be fabricated from rigid polypropylene that meets the requirements of the United States Pharmacopeia (USP) for Class IV plastics. The sheath shall be fabricated from rigid plastic. The exterior of the cannula shall be lubricated with an inert non-toxic medical grade silicone compound.

The tip of the needle point shall be ground to maximum keenness. The heel of the needle point shall be dull to prevent coring. The needle shall have a Luer-lock hub in accordance with ISO 594. The catheter shall be fabricated from polytetrafluorethylene which meets the requirements of the USP for Class IV plastics

Radiopacity test. The catheter shall be radiopaque in accordance with ASTM F640.

Sterility. The catheter set shall be sterile when tested in accordance with the USP. Sterility assurance level shall be 0.000001 for each set.

Safety. The catheter and all components coming in contact with intravenous medication and patient's blood shall contain no more than 5 parts per million of lead. The finished catheter and hub shall be pyrogen-free and non-toxic in accordance with the USP.

Workmanship. The catheterization set shall be free from other defects which detract from its appearance or impair its serviceability.

Unit. Package (PG). One package containing fifty individually packaged catheterization sets, as specified, constitutes one unit. The catheterization set shall be individually packaged in a suitable peel-open container which will allow aseptic handling and maintain sterility of the contents until opened or damaged.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

QUALITY ASSURANCE PROVISIONS.

Responsibility for inspection. Unless otherwise specified in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government upon request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Inspection. Inspection, as used herein, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Examination. Examination shall be conducted to determine compliance with specification requirements.

Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and as specified. Unit of product for examination purposes shall be one catheterization set as specified. Sampling shall be inspection level II with an Acceptable Quality Level (AQL) of 1.0.

Tests. Tests shall be conducted to determine compliance with specification requirements. Where feasible, the same sample shall be used for the determination of two or more test characteristics.

Leakage and security test. The catheters shall not pull out or loosen from the catheter hub when subjected to an axial force of 3.0 pounds. Catheters shall not exhibit leakage exceeding one clinging drop (hydrostatic test) with 20 psig applied for two minutes using appropriate fixtures and test gauges. The needle cannula shall not pull out of, or loosen from, the needle hub when subjected to an axial force of ten pounds.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Sampling for tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and as specified. The unit of product for test purposes shall be one catheterization set. Sampling shall be inspection level S-2 with an AQL of 1.0.

Contractor certification. The contractor shall certify and maintain substantiating evidence that the product offered meets the salient characteristics of this Commercial Item Description, and that the product conforms to the producers' own drawings, specifications, standards, and quality assurance practices. The government reserves the right to require proof of such conformance prior to first delivery and thereafter as may be otherwise provided for under the provisions of the contract.

Metric products. Products manufactured to metric dimensions will be considered on an equal basis with those manufactured using inch-pound units, providing they fall within the tolerances specified using conversion tables contained in the latest revision of Federal Standard No 376, and all other requirements of this document are met.

If a product is manufactured to metric dimensions and those dimensions exceed the tolerances specified in the inch/pound units, a request should be made to the contracting officer to determine if the product is acceptable.

The contracting officer has the option of accepting or rejecting the product.

Regulatory requirements.

Federal Food, Drug, and Cosmetic Act. If the product covered by this document has been determined by the U.S. Food and Drug Administration to be under its jurisdiction, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of the Federal Food, Drug, and Cosmetic Act, as amended, and regulations promulgated thereunder. In addition, the offeror/contractor shall comply, and be responsible for compliance by its subcontractors/suppliers, with the requirements of all other applicable Federal, State, and local statutes, ordinances, and regulations.

Recovered materials. The offeror/contractor is encouraged to use recovered material in accordance with Federal Acquisition Regulation Subpart 23.4 to the maximum extent practical.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

Preservation, packaging, packing, labeling, and marking. Unless otherwise specified, preservation, packaging, and packing shall be to a degree of protection to preclude damage to containers and/or contents thereof under normal shipping conditions, handling, etc., involving shipment from the supply source to the receiving activity, plus reshipment from receiving activity, and shall conform to applicable carrier's rules and regulations. Intermediate and exterior package quantities and labeling and marking shall be as specified in the contract and/or order.

Ordering data Intermediate/exterior package quantities and labeling and marking must be specified in the contract and/or order.

NOTE: The following National Stock Number and its respective Item Identification is covered by this document:

NATIONAL STOCK NUMBER	ITEM IDENTIFICATION
6515-01-166-0154	CATHETERIZATION SET, Radial Artery, Sterile, Disposable, Silicone Rubber, 20 Gauge by 1-3/4 inch, 50s.

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

PREPARING ACTIVITY:

DoD-MB

Agent:

DLA-DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
USPHS
FDA-MPQAS

Project No. 6515-5745
Location:\word\mark\cath(mab)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IM

INSTRUCTIONS

1. The preparing activity must complete blocks 1, 2, 3 and 8. In block 1, both the document number and revision letter should be given.
2. The submitter of this form must complete blocks 4, 5, 6 and 7.
3. The preparing activity must provide a reply within 30 days from receipt of the form.

NOTE: This form may not be used to request copies of documents, nor to request waivers, or clarification of requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

I RECOMMEND A CHANGE:	1. DOCUMENT NUMBER	2. DOCUMENT DATE (YYMMDD)
	A-A-55000	95/02/15

3. DOCUMENT TITLE
CATHETERIZATION SET, CARDIOVASCULAR, Radial Artery

4. NATURE OF CHANGE (Identify paragraph number and include proposed rewrite, if possible. Attach extra sheets as needed.)

5. REASON FOR RECOMMENDATION

6. SUBMITTER

a. NAME (Last, First, Middle Initial)	b. ORGANIZATION	
c. ADDRESS (Include Zip Code)	d. TELEPHONE (Include Area Code) (1) Commercial (2) AUTOVON (If applicable)	7. DATE SUBMITTED (YYMMDD)

8. PREPARING ACTIVITY

a. NAME	b. TELEPHONE (Include Area Code) (1) Commercial	(2) AUTOVON
Defense Personnel Support Center ATTN DPSC-MAP	(215) 737-8431	444-8431
c. ADDRESS (Include Zip Code) 2800 South 20th Street Philadelphia, PA 19145	IF YOU DO NOT RECEIVE A REPLY WITHIN 45 DAYS CONTACT Defense Quality and Standardization Office 5203 Leesburg Pike Suite 1403 Falls Church VA 22041-3466 Telephone (703) 756-2340 AUTOVON 289-2340	

DD Form 1426, OCT 89 — Previous editions are obsolete — 1980

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T09:41Z
Check the source to verify that this is the current version before use.