
FED. SUP CLASS
6640



MATERIAL: The press shall be made of steel and other construction materials as required for reliable service.

DESIGN: Unit is a hydraulic mounting press that will provide fingertip pumping to full pressure needed and quick ram return. A pressure gauge for direct reading shall be furnished. The mold closure system shall close fast and open quickly without striking to break the seal. Also, a thermostatically controlled heater and a finned slip-on cooler that accommodates molds for 2.5, 3.2 and 3.8 cm sample sizes. Overall dimensions are 27.9 W x 35.6 D x 58 H cm. Power requirements are 115V AC 50/60Hz.

All electrical components shall conform to the standards of Underwriters Laboratories, Incorporated.

NOTE: Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

FSCM: 08455 and 09410.

USER ACTIVITIES: ARMY - MD, CM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 82

APPROVED 20 JUL 1983 REVISED

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

P.A. EA	INTERNATIONAL INTEREST	TITLE	MILITARY STANDARD
Other Cust MS 99		PRESS, SPECIMEN PREPARATION, LABORATORY	MS36246
Procurement Specifications NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-1189

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T17:37Z
Check the source to verify that this is the current version before use.