
FED. SUP CLASS
6640



MATERIAL: The rack shall be made of polypropylene plastic or other materials of equal strength and durability.

DESIGN: The rack shall hold eight sedimentation tubes in a vertical position. Black sedimentation rate scale shall be permanently imprinted beside each tube aperture and red hematocrit scales at each end. Leveling screws and spirit level shall be included for proper positioning of the rack.

NOTE: 1. Items differing from this Military Standard by minor variations in design or dimensions are considered standard under this document, provided such variations do not affect performance and are not in conflict with the material requirements specified herein.

USER ACTIVITIES: ARMY - MD, DM
NAVY - MS

REVIEWER ACTIVITIES: ARMY - ME
NAVY - OS
AIR FORCE - 03

This military standard is approved for use by all Departments and Agencies of the Department of Defense. Selection for all new engineering and design applications and for repetitive use shall be made from this document.

APPROVED 26 JUL 1982 REVISED

P.A. EA / Other Cust MS 99	INTERNATIONAL INTEREST	TITLE: RACK, SEDIMENTATION, BLOOD	MILITARY STANDARD MS 36413
Procurement Specifications: NONE	SUPERSEDES:		SHEET 1 OF 1

DD FORM 1 MAY 73 672 (COORDINATED) PREVIOUS EDITIONS OF THIS FORM ARE OBSOLETE

PROJECT NUMBER 6640-0797
STANDARD DATE

★ U.S. GOVERNMENT PRINTING OFFICE: 1982 [illegible]

Source: https://assist.dla.mil -- Downloaded: 20260-23T17:40Z
Check the source to verify that this is the current version before use.