INCH-POUND

A-A-59164
21 March 2003

COMMERCIAL ITEM DESCRIPTION
TESTER, FLASH POINT, ASTM D56 METHOD

The General Services Administration has authorized the use of this commercial item description (CID) for all federal agencies.

1. SCOPE

1.1 Description. This Commercial Item Description (CID), describes Tester, Flash Point, ASTM D56 Method. The flash point tester is used to measure and describe the properties of materials, products, or assemblies in response to heat and flame under controlled laboratory conditions.

2.0 SALIENT CHARACTERISTICS

2.1 Equipment. The complete countertop instrument shall be capable of manually and automatically determining the flash point of volatile materials in accordance with ASTM D 56. The instrument is designed to increase the laboratory's productivity, repeatability and reproducibility. Test results include adjustments to compensate for barometric pressure differences.

2.2 Standard Test Method. The flash point tester shall automatically perform a variety of standard tests including testing in accordance with ASTM D56.

2.3 Sample Temperature Measurement Range. The sample temperature range shall be -20 degree C to +71 degree C with increments of 0.1 degree C.

2.4 Sample Heating. The tester shall be capable of automatically determining the heating rate for the sample under test or can be user defined.

Beneficial comments, recommendations, additions, deletions, clarifications, etc. and any other data which may improve this document should be sent to: WR-ALC/LEEG, 295 Byron Street, Robins AFB, GA 31098-1611.

AMSC N/A FSC 6630

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.

2.5 Sample Stirring. The tester shall be capable of automatically determining the stirring rate for the sample under test or can be user defined.

2.6 Ignition Frequency. The tester shall be capable of automatically determining the ignition frequency for the sample under test or can be user defined.

2.7 Flash Detection System. The flash point tester shall be equipped with a thermal detector which eliminates interference from water and/or silicone .

2.8 Flash Ignition System. The tester shall be equipped with a gas or electronic ignition system.

2.9 Display. The tester shall be supplied with a LCD ¼ VGA monochrome display or equivalent. The display shall be visible from a distance of 5 meters.

2.10 Keyboard. The tester shall be designed with a keyboard or similar device which will allow the users to input data to the tester.

2.11 Data Storage. The tester shall be capable of storing multiple test results and have the capability of transferring the test data to another data storage device or being able to obtain a printed copy of the test results.

2.12 Power Requirements. If the solicitation indicates destination(s) outside the United States, then the offered item(s) may have to be altered to adapt to the proper power requirements and electric plugs for the destinations indicated. The proposed suppliers shall have the resources available to provide the proper power requirements for the destination indicated.

3.0. REGULATORY REQUIREMENTS.

3.1 Recovered materials. The offeror/contractor is encouraged to use recovered materials to the maximum extent practicable, in accordance with paragraph 23.403 of the Federal Acquisition Regulation (FAR).

4.0 QUALITY ASSURANCE PROVISIONS

4.1 Product Conformance: The products provided shall conform to the producer's own drawings, specifications, standards, and quality assurance practices and be the same product offered for sale in the commercial market. The government reserves the right to require proof of such conformance. Minor changes to the commercial product, with respect to ruggedizing and environmental suitability and protection shall be permitted, subject to the approval of the procurement office, to satisfy the requirements of this requirement.

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.

4.2 Installation and Demonstration: If required by the end user, the contractor shall be responsible for user orientation of the equipment at the designated Air Force facility demonstration that the equipment is ready for use. Installation and demonstration is projected to take one day. The cost for user orientation shall be at no additional cost to the government.

4.3 Market Acceptability: The following market acceptability criteria are necessary to document the quality of the product to be provided under this requirement.

4.3.1 The manufacturer of the item shall have been producing a product meeting these requirements for at least 5 years.

4.3.2 The manufacturer of the item shall have sold at least 250 units meeting these requirements to the government and or the commercial market over the last 2 years.

4.4 Alternative offers: A current, complete instrument with all items required for use, shall be offered/provided. If an alternative/replacement item is offered, descriptive literature (commercial catalogs, specification sheets, etc.) shall be submitted with the quote, for review and evaluation by the cognizant engineer who will coordinate qualifying the instrument with the end user. If the instrument has not been purchased previously by the government and meets the evaluation process, the instrument shall be subjected to functional testing prior to final approval by the government.

4.5 Serviceability, Reliability, and Quality of Materials:

4.5.1 One copy of the commercial operation and maintenance manual(s) shall be shipped with the equipment. One additional copy of the operation and maintenance manual shall be shipped to the cognizant engineer at no additional cost to the government. The address for the cognizant engineer is found in 4.5.2.

4.5.2 All supplied items shall carry a standard commercial warranty on parts and labor. Information on warranty duration and expiration date, conditions and point of contact shall be shipped with each unit. One additional copy of the warranty information shall be shipped to the cognizant engineer at no additional cost to the government. The address for the cognizant engineer is; FSC 6630 Engineer, WR-ALC/LEEG, 295 Byron Street, Robins AFB, GA 31098-1611.

5.0 PACKAGING

5.1 The package/packaging for shipment to the Government shall be in accordance with contract or purchase order specifications. If specifications are not specified, then the package shall be standard commercial packaging. Additionally, each shipping container shall be marked with the following information or as specified by the contact or purchase order:

Name of equipment
Manufacturer

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.

Model/part number
Contract and/or purchase order number
NSN
Ship to address
Return address

6.0 NOTES

6.1 Source of Documents

6.1.1 ASTM Document A copy of ASTM D 56, Standard Test Method for Flash Point by TAG Closed Tester can be obtained from ASTM 100 Barr Harbor Drive, P.O. Box C700, West Conshohocken, PA 19428-5959.

6.2 Suggested Sources of Supply.

Petroleum Analyzer Company L.P. 3234 East Pasadena Freeway Pasadena, TX 77503	Part Number HFP 362
ISL Instrumlentation Scientifique De Laboratoire Bretteville Sur Odon France	Part Number FP 56 5G

MILITARY INTEREST:

Custodian:
Air Force - 99

Preparing Activity:
Air Force - 84

Agent:
Air Force - 99

(Project No. 6630-0593)

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T10:05Z
Check the source to verify that this is the current version before use.