MILITARY SPECIFICATION

AIRWAY-ESOPHAGEAL OBTURATOR, PLASTIC, ADULT

This amendment forms a part of Military Specification MIL-A-37934(DLA-DM) dated 22 November 1978, and is approved for use by the Defense Personnel Support Center, Defense Logistics Agency and is available for use by all Departments and Agencies of the Department of Defense.

PAGE 2

3.1.1 Airway cuff. Delete last sentence.

PAGE 3

3.2 Style, design and dimensions. Delete third sentence.

3.3 Inflation of airway cuff. Line 2, delete "30 minutes" and substitute "24 hours".

3.4 Inflation of face mask rim. Line 2, delete "30 minutes" and substitute "24 hours".

3.6 Certificate of biological safety. Delete entirely and substitute: "The manufacturer of the tube-cuff shall certify that the tube-cuff meets the USP requirements for Class II plastics as specified in 3.1.1."

PAGE 6

4.5 Biological safety. Delete entirely and substitute: "Biological safety shall be in accordance with USP requirements for Class II and as specified in 3.6."

4.6 Airway and cuff inflation. Line 3, after the word "after" delete the balance of the sentence and substitute: "several minutes for leakage. Fifteen ml of air shall then be withdrawn and the cuff shall be observed 24 hours later for leakage."

4.7 Face mask rim inflation. Line 4, delete "30 minutes" and substitute: "24 hours".

PAGE 7

5.1.2 Level A. Line 4, delete "with" and substitute "within". Line 6, following "strip" insert "of". Also, delete "or inspection label".

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing Activity:

DLA-DM

* U.S. GOVERNMENT PRINTING OFFICE: 1983—605-034-2220

Source: https://assist.dla.mil -- Downloaded: 2025-[illegible]T18:39Z
Check the source to verify that this is the current version before use.

Project No. 6515-2579

FSC 6515