MILITARY SPECIFICATION

PROBES, PERIODONTAL

This specification is approved for use by the Defense Personnel Support Center, Defense Logistics Agency, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers probes suitable for diagnosing and measuring the depth of periodontal pockets.

1.2 Classification. Probes shall be of the following types and styles, as specified (see 6.1):

Type I. Single-ended:
  Style 1. - Glickman, No. 26G.
  Style 2. - No. 4.
Type II. Double-ended:
  Style 1. - Nabers, No. 2N.

2. APPLICABLE DOCUMENTS

2.1 Issues of documents. The following documents of the issue in effect on date of invitation for bid or request for proposal form a part of this specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

GG-I-526	- Instruments, Dental and Surgical; General Specifications.
PPP-B-566	- Boxes, Folding, Paperboard.
PPP-B-636	- Boxes, Shipping, Fiberboard.
PPP-B-676	- Boxes, Setup.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:39Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-Packaging, Methods Of.

STANDARDS

FEDERAL

FED-STD-66 - Steel, Chemical Composition and Hardenability.
FED-STD-151 - Metals, Test Methods.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.

DRAWING

DEFENSE PERSONNEL SUPPORT CENTER (DPSC)

24405 - Probes, Periodontal, No. 26G, No. 4, and No. 2N.

(Copies of specifications, standards, and drawings required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bid or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS (ASTM)

ASTM B 135 - Specification for Seamless Brass Tube.
ASTM E 18 - Tests for Rockwell Hardness and Rockwell Superficial Hardness of Metallic Materials.
ASTM E 92 - Test for Vickers Hardness of Metallic Materials.

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, PA 19103.)

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:39Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 General requirements. The probes shall be in accordance with the applicable requirements of GG-I-526 with the options and additional requirements specified herein.

3.2 Material.

3.2.1 Operative ends. The operative ends shall be class 3, type 302, 303, or 304 stainless steel.

3.2.2 Handles. The handles shall be class 3, type 303 or class 4, type 416 stainless steel, or class 6 stainless steel, except the type shall be 430F of FED-STD-66 in lieu of type 430, or class 8 alloy number 360 free cutting brass, chromium over nickel plated. As an alternate to class 8 alloy number 360 composition, the brass composition only shall be in accordance with alloy number 330 of ASTM B 135. The boil test shall apply for the steel and brass handles. The copper sulfate test shall apply for handles made of types 416 and 430F steel only.

3.3 Hardness. After proper stress relieving, the Rockwell hardness of the operative ends shall be C45 to C50.

3.4 Style, design, and dimensions. Type I probes shall be suitable for use in measuring the depth of periodontal pockets, and type II probe shall be suitable for diagnosing and evaluating periodontal pockets and abnormalities in bifurcation areas. The style, design, and dimensions of the probes shall conform to Defense Personnel Support Center (DPSC) Drawing No. 24405 for the applicable type and style.

3.4.1 Type I, style 1. Probe shall be a No. 26G probe of the Glickman design.

3.4.2 Type I, style 2. Probe shall be a No. 4 probe.

3.4.3 Type II, style 1. Probe shall be a No. 2N probe of the Nabers design.

3.4.4 Drawings. The drawings forming a part of this specification are specification control drawings. The supplier shall be responsible for preparing his own shop drawings. Where tolerances prescribed could cumulatively result in improper design, the supplier shall provide a product with dimensions within the tolerances prescribed on the drawing to insure proper design and function of the item. No deviation from the prescribed dimensions or tolerances is permissible without prior approval of the Contracting Officer.

Source: https://assist.dla.mil -- Downloaded: 20260-21T11:39Z
Check the source to verify that this is the current version before use.

3.5 Construction.

3.5.1 Type I, style 1. Probe shall be single-ended, two-piece construction, consisting of an operative end and a handle. The operative ends shall be permanently secured in the handles. The millimeter (mm) markings shall be legible, sharply defined lines, of uniform width and depth. Markings shall be calibrated in 1 mm, 2 mm, 3 mm, 5 mm, 7 mm, 8 mm, 9 mm, and 10 mm increments extending around the periphery of the measuring tip. The finger grip portion of the handles shall be finely serrated or fine diamond knurled.

3.5.2 Type I, style 2. Probe shall be single-ended, two-piece construction, consisting of an operative end and a handle. The operative ends shall be permanently secured in the handles. The operative ends shall be color-coded and calibrated in 3 mm, 6 mm, 9 mm, and 12 mm increments extending around the periphery of the measuring tip. The finger grip portion of the handles shall be finely serrated or fine diamond knurled.

3.5.3 Type II, style 1. Probe shall be double-ended, three-piece construction, consisting of two operative ends and a handle. The operative ends shall be permanently secured in the handles. Operative ends shall be non-calibrated. The finger grip portions of the handles shall be finely serrated or fine diamond knurled.

3.6 Finish. The operative ends shall have a mirror finish. The handles shall have a mirror or satin finish, except for the serrated or knurled portion. The operative ends shall be free from coarse grind marks, cracks, burrs, feathers, or other defects when viewed as specified in 4.3.1.

3.7 Identification marking. Type I, style 1, Glickman probe shall be marked "26G"; type I, style 2 probe shall be marked "4"; type II, style 1, Nabers probe shall be marked "2N". Marking shall be on the front face of the instruments. All markings shall be on the smooth portion of the handles and shall be permanent and legible. As an alternate to the word "stainless", the use of a registered trade name identifying the steel as stainless is acceptable, provided prior approval is received from the cognizant contracting officer (see 6.1).

3.8 Workmanship. Workmanship shall be first class throughout. The probes shall be free from defects which detract from their appearance or may impair their serviceability.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise specified in the contract, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the Item Identification, the name of the material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105 and table I. The unit of product for examination purposes shall be one instrument.

TABLE I. Sampling for examination.

Defect	Inspection level	AQL (percent defective)
Major A	II	1.0
Major B	S-1	2.5
Minor defects	II	4.0

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:39Z
Check the source to verify that this is the current version before use.

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105 and table II. The unit of product for test purposes shall be one instrument. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

TABLE II. Sampling for tests.

Component	Characteristic	Requirement	Test procedure	Inspection level	AQL
Operative end	Hardness	3.3 and GG-I-526	4.4.2	S-2 2/	1.5
Handle(416 and 430F steel only)	Copper Sulfate	3.2.2. and GG-I-526	GG-I-526	S-2	1.0
Entire probe	Boil test	3.2.1, 3.2.2 & GG-I-526	GG-I-526	S-2	1.0
Entire probe	Composition	3.2.1 & 3.2.2	4.4.1	See 4.2.2.1	1/

1/ Acceptance number shall be zero for all sample sizes.

2/ In lieu of sampling plans for normal inspection of MIL-STD-105, sampling plans for reduced inspection may be used when the inspection lot or batch complies with both of the below listed conditions:

a. consists of units manufactured from the same melt of metal, specified chemical composition, produced by the same manufacturer at the same plant, and using the same facilities at essentially the same time.

b. consists of units subjected to the same heat treatment processes simultaneously in the same equipment; except that when continuous belt automatic heat treatment equipment is used, a heat treatment lot shall be considered no more than one day's production under constant conditions.

4.2.2.1 For raw material. Sufficient amount from each lot of each metal used in the construction of the probes shall be selected for composition analysis.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

4.3 Examination. The probes shall be examined to determine compliance with specification requirements. Examination shall be conducted in accordance with the classification of defects reflected below in table III. Any probes in the sample containing one or more defects shall be rejected, and if the number of defective probes in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE III. Classification of defects.

Categories	Defects
Major A:	
101	Style and design not as specified for the applicable type and style.
102	Instrument not constructed as specified.
103	Operative end finish not as specified.
104	Operative end not free from cracks and burrs.
Major B:	
151	Operative end dimensions not as specified.
Minor:	
201	Handle finish not as specified.
202	Instrument not permanently and legibly marked as specified.
203	Instrument identification not properly located.
204	Handle dimensions not as specified.

1/ Examination shall not be restricted to the classified possible defects listed.

4.3.1 Microscopic examination. Examination of probe defects not readily perceptible to the naked eye shall be made using a microscope at 4 to 10 power (4X to 10X) magnification (see 3.6).

4.3.2 Packaging inspection. The inspection of the preservation-packaging and interior package marking shall be in accordance with the group A and B quality conformance inspection requirements of MIL-P-116. The inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification and the marking requirements of MIL-STD-129.

4.4 Tests. Tests shall be conducted to determine compliance with specification requirements.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:39Z
Check the source to verify that this is the current version before use.

4.4.1 Chemical analysis. In the absence of certificates of quality for the raw material, chemical analysis shall be conducted in accordance with method 111 or method 112 of FED-STD-151. In case of dispute, method 111 shall govern.

4.4.2 Hardness. Hardness shall be determined in accordance with ASTM Designation E 18 (Rockwell) or E 92 (Vickers or Diamond Pyramid).

5. PACKAGING

5.1 Preservation-packaging. Preservation-packaging shall be level A or C, as specified (see 6.1).

5.1.1 Level A.

5.1.1.1 Cleaning. Probes shall be cleaned in accordance with MIL-P-116, process C-1.

5.1.1.2 Drying. Probes shall be dried in accordance with MIL-P-116.

5.1.1.3 Preservative application. Preservatives shall not be used.

5.1.1.4 Unit packaging. Each probe shall be packaged in accordance with MIL-P-116, method IC-2. As an alternate, each probe shall be packaged in a rigid plastic tube having a metal or plastic screw cap or a plastic plug-type closure. The probe shall be immobilized within the container to prevent damage to the operative ends.

5.1.1.5 Intermediate package. Twelve probes of one type and style only shall be packaged in a box conforming to PPP-B-566 or PPP-B-676. Closure shall be as specified in the appendix of the applicable box specification.

5.1.2 Level C. Probes shall be clean, dry, and individually packaged in a manner that will afford adequate protection against corrosion, deterioration, and physical damage during shipment from the supply source to the first receiving activity.

5.2 Packing. Packing shall be level A, B, or C, as specified (see 6.1).

5.2.1 Level A. Twenty-four intermediate packages containing probes of one type and style shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping, and waterproofing shall be as specified in the appendix of the box specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:39Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Twenty-four intermediate packages containing probes of one type and style shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the appendix of the box specification.

5.2.3 Level C. The packaged probes shall be packed in shipping containers in a manner that will afford adequate protection against damage during direct shipment from the supply source to the first receiving activity. These packs shall conform to the carrier's rules and regulations.

5.3 Marking. Each unit package, intermediate package, and shipping container shall be marked as specified in MIL-STD-129.

6. NOTES

6.1 Ordering data. Purchasers should select the preferred options permitted herein and include the following information in the procurement documents:

a. Title, number, and date of this specification.
b. Type and style required (see 1.2).
c. Selection of applicable levels of packaging and packing (see 5.1 and 5.2).
d. Approval of trade name, if applicable (see 3.7).

6.2 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6520-01-046-7329	PROBE, PERIODONTAL, Glickman, 26G
6520-01-047-0646	PROBE, PERIODONTAL, No. 4
6520-01-047-5629	PROBE, PERIODONTAL, Bifurcation, Nabers, 2N

Preparing activity:
DLA-DM

Project No. 6520-1503

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:39Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

INSTRUCTIONS: This form is provided to solicit beneficial comments which may improve this document and enhance its use. DoD contractors, government activities, manufacturers, vendors, or other prospective users of the document are invited to submit comments to the government. Fold on lines on reverse side, staple in corner, and send to preparing activity. Attach any pertinent data which may be of use in improving this document. If there are additional papers, attach to form and place both in an envelope addressed to preparing activity. A response will be provided to the submitter, when name and address is provided, within 30 days indicating that the 1426 was received and when any appropriate action on it will be completed.

NOTE: This form shall not be used to submit requests for waivers, deviations or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

DOCUMENT IDENTIFIER *(Number)* AND TITLE

MIL-P-37928(DLA-DM) PROBES, PERIODONTAL

NAME OF ORGANIZATION AND ADDRESS OF SUBMITTER

☐ VENDOR ☐ USER ☐ MANUFACTURER

1. ☐ HAS ANY PART OF THE DOCUMENT CREATED PROBLEMS OR REQUIRED INTERPRETATION IN PROCUREMENT USE? ☐ IS ANY PART OF IT TOO RIGID, RESTRICTIVE, LOOSE OR AMBIGUOUS? PLEASE EXPLAIN BELOW.

A. GIVE PARAGRAPH NUMBER AND WORDING

B. RECOMMENDED WORDING CHANGE

C. REASON FOR RECOMMENDED CHANGE(S)

2. REMARKS

SUBMITTED BY *(Printed or typed name and address — Optional)*	TELEPHONE NO.
	DATE

DD FORM 1 OCT 76 1426 EDITION OF 1 JAN 72 WILL BE USED UNTIL EXHAUSTED.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:39Z
Check the source to verify that this is the current version before use.