MILITARY SPECIFICATION

SPINEBOARD

This specification is approved for use by the Defense Logistics Agency, Department of Defense, and is available for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers two types of spineboards for use in the immobilization of patients with injuries or possible fractures of the cervical, thoracic, or lumbar spine.

1.2 Classification. Spineboards shall be of the following types, as specified below:

I Spineboard, Long
II Spineboard, Short

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specification and standards. Unless otherwise specified (see 6.2), the following specifications and standards of the issue listed in that issue of the Department of Defense Index of Specifications and Standards (DODISS) specified in the solicitation, form a part of this specification to the extent specified below:

SPECIFICATIONS

FEDERAL

TT-V-121 - Varnish, Spar, Water Resisting
TT-F-336 - Filler, Wood, Paste.
PPP-B-636 - Boxes, Shipping, Fiberboard.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-ATT, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1425) appearing at the end of this document or by letter.

FSC 6530

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

MILITARY

MIL-P-116 - Preservation-Packaging, Methods Of.
MIL-L-19140 - Lumber and Plywood, Fire Retardent Treated.
MIL-A-22397 - Adhesive, Phenol And Resorcinol Resin Base.

STANDARDS

FEDERAL

FED-STD-123 - Marking for Domestic Shipment.

MILITARY

MIL-STD-105 - Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129 - Marking for Shipment and Storage.
MIL-STD-130 - Identification Marking of U. S. Military Property.
MIL-STD-171 - Finishing of Metal and Wood Surfaces.

DRAWINGS

DPSC Drawing No. 25503 Long Spineboard
DPSC Drawing No. 25502 Short Spineboard

(Copies of specifications and standards required by manufacturers in connection with specific acquisition functions should be obtained from the contracting activity or as directed by the contracting officer.)

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. The issues of the documents which are indicated as DoD adopted shall be the issue listed in the current DoDISS and the supplement thereto, if applicable.

DEPARTMENT OF COMMERCE

Product Standard: PS-1-Construction and Industrial Plywood.

(Application for copies should be addressed to the Superintendent of Documents, U. S. Government Printing Office, Washington, DC 20402.)

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence.

3. REQUIREMENTS

3.1 First Article. When specified, a sample shall be subjected to first article inspection (see sections 4 and 6).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

3.2 Fire and casualty hazards

3.2 The contractor shall submit proof, prior to delivery, that the items to be delivered under this specification conform to the applicable requirements of MIL-L-19140 (see 4.1.3).

3.3 Materials. Except as herein specified, materials for both types of spineboards shall conform to the applicable portions of the specifications, standards, and drawings listed in section 2, and shall meet the requirements shown therein. When materials to be used are not specifically designated, those used shall be entirely suitable for the purpose intended.

3.3.1 Plywood. The plywood shall conform to U. S. Commerce Dept. Product Standard PS-51, Type I, Grade 2 and be fire retardant treated in accordance with MIL-L-19140. The wood shall have at least five plies and shall be finished smooth on all surfaces.

3.3.2 Runners (Type I Spineboard Only). The runners shall be of red or white oak and have the dimensions indicated in the applicable drawing.

3.3.3 Screws (Type I Spineboard Only). The screws used in attaching the runners to the spineboard shall be commercial grade #10 oval head, brass, 1-3/4 inches long, and shall conform to the applicable drawing.

3.3.4 Adhesive (Type I Spineboard Only). The adhesive used in attaching the runners to the spineboard shall be high in both wet and dry strength and conform to MIL-A-22397.

3.4 Design and construction.

3.4.1 Long Spineboard. The long spineboard, Type I, consists of a plywood sheet with handholes and with runners, for ground clearance, fastened to the bottom. The board shall be constructed in accordance with the applicable drawing (see 2.1).

3.4.2 Short Spineboard. The short spineboard, Type II, consists of a plywood sheet, with handholes, and shall be constructed in accordance with the applicable drawing (see 2.1).

3.5 Finish and Color. The color of the spineboards shall be that of the natural wood from which they are made. The boards shall be medium sanded, sealed with paste wood filler which meets TT-F-336, and coated with water resisting spar varnish which meets TT-V-121. This finish conforms to finish 28 2 of MIL-STD-171.

3.5.1 Alternate Finish. Prior to the application of the finish on the spineboards, a contractor may submit an alternate method or process for finishing. The contractor must submit proof (see 4.1.3) that such methods or processes are equivalent to that specified in 3.6. In all cases the contractor shall obtain approval for the use of such alternate process from the contracting officer prior to its application to the spineboards (see 6.2).

3.6 Workmanship. The spineboards shall be free from defects which detract from their appearance or may impair their serviceability. The finished boards shall be smooth and free of splinters.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

4. QUALITY ASSURANCE PROVISIONS

4.1 Responsibility for inspection. Unless otherwise in the contract or purchase order, the contractor is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or purchase order, the contractor may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribed requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual and auditory investigation without the use of special laboratory appearances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2.1 Classification of inspection. The inspection requirements specified herein are classified as follows:

a. First article inspection
b. Quality conformance inspection.

4.1.2.1.1 First article inspection. First article inspection shall be performed on the spineboard when a first article sample is required (see 3.1). This inspection shall include examination of 4.3.

4.1.2.1.2 Quality conformance inspection. Quality conformance inspection shall consist of the examination of 4.3.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the materials may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the NSN, the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Inspection lot

4.2.1 Raw material. An inspection lot for raw material inspection shall be all material produced under the same conditions and controlled to the degree that uniformity of quality within the designated lot can be assured. Lot identification of the raw material by the manufacturer is one method of control, as cited above, which is considered acceptable.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

4.2.2 End product. An inspection lot for inspection of end product shall consist of all spineboards of the same type produced under essentially the same conditions and offered for inspection at one time.

4.3 Sampling.

4.3.1 Sampling for examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and table I. The unit of product for examination purposes shall be one spineboard.

TABLE I. Sampling for examination.

	Inspection level	AQL (defects per hundred unit)
For visual examination		
Major defects	S-4	2.5
Total defects (Major and Minor combined)	S-4	6.5
For dimensional examination	S-4	2.5

4.4 Examination. The spineboards shall be examined to determine compliance with all requirements contained in this specification, including examination of the spineboards to insure that they are flame retardent

4.4.1 Classification of defects. Examination shall be conducted in accordance with the classification of defects listed in table II. Examination shall not be restricted to the classified possible defects listed below

TABLE II. Classification of defects.

Categories	Defects
Major	
101	Not finished as specified (sanded, sealed, varnished)
102	Runners not properly attached (Type I only).
Minor	
201	Not free of slivers and sharp edges.
202	Not smooth.

4.4.2 Dimensional examination. The spineboards shall be examined for defects in dimensions. Any dimension not within the tolerances specified herein shall be classified as a defect.

5. PACKAGING

5.1 Packaging. Packaging shall be Level A or Level C as specified (see 6.2).

5.1.1 Unit of issue. One spineboard of the type specified, constitutes a unit of issue.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

5.1.2 Level A.

5.1.2.1 Unit packaging. Each spineboard shall be clean, dry and wrapped in kraft paper and the wrapping secured with kraft tape.

5.1.3 Level C. Each unit shall be packaged as specified in 5.1.2.1.

5.2 Packing.

5.2.1 Level A. One spineboard of the type specified shall be packed in a shipping container designed for a type 2 load and conforming to PPP-B-636, class weather-resistant. Closure, strapping and waterproofing shall be as specified in the box specification.

5.2.2 Level B. One spineboard of the type specified shall be packaged in a box designed for a type 2 load and conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.2.3 Level C. One spineboard of the type specified shall be packed in a manner that will afford adequate protection against damage during shipment from the supply source to the receiving activity. This pack shall conform to the carrier's rules and regulations.

5.3 Marking. Marking shall be as specified in MIL-STD-129.

6. NOTES

6.1 Intended Use. The spineboards covered by this specification are intended to be used, stored, and transported in air and ground ambulances, and to be used, stored, and transported as a part of the equipment complement of field medical facilities for the care and treatment of patients with back or neck injuries.

6.1.1 Intended Use, Type I. The long spineboard is intended for use in the immobilization of patients having upper and or lower back injuries.

6.1.2 Intended Use, Type II. The short spineboard is intended for use in the extrication and immobilization of patients having neck injuries.

6.2 Ordering data. Purchasers should select the preferred options permitted herein and include the following information on the procurement documents.

6.2.1 Title, Number and Date of this Specification.

6.2.2 National Stock Number.

6.2.2.1 Dimension.

6.2.3 Selection of Applicable Levels of Packaging and Packing (see 5.1 and 5.2).

6.2.4 First Article Inspection (see 3.1, 4.1.2.1.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

6.3 This specification covers the following items appearing in the Federal Supply Catalog:

National Stock Number	Item Identification
6530-01-119-0012	Spineboard, Long, Type I
6530-01-119-0011	Spineboard, Short, Type II

MILITARY INTERESTS:

Custodians:

Army-MD
Navy-MS
Air Force-03

Preparing activity:

DoD-MB

Agent:

DLA-DM

Project No. 6530-1592

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements All users of military standardization documents are invited to provide suggestions This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems Enter in block 6 any remarks not related to a specific paragraph of the document If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered

NOTE This form may not be used to request copies of documents nor to request waivers deviations, or clarification of specification requirements on current contracts Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements

(Fold along this line

(Fold along this line

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL

FIRST CLASS PERMIT NO 16[illegible] WASH[illegible]

POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

DEFENSE PERSONNEL SUPPORT CENTER
DIRECTORATE OF MEDICAL MATERIEL

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions - Reverse Side)

1 DOCUMENT NUMBER	2 DOCUMENT TITLE
MIL-S-37952	

3a. NAME OF SUBMITTING ORGANIZATION	4 TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR
	☐ USER
b. ADDRESS *(Street, City, State, ZIP Code)*	☐ MANUFACTURER
	☐ OTHER *(Specify)* _____

5 PROBLEM AREAS

a. Paragraph Number and Wording

b. Recommended Wording

c. Reason/Rationale for Recommendation

6 REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* — Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* — Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* — Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 1426 82 MAR PREVIOUS EDITION IS OBSOLETE

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:40Z
Check the source to verify that this is the current version before use.