MILITARY SPECIFICATION SHEET

SPLINT, LEG THOMAS, HALF-RING, TELESCOPING

1. SCOPE

1.1 Scope. This specification covers a telescoping splint, suitable for use in immobilization of fractures and treatment of injuries to the lower extremeties.

2. APPLICABLE DOCUMENTS

2.1 The following documents of the issue in effect on date of invitation for bids or request for proposal, form a part of the specification to the extent specified herein.

SPECIFICATIONS

FEDERAL

QQ-A-225/8	- Aluminum Alloy 6061, Bar Rod, Wire and Special Shapes, Rolled, Drawn, or Cold Finished
V-T-285	- Thread, Polyester
WW-T-700/6	- Tube, Aluminum Alloy, Drawn, Seamless, 6061
PPP-B-636	- Boxes, Shipping, Fiberboard

MILITARY

MIL-P-116	- Preservation, Methods of
MIL-F-495	- Finish, Chemical, Black, For Copper Alloys
MIL-W-530	- Webbing, Textile, Cotton, General Purpose, Natural Or In Colors
MIL-T-5550	- Thread And Twine Mildew-Resistant or Water Repellent Treated
MIL-T-5038	- Tape, Textile and Webbing, Textile Reinforcing Nylon
MIL-R-6130	- Rubber, Cellular, Chemically Blown
MIL-A-8625	- Anodic Coatings, For Aluminum And Aluminum Alloys
MIL-C-13924	- Coating, Oxide, Black, For Ferrous Metals
MIL-T-43566	- Tape, Textile, Cotton, General Purpose, Natural Or In Colors

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

STANDARDS

FEDERAL

*FED-STD-595 - Color

FED-STD-751 - Stitches, Seams and Stitchings

MILITARY

MIL-STD-105 - Sampling Procedures and Tables For Inspection By Attributes

MIL-STD-129 - Marking For Shipment and Storage

MIL-STD-794 - Parts And Equipment, Procedures for Packaging of

DRAWINGS

Defense Personnel Support Center (DPSC)

DWG NO. 25084 - SPLINT, LEG, THOMAS, HALF-RING, TELESCOPING

(Copies of specifications, standards, drawings and publications required by suppliers in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

*NOTE: Color chips are available from:

Commissioner
Federal Supply Service
General Services Administration
Standardization Division, Paints Branch
Washington, DC 20406

2.2 Other publications. The following documents form a part of this specification to the extent specified herein. Unless otherwise indicated, the issue in effect on date of invitation for bids or request for proposal shall apply.

AMERICAN SOCIETY FOR TESTING AND MATERIALS

ASTM D2240-68 - Indentation Hardness of Rubber and Plastics By Means Of A Durometer, Test For

(Application for copies should be addressed to the American Society for Testing and Materials, 1916 Race Street, Philadelphia, Pa 19103).

COLOR ASSOCIATION OF THE UNITED STATES, INC.

United States Army Standard Color Card For the Official Standardized Shapes Of Sewing Threads

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

(Application for copies should be addressed to the Color Association of the United States, 200 Madison Avenue, New York, NY 10016).

3. REQUIREMENTS

3.1 Material.

3.1.1 Half-ring, and telescoping rod. The half-ring, and telescoping rod shall be aluminum alloy 6061, temper T-6 in accordance with WW-T-700/6.

3.1.2 Long and short extension tubing. The long and short extension tubing shall be aluminum alloy 6061, temper T-6 in accordance with QQ-A-225/8.

3.1.3 Collet and sleeve. The collet and sleeve shall be of C1015 carbon steel.

3.1.4 Rivets. The rivets used to connect the long and short extension tubing to the half-ring and the strap and buckle assembly shall be 2017-T4.

3.1.5 Cushion, foam. The closed cell foam cushion shall be homogeneous and shall conform to the applicable drawing and with MIL-R-6130, Type II, Grade C, Soft. The foam cushion may be cut or molded to meet the applicable specification and drawing.

3.1.6 Cover, cushion. The cushion cover shall be a dip coat of polychloroprene and shall conform to the applicable drawing.

3.1.7 Strap. The strap shall be type IV nylon webbing in accordance with MIL-T-5038.

3.1.8 Buckle. The buckle shall be brass.

3.1.9 Thread. The thread shall be in accordance with V-T-285. The bobbin thread shall be type I, Class 1, size FF, and the needle thread shall be the same.

3.1.10 Sling. The sling shall be cotton textile tape, type I, Class 4, in accordance with MIL-T-43566.

3.1.11 Fastener, hook and pile. The sling shall have hook and pile fasteners in accordance with MIL-F-21845. Hook shall be Type II, Class 1, and pile shall be Class 1.

3.1.12 Fastener, Snap. The snap fastener shall be of brass with steel reinforcement and composed of button and socket which mates with the stud and eyelet. The finish shall be black in accordance with MIL-F-495.

3.1.13 Spacer, nylon. The spacer between the half-ring and the long and short extension tubing shall be of nylon type 101.

3.1.14 Eyelet and washer. The eyelet and washer on the ends of the straps where they are fastened to the half-ring and extension tubing shall be brass.

3.2 Design and construction. The splint shall be designed and constructed in accordance with DPSC drawing 25084 and as specified herein.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

3.2.1 Assembly of components. Assembled components shall be finished smooth and shall be free of cracks or other irregularities. Screws and rivets shall be set securely. The finished half-ring shall be capable of rotating freely in each direction to approximately 90°.

3.2.2 Cushion. The cushion shall be of foam construction and covered with a seamless dip coating of polychloroprene rubber. The coating shall be smooth and continuous. The fit between the cushion and the half-ring shall be tight.

3.2.3 Straps. There shall be two straps on each splint. The webbing used for the straps shall be 1-1/2 inches wide uniformly woven with selvaged edges. Each strap shall be secured to the half-ring and extension tubing by a rivet in accordance with the applicable drawing. The strap to which the three prong buckle with guard is sewn shall be attached on the side of the splint with the long extension tubing. The free end of the strap attached to the shorter side of the splint shall have a diagonal cut as specified on the applicable drawing.

3.2.4 Slings. There shall be four slings for each splint, three large and one small. The tape used for the slings shall be 3 inches wide, uniformly woven with selvaged edges. The slings shall conform to the applicable drawing and shall be secured to the splints with snap fasteners. The slings shall be free to slide on the extension tubing and telescoping rod.

3.2.5 Stitching on webbbing and tape. All stitching on webbing and tape shall be stitch type 301 in accordance with FED-STD-751. There shall be not less than 7 stitches per inch. All stitching shall be backstitched at the ends to prevent raveling.

3.3 Performance.

3.3.1 Hardness. The cushion with cover shall have a durometer hardness of 25 ± 5 when tested as specified in 4.4.1.

3.3.2 Flammability. The plastic cushion shall be self-extinguishing within five seconds when tested as specified in 4.4.2.

3.3.3 Security of Strap Attachment. Both straps shall withstand a 150 pound pull without becoming detached from the splint, when tested as specified in 4.4.3.

3.4 Finish and Color.

3.4.1 Cushion. The color of the finished cushion shall be black, or olive drab approximating color 34087 of FED-STD-595.

3.4.2 Webbing. Nylon webbing for the straps shall be dyed olive drab No. 7. The diagonally cut free end of the strap shall be singed to prevent unraveling.

3.4.3 Slings. The tape for the slings shall be vat-dyed olive drab No. 7 and treated with mildew inhibitor in accordance with Class 4 tape of MIL-T-43566.

3.4.4 Buckle. The buckle shall have a black chemical (oxide) finish in accordance with MIL-F-495.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

3.4.5 Thread. Thread shall be vat-dyed olive drab shade S-1 conforming to U.S Army Standard Color card.

3.4.6 Collets and Sleeves. The collets and sleeves shall have a finish consisting of a black oxide coating in accordance with MIL-C-13924.

4.2.3 For Tests on Components. Sufficient samples of the components shall be selected from each lot of material used in the manufacture of the splints to perform tests as required by the applicable subsidiary specifications or by tests specified herein. When sample size for components covered by the subsidiary specifications is not specified, the sample size for each test shall be as indicated for inspection level S-1 in accordance with MIL-STD-105, except that the acceptance number shall be zero for all sample sizes, and a minimum of two samples shall be taken regardless of lot size.

4.3 Examination.

4.3.1 Splints. The splints shall be examined to determine compliance with the requirements contained in this specification. The splint shall be examined for possible defects, including, but not restricted to, those listed in Table III.

TABLE III. CLASSIFICATION OF DEFECTS.

Classification	Defect
MAJOR	
101	Parts missing or not correct size.
102	Parts not made of specified material.
103	Parts not secured as required.
104	Cushion not seamless.
105	Tubing bent, dented, cracked or otherwise damaged.
106	Half-ring cannot be rotated easily.
107	Stitching on webbing not as specified.
MINOR	
201	Finish of any part not as specified.
202	Strap on short side of splint does not have diagonal cut.
203	Buckle not design specified.
204	Parts not free of flash, burs, or sharp edges.

4.3.2 Inspection of packaging. The sampling and inspection of the preservation and interior package marking shall be in accordance with groups A and B quality conformance inspection requirements of MIL-P-116. The sampling and inspection of the packing and marking for shipment and storage shall be in accordance with the quality assurance provisions of the applicable container specification shown in section 5 and the marking requirements of MIL-STD-129.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

4.4 Tests on the Finished Product. Tests shall be conducted to determine compliance with specification requirements.

4.4.1 Hardness. Hardness of the cushion shall be tested in accordance with ASTM Standard D2240-68, except that the test sample paragraph does not apply. The indentation point of the durometer shall be made at the highest point of the curve of the sample and the hardness reading taken when the entire length of the presser foot just makes contact with the sample. The average of 10 readings shall be used to determine compliance with 3.3.1.

4.4.2 Flammability. The test sample shall be suspended in a vertical plane, in a draft free area, with the ends of the cushion pointing downward and equidistant from a horizontal plane. A Bunsen Burner with a flame 1/2 to 3/4 inch in height shall be placed under one end of the cushion. The position of the Bunsen Burner shall be adjusted so that the flame tip just makes contact with the cushion. At the end of 15 seconds, the flame shall be removed. By means of a stopwatch, the time for the plastic to stop burning shall be measured in order to determine compliance with 3.3.2.

4.4.3 Security of Strap Attachment. With the splint secured, each strap shall be subjected to a 150 pound static load for a period of 1/2 hour, tangent to the half-ring at the point of attachment to the half-ring to determine compliance with 3.3.3.

5. Packaging

5.1 Preservation Preservation shall be level A or C as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Each splint shall be wrapped in kraft paper or other noncorrosive material. The wrap shall be secured with kraft or other suitable tape.

5.1.1.2 Intermediate package. Two unit packages shall be packaged in a snug fitting box constructed in accordance with PPP-B-636, class domestic. Closure of the box shall be as specified in the box specification.

5.1.2 Level C. The level C preservation for leg splints shall conform to the MIL-STD-794 requirements for this level.

5.2 Packing. Packing shall be level A, B or C as specified (see 6.2).

5.2.1 Level A. Twelve intermediate packages (24 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class weather-resistant. Closure, waterproofing and strapping shall be as specified in the appendix of PPP-B-636.

5.2.2 Level B. Twelve intermediate packages (24 units) shall be packed in an exterior container designed for a type 2 load and constructed in accordance with PPP-B-636, class domestic. Closure shall conform to method I to II as specified in the appendix of PPP-B-636.

5.2.3 Level C. The level C packing for leg splints shall conform to the MIL-STD-794 requirements for this level.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

5.2.4 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Quantities less than 40 cubic feet need not be unitized. Unitized loads shall be uniform in size and quantities to the greatest extent practicable.

5.2.4.1 Level A. Splints packed as specified in 5.2.1, shall be unitized on pallets in conformance with MIL-STD-147, load type 1, with a fiberboard cap (storage aid 4) positioned over the load.

5.2.4.2 Level B. Splints, packed as specified in 5.2.2, shall be unitized as specified in 5.2.4.1 except that the fiberboard caps shall be class domestic.

5.2.4.3 Level C. Splints, packed as specified in 5.2.3, shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. In addition to any special of other identification marking required by the contract (see 6.2), each unit pack, intermediate and exterior container and unitized load shall be marked in accordance with MIL-STD-129.

5.4 General.

5.4.1 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.3.2.

6.3 This specification is intended to cover the following item appearing in the Federal Supply Catalog:

National Stock No.	Item Identification
6515-01-071-6645	SPLINT, LEG, THOMAS, HALF-RING, TELESCOPING

Preparing activity:
DLA-DM

Project No. 6515-3232
TXAREL NS6545

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.