MILITARY SPECIFICATION

SURGICAL KIT, ORTHOPEDIC

This specification is approved for use by all Departments and Agencies of the Department of Defense.

1. SCOPE

1.1 Scope. This specification covers a battery operated, with charges, multipurpose, orthopedic surgical kit suitable for use in orthopedic surgery.

2. APPLICABLE DOCUMENTS

2.1 Government documents.

2.1.1 Specifications and standards. The following specifications and standards form a part of this specification to the extent specified herein. Unless otherwise specified, the issues of these documents shall be those listed in the issue of the Department of Defense Index of Specifications and Standards (DoDISS) and supplement thereto, cited in the solicitation.

Beneficial comments (recommendations, additions, deletions) and any pertinent data which may be of use in improving this document should be addressed to: Defense Personnel Support Center, Directorate of Medical Materiel, DPSC-RST, 2800 South 20th Street, Philadelphia, PA 19101, by using the self-addressed Standardization Document Improvement Proposal (DD Form 1426) appearing at the end of this document or by letter.

AMSC N/A FSC 6515
DISTRIBUTION STATEMENT A. Approved for public release; distribution is unlimited.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

SPECIFICATIONS

Federal

GG-I-526	- Instruments, Dental and Surgical; General Specification.
PPP-B-601	- Boxes, Wood, Cleated-plywood.
PPP-B-621	- Boxes, Wood, Nailed and Lock-corner.
PPP-B-636	- Boxes, Shipping, Fiberboard.

MILITARY

MIL-L-10547	- Liners, Case and Sheet, Overwrap; Water-Vaporproof or Waterproof, Flexible.

STANDARDS

MILITARY

MIL-STD-105	- Sampling Procedures and Tables for Inspection by Attributes.
MIL-STD-129	- Marking for Shipment and Storage.
MIL-STD-147	- Palletized Unit Loads.
MIL-STD-794	- Parts and Equipment, Procedures for Packaging of.

2.1.2 Drawing. The following drawing forms a part of this specification to the extent specified herein. Unless otherwise specified, the issues shall be those in effect on the date of the solicitation.

DRAWING

Defense Personnel Support Center (DPSC)

No. 22359	-	Surgical Kit, Orthopedic, Battery Operated with Charger, Multipurpose.

(Copies of specifications, standards, and drawing required by contractors in connection with specific acquisition functions shall be obtained from the contracting activity or as directed by the contracting activity).

2.3 Order of precedence. In the event of a conflict between the text of this specification and the references cited herein, the text of this specification shall take precedence. Nothing in this specification, however, shall supersede applicable laws and regulations unless a specific exemption has been obtained.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

3. REQUIREMENTS

3.1 General requirements. The orthopedic system shall be in accordance with Federal Specification GG-I-526 and as specified herein.

3.2 The orthopedic system shall consist of the following component items:

a) 750-800 RPM Drill with ¼ -inches Jacobs Chuck
b) Sagittal Saw
c) 1250-1350 RPM Wire Driver
d) 200-250 RPM Reamer
e) Battery Charger (For Dual Batteries)
f) Sterilization Case
g) 9 Drill Bits - 3/32, 7/64, 1/8, 9/64, 5/32, 11/64, 1/16, 5/64 and 3/16 inches.
h) 1/4-inch Jacobs Chuck Adapter
i) 4 "Extra Power" Battery Packs
j) 10 Sagittal Saw Blades:
   2 - 2.5 inches by 1.25 inches by 0.025 inches.
   2 - 2.5 inches by 0.60 inches by 0.025 inches.
   2 - 1.25 inches by 0.60 inches by 0.025 inches.
   2 - 2.5 inches by 1.25 inches by 0.036 inches.
   2 - 1.25 inches by 0.38 inches by 0.025 inches.

3.2.1 The 750-800 RPM Drill. The 750-800 RPM drill shall be cordless, self contained, battery power driven, supplying sufficient power for most orthopedic drilling, wire driving, and pinning procedures.

The shape, style and nominal dimensions of the drill shall be in accordance with DEFENSE PERSONNEL SUPPORT CENTER Drawing #22359.

All the stainless steel parts of the drill shall be properly passivated.

The drill shall be small and lightweight in design, facilitating use, maneuverability, and surgeon comfort. The weight shall be approximately 37-1/2 ounces with an overall length of 5-3/4 inches nominal. Exterior surface shall be a high impact thermoplastic.

The drill shall have a "no load" speed operation of 750-800 RPM. The unit shall be powered from a nickel cadmium battery pack which locks into a base of the working head and serves as the instrument handle.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:45Z
Check the source to verify that this is the current version before use.

The battery pack shall be fully sealed allowing ethylene oxide gas sterilization and steam sterilization in temperatures up to 275^{0}F. The shape, style, and nominal dimensions of the battery shall be in accordance with DPSC Drawing #22359. The sealed battery pack shall be able to be immersed for easy cleaning.

The 750-800 RPM Drill shall deliver approximately 25-35 inch-pounds of torque for up to 20 minutes of continuous operation on a single fully-charged battery pack. The battery shall be capable of being fully recharged within 40 minutes.

The drill shall be fully cannulated to accept wires and pins from 1/32 inch (0.8 mm) to 3/16 inch (4.7 mm).

The drill shall have a keyless Jacobs type chuck which can accept drill bits, wires, and pins ranging from 1/32 to 3/16 inch (0.8 to 4.7 mm).

The rotational direction of the drill shall be both forward and reverse and be activated by trigger finger-tip controls located at the inward base of the working head. This shall permit one-handed operation as well as simple shift from forward to reverse.

The Jacobs chuck shall be an integral part of the working head. It shall be a quick insert-release type chuck and be supplied with a chuck key for secure tightening and loosening of the pins, wires, or drill bits.

The base of the working head shall have a quick insertion-release latch. This shall securely seat the battery pack handle onto the working head. Interchangeable battery power packs shall snap off-and-on within seconds when extended operation is desired.

The cannulation of the drive shaft shall allow pins and wires up to 3/16 inch (4.7 mm) diameter to pass through the keyless chuck and drill housing.

3.2.2 The Sagittal Saw. The sagittal saw shall be a cordless, self-contained battery power driven unit supplying sufficient power to perform most orthopedic cutting procedures.

The shape, style and nominal dimensions of the sagittal saw shall be in accordance with DPSC Drawing #22359.

All the stainless steel parts of the instrument shall be properly passivated.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

The sagittal saw shall be small and lightweight in design facilitating use, maneuverability, and surgeon comfort. The weight shall be approximately 33 ounces with an overall length of 6-1/8 inches nominal. The exterior surface shall be of high impact ultem thermoplastic. The power head shall enclose the electrical motor and other mechanical components of the saw. The working nose tip and some inner components shall be constructed of corrosion-resisting or anodized aluminum.

The sagittal saw shall have a "no load" speed operation of approximately 13,000-15,000 strokes per minute.

The unit shall be powered from a nickel cadmium battery pack which locks into the base of the working head and serves as the instrument handle. The battery pack shall be fully sealed allowing ethylene oxide gas sterilization in temperatures up to 275^{0}F. The sealed battery pack shall be able to be immersed for easy cleaning. Shall deliver approximately 20 minutes of power on a single fully-charged battery pack. The battery shall be capable of being fully recharged within 40 minutes. The battery pack shall be made of ultem thermoplastic which houses the rechargeable nickel cadmium energy cells that supply electrical power to the power head.

When activated, the sagittal saw blades shall oscillate at 11,000 or more strokes per minute for quick, efficient operation.

The sagittal saw blades shall have an operational stroke capability of a 4-1/2 degrees total.

The sagittal saw shall be activated by a trigger finger-tip control located at the inward base of the working head. This shall permit one-handed operation.

The blade orientation shall be infinitely variable throughout a 360 degree swivel of the instrument nose. This blade connecting nose shall rotate to at least 12 positions. The blade shall be able to be connected at the nose in at least five different positions. This variable blade orientation and rotating nose shall facilitate making different cuts by eliminating awkward operating angles or positions.

The blade connecting nose shall be of a keyless design allowing rapid and various longitudinal and lateral blade positioning without the need for special tools and eliminating the danger of loose parts. Once locked in place, the blades shall be secure and not become loosened even under maximum operating power.

The base of the working head shall have a quick insertion-release latch. This shall securely seat the battery pack handle onto the operating head. Interchangeable battery power packs shall snap off-and-on within seconds when extended operation is desired.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:45Z
Check the source to verify that this is the current version before use.

3.2.3 The 1250-1350 RPM Drive The 1250-1350 RPM wire drive shall be a cordless, self-contained, battery powered driven unit supplying sufficient power for most orthopedic wire driving procedures.

The shape, style and nominal dimensions of the wire driver shall be in accordance with DPSC Drawing #22359.

All the stainless steel parts of the instrument shall be properly passivated.

The wire driver shall be small and lightweight in design facilitating use, manueverability, and surgeon comfort. The weight shall be approximately 37 ounces with an overall length of 5-3/4 inches nominal.

The exterior surface shall be made of high impact ultem thermoplastic. The power head encloses the electric motor and the other mechanical components of the wire driver. The working nose tips and some inner components shall be constructed of stainless steel or anodized aluminum.

The wire driver shall have a "no load" speed operation of approximately 1250-1350 RPM.

The unit shall be powered from a nickel cadmium battery pack which locks into the base of the working head and serves as the instrument handle. The battery power pack shall be fully sealed allowing ethylene oxide gas sterilization in temperatures up to 275°F. The sealed battery power pack shall be able to immersed for easy cleaning. Shall deliver approximately 20 minutes of power on a single fully-charged battery power pack. The battery shall be capable of being fully recharged within 40 minutes. The battery power pack shall be made of ultem thermoplastic, which houses the rechargeable nickel cadmium energy cells that supply electrical power to the power head.

The wire driver shall deliver 15-20 inch-pounds of torque for up to 20 minutes of continuous operation on a single fully-charged battery pack. The battery shall be capable of being fully recharged within 40 minutes.

The wire driver shall be fully cannulated to accept wire and pins from 0.028 inch to 0.062 inch.

The chuck design nose shall accept wires and pins from 0.028 inch to 0.062 inch.

The rotational direction shall be forward and shall have a quick release mechanism. The trigger finger-tip controls shall be located at the inward base of the working head. This shall permit one-handed operation as well as a simple shift from forward to open-release.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:45Z
Check the source to verify that this is the current version before use.

The fully cannulated chuck shall be of a keyless automatic design allowing single-handed use and eliminating the need to remove hands from the procedure.

The base of the working head shall have a quick insertion-release latch. This shall securely seat the battery pack handle onto the working head.

Interchangeable battery power packs shall snap off-and-on within seconds when extended operation is desired.

3.2.4 THE 200-250 RPM REAMER. The 200-250 RPM reamer shall be a cordless, self-contained, battery power driven unit supplying sufficient torque power for even the most difficult surgical reaming procedures, including acetabular reaming.

The shape, style, and nominal dimensions of the reamer shall be in accordance with DPSC Drawing #22359.

All the stainless steel parts of the instrument shall be properly passivated.

The reamer shall be small and lightweight in design facilitating use, maneuverability, and surgeon comfort. The weight shall be 6-3/4 approximately 35-1/2 ounces with an overall length of inches nominal. The exterior surface shall be of high impact ultem thermoplastic. The power head shall enclose the electrical motor and the other mechanical components of the reamer. The working Hudson head and some inner components shall be constructed of stainless steel or anodized aluminum.

The reamer shall have a "no-load" speed operation of approximately 200-250 RPM.

The unit shall be powered from a nickel cadmium battery power pack which locks into the base of the working head and serves as the instrument handle. The battery pack shall be fully sealed allowing ethylene oxide gas sterilization in temperatures up to 275^{0}F. The sealed battery pack shall be able to be immersed for easy cleaning. Shall deliver approximately 20 minutes of power on a single fully-charged battery pack. The battery shall be capable of being fully recharged within 40 minutes. The battery pack shall be made of polysulfone thermoplastic which houses the rechargeable nickel cadmium energy cells that supply electrical power to the power head.

The reamer shall be able to deliver approximately 65-80 inch-pounds of torque for up to 20 minutes of continuous operation on a single, fully-charged battery pack.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

The reamer shall be fully cannulated to accept all standard guide wires of up to 5/32 inch (4 mm) in diameter.

The reamer shall have a Hudson type chuck which will accept all Hudson-end reamers, and attachments, requiring no tools to assure quick positive attachment of Hudson shank accessories.

The rotational direction of the reamer shall be both forward and reverse and be activated by trigger finger-tip controls located at the inward base of the working head. This shall permit one-handed operation as well as a simple shift from forward to reverse.

The Hudson chuck shall be an integral part of the working head. It shall be a quick insert-release type chuck which shall securely seat all Hudson type reamers or attachments. Shall also accept and securely seat and lock the component 1/4-inch Jacobs chuck adapter.

The base of the working head shall have a quick insertion release latch. This shall securely seat the battery pack handle onto the working head. Interchangeable battery power packs shall snap off-and-on within seconds when extended operation is desired.

3.2.5 THE BATTERY CHARGER. The battery charger shall be capable of simultaneously charging two batteries.

The shape, style and nominal dimensions shall be in accordance with DPSC Drawing #22359.

The charger housing shall be made out of durable polycarbonate thermoplastic and it shall enclose the electrical, electromechanical, and electronic components of the charger.

The charger shall be micro-processor controlled producing a safe, automatic charge and battery testing capability.

Each terminal on the charger shall have a built-in circuit breaker which shall protect the unit from any power pack with an open circuit or shorted cell.

The charger shall tolerate storage for extended periods at 130^0F (57^0C) 40% relative humidity and -10^0F (-23^0C) 0% relative humidity.

Shall have a "power-on" button which will activate charging and illuminate following depression.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

Shall have dual indicator light panels depicting the charging status of each individual battery. Three different colored lights shall be used on the indicator panels. One shall illuminate indicating the need for battery replacement; another shall illuminate when a battery is being charged; a different colored light shall illuminate when a battery is fully charged.

Shall have a hospital grade power cord of at least six feet in length.

The electrical characteristics shall be consistent with the requirements contained in specification DPSC-MWRA-AT(DM).

3.2.6 THE STERILIZATION CASE. The shape, style and nominal dimensions of the sterilization case shall be in accordance with DPSC Drawing #22359.

The sterilization case shall be in accordance with GG-I-526 and as specified herein.

The nominal dimensions of the case shall be 11-5/8 inches in length by 7-1/4 inches wide and 2-3/4 inches high.

Shall accommodate 4 battery power packs and the drill, reamer, wire driver, and sagittal saw working heads.

The top, bottom and sides of the tray shall be perforated with hundreds of minute holes for maximum steam or ethylene oxide penetration of case contents.

The case shall be flat surfaced without a handle facilitating stacking and storage. Shall have a hinged top with a simple flip lock closure mechanism for securing the lid during handling and sterilization. The lock shall be located on the front panel.

Shall be fabricated from corrosion-resistant austenitic steel. The finish of the instrument shall be mirror. The complete instrument shall be properly passivated.

3.2.7 THE DRILL BITS. The drill bits shall be in accordance with GG-I-526 and as specified herein.

The shape, style, and nominal dimensions shall be in accordance with DPSC Drawing #22359.

All nine drill bits (3/32, 7/64, 1/8, 9/64, 5/32, 11/64, 1/16, 5/64 and 3/16 inch) shall be compatible for use with the component 750-800 RPM drill with the 1/4 inch Jacobs chuck.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:45Z
Check the source to verify that this is the current version before use.

The proximal end shall be round for use with the compound drill and other type drills with Jacobs chuck.

The distal portion shall have sharp edged flutes. The extreme distal tip shall be sharp.

The drills shall be able to withstand repeated sterilization by steam or ethylene oxide or long (10 hour) soaking in high level disinfectant solution.

Shall have the drill size or the company stock number clearly imprinted on the drill bit shaft.

3.2.8 THE JACOBS CHUCK ADAPTER. The Jacobs chuck adapter shall be in accordance with GG-I-526 and as specified herein.

The shape, style and dimensions shall be in accordance with DPSC Drawing #22359.

The instrument shall be a Jacobs chuck on the distal end with a Hudson proximal end. Shall securely seat and lock onto the component Hudson chuck 200-250 RPM reamer.

The adapter shall accept all pins, drills, and reamers up to 1/4 inch.

3.2.9 THE BATTERY PACKS. The shape, style and nominal dimensions of the battery packs shall be in accordance with DPSC Drawing #22359.

The nickel cadmium battery packs shall be sealed and shall withstand ethylene oxide gas sterilization and steam sterilization of temperatures up to 275^{0}F.

The sealed battery pack shall be able to be completely immersed for easy cleaning. The battery shall lock onto the working heads of the component drill, reamer, wire driver, and sagittal saw. The packs shall be designed to serve as handles when attached to the working heads. When fully charged, each battery pack shall provide 20 minutes of continuous operation to the attached component working head. Shall be capable of being fully recharged within 40 minutes.

The batteries shall be capable of being stored indefinitely at temperatures ranging from minus 50 degrees F to 122 degrees F without damage. Shall also tolerate storage at temperatures above 122 degrees F for brief periods.

If a battery loses all or part of its charge through self discharge, it shall be returned to full service with one or two charge/discharge cycles.

The battery packs shall not be affected adversely in high altitude environments.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:45Z
Check the source to verify that this is the current version before use.

3.2.10 THE SAGITTAL SAW BLADES. The sagittal saw blades shall be in accordance with GG-I-526 and as specified herein.

The shape, style and dimensions shall be in accordance with DPSC Drawing #22359.

All ten sagittal saw blades shall be compatible for use with the component sagittal saw. The proximal end shall have a hexagonal shaped opening with one area open to permit attachment to the distal operating nose of the sagittal saw. The five full angles of the hexagonal shaped opening shall permit blade attachment in five different positions. Shall have sharp cutting teeth at the distal portion of the blade.

Shall be able to withstand repeated sterilization be steam or ethylene oxide or long term (10 hour) soaking in high level disinfectant solution. Shall have the blade size or company stock number clearly imprinted on the blade shaft.

The orthopedic kit shall be provisioned in accordance with the requirements contained in specification DPSC-MWRA-AT(DM).

All component power heads -- the 750-800 RPM drill, sagittal saw, 1250-1350 RPM wire driver, 200-250 RPM reamer -- shall be capable of being stored for extended periods at temperatures ranging from -10^0F (-23^0C) 0% relative humidity to 135^0F (57^0C), 40% humidity.

All component items shall comply with the specification of this solicitation and with the manufacturer's own commercial specifications.

Shall be furnished with a minimum of five operative manuals which shall include: product diagrams, power pack installation and removal; operating instructions on every aspect of each individual instrument; cleaning and sterilization instruction; trouble shooting; and service/repair procedures.

3.3 Identification marking. The drill proper and chuck shall be permanently marked with the name or registered trade mark of the contractor, and the country or origin, if other than the U.S.

3.4 Service data. The contractor shall enclose, with each kit assembly two copies each of one or more booklets which contain complete instructions for installation, operation, maintenance, and a list of component parts. Each part of the unit shall be identified in the service data and parts list by the manufacturer's part number. As soon as practicable after award of the contract or purchase order, the contractor shall furnish to the contracting officer, via the cognizant Government inspector, two preliminary copies of his proposed service

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

data. for review and comment by the Defense Personnel Support Center, Directorate of Medical Materiel. The cover of each service manual should indicate the following:

NSN: 6515-01-177-4928
Surgical Kit, Orthopedic, battery operated, with charger, multipurpose
Contractor's trademark, Model Number, Name and Address

The preliminary service data shall be submitted sufficiently early to permit adequate review by the Defense Personnel Support Center (based on a maximum period of 30 days for such review) and to allow incorporation by the contractor of required revisions and corrections in the final service data without delaying delivery under the contract or purchase order. In addition, the contractor shall mail 15 copies of the approved service data to the contracting officer for distribution.

3.5 Workmanship. Workmanship shall be first class throughout. Drill assembly shall be free from burrs, cracks, pits, sharp edges, scale, corrosion and other defects which detract from its appearance or impair its serviceability.

4. QUALITY ASSURANCE PROVISIONS

4.1 Supplier responsibility for inspection. Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified in the contract or order, the supplier may use his own or any other facilities suitable for the performance of the inspection requirements specified herein, unless disapproved by the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure supplies and services conform to prescribe requirements.

4.1.1 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of 3 years after delivery of the supplies to which such records relate.

4.1.2 Inspection. Inspection, as used in this specification, is defined as both examination (such as visual or auditory investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

Source: https://assist.dla.mil -- Downloaded: 202640-21T11:45Z
Check the source to verify that this is the currentversion before use.

4.1.3 Certificates of quality. Certificates of quality, supplied by the manufacturer of the steel alloys and copper alloys, may be furnished in lieu of actual performance of such testing by the contractor, provided lot identity has been maintained and can be demonstrated to the Government. The certificate shall include the name of the contractor, the contract number, the name of the manufacturer or supplier, the National Stock Number (NSN), the item identification, the name of the component/material, the lot number, the lot size, the sample size, the date of testing, the test method, individual test results, and the specification requirements.

4.2 Sampling.

4.2.1 For examination. Sampling for examination shall be conducted in accordance with MIL-STD-105, and as indicated in Table II. The unit of product for examination purposes shall be one complete kit assembly (see 6.2).

TABLE II. Sampling criteria for examination

Type Defect	Inspection Level	AQL (percent defective)
Major	II	1.0
Minor	II	2.5

4.2.2 For tests. Sampling for tests shall be conducted in accordance with MIL-STD-105, and as indicated in Table III. The unit of product for test purposes shall be one complete kit assembly, (see 6.2).

TABLE III. Sampling criteria for tests

Component	Characteristic	Requirement	Test Procedure	Inspection Level	AQL
Surgical Kit	Dimensions	3.1 & 3.2	---	S-2	2.5
Stainless Steel Parts	Hardness	3.1, 3.2 & GG-I-526	GG-I-526	S-2	1.5
	Boil Test	3.1, 3.2 & GG-I-526	GG-I-526	S-2	1.0
	Copper Sulfate test (class 4 material)	3.1, 3.2 & GG-I-526	GG-I-526	S-2	1.0

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

4.2.3 Raw materials. A sufficient quantity of each lot of each metal used in the construction shall be selected for chemical analysis.

4.2.4 Sample unit usage. Where feasible, the same sample unit shall be used for the determination of two or more test characteristics.

4.3 Examination of assemblies. The orthopedic kit shall be examined to determine compliance with all requirements continued in this specification. Examination shall be conducted as specified in Tables II and IV. Examination, however, shall not be restricted to defects listed in Table IV. Any items in the sample containing one or more defects shall be rejected and if the number of defective orthopedic kits in any sample exceeds the acceptance number for that sample, the lot represented by the sample shall be rejected.

TABLE IV. Classification of defects

Categories	Defects
Critical	Non defined.
Major	
101	Orthopedic kit not of the design specified.
102	Components not free from cracks, pits, burrs, sharp edges, scale and corrosion.
103	Kit not furnished with all specified components.
104	Finish not as specified.
Minor	
201	Identification markings not legible, not complete, not correct, or missing.
202	Service data not functional.

4.3.1 Examination for preparation for delivery. An examination shall be made to determine whether the packaging, packing, and marking comply with the requirements of Section 5. Defects shall be scored as specified in Table V. sampling shall be in accordance with MIL-STD-105. The sample unit shall be one container fully prepared for delivery. The lot shall be the number of containers offered for inspection at one time. The inspection level shall be S-2 and the acceptable quality level shall be 4.0 defects per hundred units.

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

TABLE V. Examination for preparation for delivery

Examination	Defect
Containers	Not as specified
Quantity	Quantity in interior packages and shipping containers not as specified.
Packing and packaging materials	Any component missing or damaged.
Workmanship	Inadequate application of components, such as incomplete closure of container flaps, tape improperly applied, or containers distorted.
Markings	Omitted, incorrect, illegible, improper size, location, sequence, or method of application.

5. PACKAGING

5.1 Preservation. Preservation shall be level A or C, as specified (see 6.2).

5.1.1 Level A.

5.1.1.1 Unit package. Components comprising each orthopedic surgical kit shall be clean, dry and positioned within their respective compartments within the carrying case. The case shall be closed, secured and packaged in a snug fitting box conforming to PPP-B-636, class domestic. Closure shall be as specified in the box specification.

5.1.2 Level C. Each orthopedic surgical kit with components positioned within the carrying case shall be packaged in accordance with the requirements of MIL-STD-794 for level C.

5.2 Packing. Packing shall be level A, B or C, as specified (see 6.2).

5.2.1 Level A. Each orthopedic surgical kit packaged as specified in 5.1.1.1 shall be further packed in an exterior container designed for a type 2 load and conforming to PPP-B-601, overseas type or PPP-B-621, class 2. Closure and strapping shall be as specified in the applicable box specification.

5.2.1.1 Case liner. Each level A wood box shall be lined with a waterproof case liner conforming to MIL-L-10547. Closure and sealing shall be as specified in the liner specification.

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:45Z
Check the source to verify that this is the current version before use.

5.2.2 Level B. Each orthopedic surgical kit packaged as specified in 5.1.1.1 shall require no further packing.

5.2.3 Level C. Orthopedic surgical kits shall be packed in accordance with the requirements of MIL-STD-794 for level C.

5.2.4 Unitized loads. Unitized loads, commensurate with the level of packing specified in the contract or order, shall be used whenever total quantities for shipment to one destination equal 40 cubic feet or more. Loads shall be unitized on pallets conforming to MIL-STD-147. Quantities less than 40 cubic feet need not be unitized. Pallet loads including the pallet shall not exceed 43 inches in height, 48 inches in width, and 40 inches in length. Pallet shall conform to Type IV or Type V.

5.2.4.1 Levels A and B. Orthopedic surgical kits packed as specified in 5.2.1 and 5.2.2 shall be unitized on pallets as specified in 5.2.5.

5.2.4.2 Level C. Orthopedic surgical kits packed as specified in 5.2.3 shall be unitized as specified in MIL-STD-794.

5.3 Marking.

5.3.1 Levels A, B and C. Each unit pack, exterior container and unitized load shall be marked as specified in MIL-STD-129.

5.4 General.

5.4.1 Exterior container. Exterior container (see 5.2.1, 5.2.2 and 5.2.3) shall be of minimum tare and cube consistent with the protection required and shall contain equal quantities of identical stock numbered items to the greatest extent practicable.

5.4.2 Packaging inspection. The inspection of these packaging requirements shall be in accordance with 4.3.1.

6. NOTES

6.1 Intended use. The surgical kit is intended for use in orthopedic surgery.

6.2 Ordering data.

6.2.1 Acquisition requirements. Acquisition documents should specify the following:

a. Title, number, and date of this specification.
b. Requirements of components (see 3.2).
c. Preservation, packaging, packing and marking (see 5.1).

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

6.3 National Stock Number (NSN) coverage. This specification covers the following National Stock Number:

NSN	Item Identification
6515-01-177-4928	SURGICAL KIT, ORTHOPEDIC, Battery Operated, With Charger, Multipurpose.

6.4 Subject term (key word) listing.

Kit
Orthopedic kit
Surgical kit

MILITARY INTERESTS:

Custodians:

Army - MD
Navy - MS
Air Force - 03

Preparing Activity:

DoD - MB

Agent:

DLA - DM

CIVIL AGENCY COORDINATING ACTIVITIES:

VA-OSS
PHS
FDA-MPQAS

Project No. 6515-3372

Source: https://assist.dla.mil -- Downloaded: 202610-21T11:45Z
Check the source to verify that this is the current version before use.

INSTRUCTIONS: In a continuing effort to make our standardization documents better, the DoD provides this form for use in submitting comments and suggestions for improvements. All users of military standardization documents are invited to provide suggestions. This form may be detached, folded along the lines indicated, taped along the loose edge *(DO NOT STAPLE)*, and mailed. In block 5, be as specific as possible about particular problem areas such as wording which required interpretation, was too rigid, restrictive, loose, ambiguous, or was incompatible, and give proposed wording changes which would alleviate the problems. Enter in block 6 any remarks not related to a specific paragraph of the document. If block 7 is filled out, an acknowledgement will be mailed to you within 30 days to let you know that your comments were received and are being considered.

NOTE: This form may not be used to request copies of documents, nor to request waivers, deviations, or clarification of specification requirements on current contracts. Comments submitted on this form do not constitute or imply authorization to waive any portion of the referenced document(s) or to amend contractual requirements.

(Fold along this line)

(Fold along this line)

DEFENSE LOGISTICS AGENCY

NO POSTAGE NECESSARY IF MAILED IN THE UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE $300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 4966 WASHINGTON D. C.
POSTAGE WILL BE PAID BY THE DEFENSE LOGISTICS AGENCY

Commander
Defense Personnel Support Center
Attn: DPSC-QTS
2800 South 20th Street
Philadelphia, PA 19101

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.

STANDARDIZATION DOCUMENT IMPROVEMENT PROPOSAL

(See Instructions – Reverse Side)

1. DOCUMENT NUMBER	2. DOCUMENT TITLE
MIL-S-37972	SURGICAL KIT, ORTHOPEDIC

3a. NAME OF SUBMITTING ORGANIZATION	4. TYPE OF ORGANIZATION *(Mark one)*
	☐ VENDOR ☐ USER ☐ MANUFACTURER ☐ OTHER *(Specify)*: _____
b. ADDRESS *(Street, City, State, ZIP Code)*	

5. PROBLEM AREAS

a. Paragraph Number and Wording:

b. Recommended Wording:

c. Reason/Rationale for Recommendation:

6. REMARKS

7a. NAME OF SUBMITTER *(Last, First, MI)* – Optional	b. WORK TELEPHONE NUMBER *(Include Area Code)* – Optional
c. MAILING ADDRESS *(Street, City, State, ZIP Code)* – Optional	8. DATE OF SUBMISSION *(YYMMDD)*

DD FORM 82 MAR 1426 PREVIOUS EDITION IS OBSOLETE.

(TO DETACH THIS FORM, CUT ALONG THIS LINE.)

Source: https://assist.dla.mil -- Downloaded: 2026-10-21T11:45Z
Check the source to verify that this is the current version before use.