MIL-G-83794/2(USAF)
18 JANUARY 1974
Superseding
AFPID 6806
21 June 1966

MILITARY SPECIFICATION SHEET

# Gases, Calibrating, and Mixtures; Hydrocarbon (90 ppm) Calibrating in Oxygen

The complete requirement for procuring the calibrating gas described herein shall consist of this document and the issue in effect of specification MIL-G-83794(USAF).

Grade Designation: B

Major Component:

Oxygen, MIL-P-25508, Type 1, Purity and Moisture Requirements Only.

| Minor Component | Requirement | Test Method |
|---|---|---|
| Total Hydrocarbons as Methane | 90.0 ppm v/v | 4.6.2 |

Cylinder: DOT 3A or 3AA

Valve: MIL-V-2/39

Special Marking Requirement. The modifier "calibrating gas" shall be used in addition to the requirement in MIL-STD-101.

Intended Use: This calibrating gas is used as a span gas for total hydrocarbon analyzers.

Custodian:

Air Force - 68

Preparing Activity:

Air Force - 68

Project No. 6830-F083

☆ U. S. GOVERNMENT PRINTING OFFICE: 1974-713-151 3392

FSC 6830

Source: https://assist.dla.mil -- Downloaded: 2026-10-23T09:41Z
Check the source to verify that this is the current version before use.