**NNN–C–20**
**March 17, 1965**

*SUPERSEDING*
**Int. Fed. Spec. NNN–B–00595 (GSA–FSS)**
**March 2, 1964 and**
**Fed. Spec. RR–B–638**
**February 3, 1950**

FEDERAL SPECIFICATION

# CAN, PIPET STERILIZING

*This specification was approved by the Commissioner, Federal Supply Service, General Services Administration, for the use of all Federal agencies.*

## 1. CLASSIFICATION

**1.1 Types.** The pipet sterilizing cans covered by this specification shall be of the following types, as specified:

Type I—Tubular box.

Type II—Rectangular box.

**1.2 Compositions.** The pipet sterilizing cans shall be of the following compositions, as specified:

Composition A—Nickel-copper-alloy.

Composition B—Corrosion-resisting (stainless) steel.

## 2. APPLICABLE SPECIFICATIONS AND STANDARDS

**2.1 Specifications and standards.** The following specifications and standards, of the issues in effect on date of invitation for bids, form a part of this specification:

*Federal Specifications:*

QQ–N–281 — Nickel - Copper - Alloy (Monel and R-Monel) Bars, Plates, Rods, Sheets, Strips, Wire, Forgings, and Structural and Special Shaped Sections.

QQ–S–766—Steel Plates, Sheets, and Strip—Corrosion Resisting.

PPP–B–566—Boxes, Folding, Paperboard.

PPP–B–585—Boxes, Wood, Wirebound.

PPP–B–591—Boxes, Fiberboard, Wood-Cleated.

PPP–B–601—Boxes, Wood, Cleated-Plywood.

PPP–B–621—Boxes, Wood, Nailed and Lock-Corner.

PPP–B–636—Box, Fiberboard.

PPP–B–676—Boxes, Set-Up Paperboard.

PPP–T–76—Tape, Pressure - Sensitive Adhesive Paper, (For Carton Sealing).

*Federal Standards:*

Fed. Std. No. 102—Preservation, Packaging, and Packing Levels.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

Fed. Std. No. 123—Marking for Domestic Shipment (Civilian Agencies).

(Activities outside the Federal Government may obtain copies of Federal Specifications, Standards, and Handbooks as outlined under General Information in the Index of Federal Specifications and Standards and at the prices indicated in the Index. The Index, which includes cumulative monthly supplements as issued, is for sale on a subscription basis by the Superintendent of Documents, U.S. Government Printing Office, Washington, D.C. 20402.

(Single copies of this specification and other product specifications required by activities outside the Federal Government for bidding purposes are available without charge at the General Service Administration Regional Offices in Boston, New York, Washington, D.C., Atlanta, Chicago, Kansas City, Mo., Dallas, Denver, San Francisco, Los Angeles, and Seattle, Wash.

(Federal Government activities may obtain copies of Federal Specifications, Standards, and Handbooks and the Index of Federal Specifications and Standards from established distribution points in their agencies.)

*Military Specification:*

MIL–L–10547—Liners, Case, and Sheet, Overwrap, Water - Vaporproof or Waterproof, Flexible.

*Military Standards:*

MIL–STD–105—Sampling Procedures and Tables for Inspection by Attributes.

MIL–STD–129—Marking for Shipment and Storage.

(Copies of Military Specifications and Standards required by contractors in connection with specific procurement functions should be obtained from the procuring activity or as directed by the contracting officer.)

## 3. REQUIREMENTS

### 3.1 Material.

3.1.1 *Composition A.* Nickle-copper-alloy shall conform to the chemical requirements for class A sheets of QQ–N–281, shall be not less than 0.018 inch in thickness, and shall be given a satin finish.

3.1.2 *Composition B.* Corrosion-resisting (stainless) steel shall conform to the chemical composition for classes 302, 304, or 316 of QQ–S–766, shall be not less than 0.017 inch in thickness, and shall be given a No. 4 commercial finish.

### 3.2 Dimensions.

3.2.1 *Type I.* Type I pipet cans shall be of the dimensions specified in table I.

TABLE I. *Dimensions*

| Overall length | Outside diameter | Length of body | Length of lid |
|---|---|---|---|
| *Inches* | *Inches* | *Inches* | *Inches* |
| 13½ ± ¼ | 2½ ± 1/16 | 11 ± ⅛ | 4¾ ± ⅝ |
| 16¼ ± ¼ | 2½ ± 1/16 | 14 ± ⅛ | 4¾ ± ⅝ |

3.2.2 *Type II.* Type II pipet cans shall be of the dimensions specified in table II.

TABLE II. *Dimensions*

| Overall length | Width and depth | Length of body | Length of lid |
|---|---|---|---|
| *Inches* | *Inches* | *Inches* | *Inches* |
| 13½ ± ¼ | 2½ ± 1/16 | 11 ± ⅛ | 4¾ ± ⅝ |
| 16¼ ± ¼ | 2½ ± 1/16 | 14 ± ⅛ | 4¾ ± ⅝ |

### 3.3 Design.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

3.3.1 *Body.* The body shall have a bead near the top edge to serve as a stop for the lid. The body shall be joined with an inside lock seam and the bottom shall be double seamed on the body. The open end shall be finished so that sharp cutting edges are smooth.

3.3.2 *Lid.* The lid shall fit smoothly and snugly over the body. The lid shall be joined with an outside lock seam and the top shall be double seamed in the cover. The open end shall be finished so that sharp cutting edges are smooth.

**3.4 Marking.** Each pipet can shall be legibly and permanently marked with the name or registered trademark of the manufacturer.

**3.5 Workmanship.** The pipet cans shall be uniformly made. The pipet cans shall be free of holes, cracks, corrosion, dents, nicks, abrasions, dirt, grease, scratches, or other defects which detract from their appearance or which may impair their serviceability.

## 4. SAMPLING, INSPECTION, AND TEST PROCEDURES

4.1 Unless otherwise specified in the contract or purchase order, the supplier is responsible for the performance of all inspection requirements as specified herein. Except as otherwise specified, the supplier may utilize his own facilities or any commercial laboratory acceptable to the Government. The Government reserves the right to perform any of the inspections set forth in the specification where such inspections are deemed necessary to assure that supplies and services conform to prescribed requirements.

4.1.1 *Inspection.* Inspection, as used in this specification, is defined as both examination (such as visual, auditory, etc., investigation without the use of special laboratory appliances or procedures) and testing (determination by technical means of physical and chemical properties) of the item.

4.1.2 *Certificates of quality.* When available, certificates of quality, supplied by the manufacturer of the component or material, listing the specified test method and test results obtained, may be furnished in lieu of actual performance of such testing by the contractor.

**4.2 Sampling.**

4.2.1 *For examination.* Sampling for examination shall be conducted in accordance with MIL-STD-105.

| | Inspection level | AQL (percent defective) |
|---|---|---|
| **For visual examination:** | | |
| **Major defects ....** | **II** | **4.0** |
| **Minor defects ....** | **II** | **6.5** |
| **For dimensional examination.** | **S-2** | **2.5** |

4.2.2 *For tests.* Sampling for tests shall be conducted in accordance with MIL-STD-105.

**4.3 Examination.** Sample pipet cans shall be examined in accordance with the classification of defects and with MIL-STD-105.

4.3.1 *Classification of defects.*

4.3.1.1 *Pipet cans*



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

| Categories | Defects |
|---|---|
| *Critical:* | None defined. |
| *Major:* | *AQL 4.0 percent defective* |
| 101 | Bead of body does not serve as a stop for lid. |
| 102 | Edges of open end of body and lid not smooth. |
| 103 | Lid does not fit smoothly and snugly over body. |
| 104 | Pipet can not uniformly made. |
| 105 | Pipet can not free of holes, cracks, corrosion, dents, nicks, abrasions, dirt, grease, etc. |
| *Minor:* | *AQL 6.5 percent defective* |
| 201 | Pipet can not free of scratchs. |
| 202 | Identification marking missing, illegible, incorrect, or incomplete. |

4.3.1.2 *Inspection of preparation for delivery requirements.* An inspection shall be made to determine that the packaging, packing, and marking requirements comply with section 5. Defects shall be scored in accordance with table III. For examination of interior packaging, the sample unit shall be one shipping container fully prepared for delivery selected at random just prior to the closing operations. Sampling shall be in accordance with MIL–STD–105. Defects of closure listed shall be examined on shipping containers fully prepared for delivery. The lot size shall be the number of shipping containers in the end item inspection lot. The inspection level shall be II and the AQL shall be 4.0 defects per hundred units.

TABLE III. *Classification of preparation for delivery defects*

| Examine | Defects |
|---|---|
| Markings (exterior and interior) | Omitted; incorrect; illegible; improper size, location, sequence, or method of application. |
| Materials | Any component missing or damaged. |
| Workmanship | Inadequate application of components such as incomplete closure of container flaps, loose strapping, inadequate sealing. Bulging or distortion of container. |

4.3.1.3 *Dimensional examination.* The pipet cans shall be examined for defects in dimensions. Any dimension not within the tolerance specified herein shall be classified as a defect.

4.4 **Tests.** Tests for material shall be as specified in 3.1.1 and 3.1.2.

## 5. PREPARATION FOR DELIVERY.

The definitions and applications of the various levels of packaging and packing for civil agency procurements shall be in accordance with Fed. Std. No. 102.

### 5.1 Packaging.

5.1.1 *Levels A and B.* Pipet boxes shall be individualy packaged in paperboard folding or set up boxes conforming to PPP–B–566 and PPP–B–676, respectively. Each box shall be closed in a manner to prevent accidental opening.

5.1.2 *Level C.* Pipet boxes shall be packaged in accordance with the suppliers commercial practice.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

**5.2 Packing.**

5.2.1 *Level A.* Thirty-six pipet boxes of the same type and composition, packaged as specified in 5.1.1, shall be packed in any of the following containers at the option of the contractor:

| Federal Specification | Type and class |
|---|---|
| PPP–B–585 | Class 3 use |
| PPP–B–591 | Overseas type |
| PPP–B–601 | Overseas type |
| PPP–B–621 | Class 2 |
| PPP–B–636 | Class weather-resistant |

Containers shall be provided with case liners conforming to MIL–L–10547, closed and sealed in accordance with the appendix thereto. Case liners may be omitted for fiber boxes conforming to PPP–B–636 provided all corners, edge seams and manufacturer's joints are sealed with minimum 1½ inch wide tape conforming to PPP–T–76. Boxes shall be closed and strapped in accordance with the applicable box specification or appendix thereto, except fiberboard boxes may be banded with filament reinforced tape in accordance with the appendix to PPP–B–636. The gross weight of wood or wood cleated boxes shall not exceed 200 pounds.

5.2.2 *Level B.* Thrity-six pipet boxes of the same type and composition, packaged as specified in 5.1.1, shall be packed in a fiberboard box conforming to PPP–B–636, type CF, class domestic, style RSC, bursting strength of not less than 275 p.s.i. Closure shall be in accordance with the appendix to the box specification.

5.2.3 *Level C.* Pipet boxes shall be packed in accordance with the manufacturer's standard commercial practice in containers conforming to carrier rules and regulations applicable to the mode of transportation.

**5.3 Marking.**

5.3.1 *Military activities.* In addition to any special marking required by the contract or order, unit and exterior shipping containers shall be marked in accordance with the applicable paragraphs of MIL–STD–129.

5.3.2 *Civil agencies.* In addition to any special marking required by the contract or order, unit and exterior shipping containers shall be marked in accordance with Fed. Std. No. 123.

**6. NOTES**

**6.1 Ordering data.** Purchasers should exercise any desired options offered herein, and procurement documents should specify the following:

(*a*) Title, number, and date of this specification.

(*b*) Type and composition required (see 1.1 and 1.2).

(*c*) Selection of the applicable levels of packaging and packing (see 5.1 and 5.2).

**6.2 Transportation description.** Transportation description and minimum weights applicable to this item are:

*Types I and II, composition A.*

Copper articles, not otherwise indexed by name, not nested.

Carload minimum weight 14,000 pounds, subject to Rule 34, Uniform Freight Classification.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

Copper articles, not otherwise indexed, not nested.

Truckload minimum weight 14,000 pounds, subject to Rule 115, National Motor Freight Classification.

*Types I and II, composition B.*

Sheet steel ware, not otherwise indexed by name, without copper bottoms.

Carload minimum weight 20,000 pounds, subject to Rule 34, Uniform Freight Classification.

Sheet steel articles, not otherwise indexed.

Truckload minimum weight 20,000 pounds, subject to Rule 115, National Motor Freight Classification.

**MILITARY CUSTODIANS:**

**Army—MD**

**Navy—MS**

**Air Force—03**

*Review activities:*

**Army—MD**

**Navy—MS**

**Air Force—03**

**Defense Supply Agency—DM**

Military coordinating activity:

**Defense Supply Agency—DM**

Review information is current as of the date of this document. For future coordination of changes to this document, draft circulation should be based on the information in the current Federal Supply Classification Listing of DOD Standardization Documents.



Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.