INTERIM AMENDMENT

TO

FEDERAL SPECIFICATION

CAN, PIPET STERILIZING

This interim amendment was developed by the Defense Personnel Support Center, 2800 South 20th Street, Philadelphia, PA 19101, based upon currently available technical information. It is recommended that Federal agencies use it in procurement and forward recommendations for changes to the preparing activity at the address shown above.

The General Services Administration has authorized the use of this interim amendment as a valid exception to Federal specification NNN-C-20.

Page 1:

1.1 Types. Shall be type I.

1.2 Compositions. Shall be composition B.

Page 2:

3.1.2 Composition B. Line 6, delete "4 commercial finish." and substitute "2B or No. 4 commercial finish."

3.2.1 Type I. Table I. Dimensions. Shall be 16-1/4 ± 1/4 inches long; overall, outside diameter shall be 2-1/2 ± 1/16 inches, length of body shall be 14 ± 1/8 inches; and length of lid shall be 4-3/4 ± 3/8 inches.

FSC 6640

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.

Page 3:

Add the following paragraph:

"4.1.3 Records. Records of examinations and tests performed by or for the contractor shall be maintained by the contractor and made available to the Government, upon the Government's request, at any time, or from time to time, during the performance of the contract and for a period of three years after delivery of the supplies to which such records relate.

Preparing activity:
DSA-DM

Project No. 6640-0992

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.