AMENDMENT

TO

INTERIM FEDERAL SPECIFICATION

RACK, TEST TUBE, LABORATORY

This amendment was developed by the General Services Administration, Federal Supply Service, Washington, DC 20406, based upon currently available technical information. The General Services Administration has authorized this amendment as a part of Interim Federal Specification NNN-R-0055 (DM), dated January 23, 1968.

PAGE 8

Paragraph 3.2.3: Delete the first and second sentences and substitute the following:

"The test tube support rack shall be of a steel wire assembly 4 inches high ± 1/16 inch by 4 inches wide ± 1/16 inch by 10 inches long ± 1/16 inch (4 openings wide by 10 openings long). The measurements for the width and length shall be considered as inside dimensions, and the height shall be considered as the outside dimension."

FSC 6640

GSA DC 71-14051

Source: https://assist.dla.mil -- Downloaded: 2026-10-26T12:48Z
Check the source to verify that this is the current version before use.